



PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
 BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: PHC KEEZHPALLY	Name of District: KANNUR
Name of Equipment: CENTRIFUGE	Equipment ID & Barcode: #140956-1343461
Make: REMI	Model: R-4C
Serial Number: AECC688	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: In ticket master installation date is 10.02.2020, as per stock book installation date is 25.05.2018(stock book copy attached for reference)	Present status of the equipment: Fully damaged
Date of breakdown: 16.10.2024 Date of registration of complaint through email/ Toll free: Toll free	*AMC/CAMC Period agreed at the time of purchase: NO AMC/ CAMC
Action taken: Checked the Equipment found Equipment is not Working, found that Motor, speed regulator defective. Need to replace these spares for further checking.	
Recommendations for repair (required service details): No recommendations	
Cost of spares (specify parts and cost): KNOB FOR SPEED REGULATOR-58/- ELECTRIC MOTOR-6443/- TOTAL- 6501/-	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 67.36%	Asset Value: 9651/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix Service report & OEM quotation attached	
Reasons for recommending the equipment as BER: Checked the Equipment found Equipment is not Working, found that Motor, speed regulator defective. Need to replace these spares for further checking. In software Equipment Installed On 10.02.2020, as per stock book installation date is 25.05.2018 aged up to 6 years 5 months. Spare cost is 67.36% Both criteria for RBER met. As per tender clause 5.3.14.1 we recommend the Equipment for condemnation.	
Name & Signature of CYRIX Authority with date	
Remarks of Junior Consultant (Biomedical) NHM: checked & identified that machine not working due to the motor, regulator are defective. its service cost not economically feasible. and aged up to 6 years & above. so recommended for BER.	
Recommended for BER (Yes/ No):	 11/11/24 Signature of JC BM (NHM)
Date: 11/11/24	
Date:	 Signature of Superintendent / Medical Officer (M/O)

*Not mandatory

Based on the period of life and value as per the BER guidelines

*Attach Photograph



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1011597

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility...PHC...KEEZHPALLY.....	Call Registration Date : ...15/10/2024.....
Address.....KEEZHPALLY.....	Caller ID :140956.....
.....KANNUR.....	Date of Visit :17/10/2024.....
Ph :9848131420.....	Asset No :1343461.....
	EQPT Name :Centrifuge.....
	Manufacture.....REM..... Model :R-4C.....
	S. No.....96668..... Dept.....Laboratory.....

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified :Equipment is not switching ON.....

Action Taken :checked the equipment. found equipment is not switching ON. Found that motor, speed regulator defective. Need to replace these spares for further checking.....

Completed Date : ...17/10/2024..... Time : ...1:00pm..... Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Pradyush T. Bhaskar	17/10/2024	12:20pm	1:00pm

Customer Remark Completed Pending

Service Engineer Name : Pradyush T. Bhaskar

Signature: *[Signature]*

Date: 17/10/2024

Contact Number: 9947028643

Customer Name : *[Signature]*

Signature: *[Signature]*

Date: 17/10/24

Contact Number: 9848131420

Designation : Lab technician grade I

Hospital Seal

MEDICAL OFFICER
PHC, KEEZHPALLY
KANNUR DT. - 670704

REMI SALES & ENGINEERING LIMITED



48/1103, Puthiyedath Road, Vytilla P.O, Ernakulam Kochi-682019
 Phone : 0484-682019, Mob: 8921409347 Email Id.: instservicekochi@remigroup.com/service@remilabworld.com

Corporate Off : Remi House, 3rd floor, 11 Cama Indl. Estate, Goregaon (East) Mumbai 400 063. INDIA
 • Board: +91 22 4058 9888 SALES SUPPORT NO.8879448682 • Email Id.: sales@remilabworld.com

QUOTATION CUM PROFORMA INVOICE

QN- OCT-25/24-25

Complaint No.	Customer Name & Address.:					Quotation No	PRINTED		
Dated	26/10/2024	CYRIX HEALTHCARE (P) LTD					Date	26/10/2024	
ENGINEER NAME	vipin	30/641B Petta Punitthura Kochi Kerala- 682038					REF	Phone	
REF.	Phone	Avinash T						DATE	26/10/2024
Email	avinash.t.cavinash@cyrix@gmail.com		6364064948						
Model no.	REMI : R-4C			SR NO. /MFG DT.					
Sr No.	Description	Part code	HSN/SAC	Rate Rs.	Qty	Amount Rs. (A)	GST % (B)	GST Amt (C)	Total Amt (A+C)
	KNOB FOR SPEED REGULATOR (NRT)	L22000116	392390	49	1	49	18%	9	58
	ELECTRIC MOTOR (RT)	ML25000140	850140	5460	1	5460	18%	983	6443
TOTAL(D)									6501
	CALIBERATION CHARGE	0	0	0	0	0	18%	0	0
	SERVICE CHARGE	0	0	0	0	0	18%	0	0
TOTAL (E)									6501
Rupees in Words:		Rs: Six thousand Five hundred one only .							

Terms & Conditions:

- Spares are suitable only to use with REMI instruments & not for any other make.
- We will accept return of spares only if they are unused and not damaged. At our discretion, the returned spare will be replaced. The difference in prices. If any will have to be borne by you.
- To ensure supply of correct spares, we suggest you to send us defective spare in advance with your request for which you want to purchase new spares.
- 100 % Advance along with your order confirmation
- Payment can be made by Cheque/ NEFT in our Bank account. For Cash Payment you must collect Receipt from recipient of cash failing which REMI will not be held responsible for any discrepancy.
- Courier/ transportation charges are payable at destination & not included in our price.
- Validity. The quotation is valid for 30 days from the date of this quotation / Proforma Invoice
- Subject to Ernakulam Jurisdiction only

REMARKS : PAYMENT 100% ADVANCE

For REMI SALES & ENGINEERING LTD

Retheesh PR
 Authorised Signatory

BANK DETAILS FOR REMI SALES & ENGINEERING LTD.

- A/C NO: 35699154671
- Bank Name: STATE BANK OF INDIA
- BRANCH: ELAMAKKARA
- IFS CODE: SBIN0013223
- MICR CODE: 682002053
- SWIFT CODE: SBININBB395
- GST NO: 32AAACR0421M1Z5
- PAN NO: AAACR0421M
- CIN NO. U31100MH198PL022314



QUOTATION REQUIRED FOR CENTRIFUGE #1343461,140956 @PHC KEEZHPALLY

2 messages

avinash t <avinashtcyrix@gmail.com>
To: Instservicekochi@remigroup.com

Fri, 25 Oct, 2024 at 1:04 pm

Dear sir,

Please provide the stock availability and quotation of below mentioned part of centrifuge(make-REMI, model-R4C, SL NO-AECC688)

1. MOTOR
2. SPEED REGULATOR

Avinash T
District charge, kannur
bemp kerala
mobile: +91 6364064948
E mail: avinashtcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B | Petta | Punithura | Kochi | Kerala- 682038
www.cyrixhealthcare.com

instservicekochi <instservicekochi@remigroup.com>
To: avinashtcyrix@gmail.com
Cc: V. Raman <raman@remilabworld.com>, kochisales.hcd@remigroup.com, kochi@remigroup.com

Sat, 26 Oct, 2024 at 11:28 am

Dear Sir,

Please find the attached PERFORMA INVOICE for your reference. **PLEASE RELEASE THE PAYMENT IN ADVANCE.**

Regards

REMI KOCHI

INST DIV.

[Quoted text hidden]

10

Centrifuge

		Revised	used	Balance
31-10-09		01	01	
25/5/18	KMC - Kochin lab equipment Invoice No - K1/0389/18-19	01		
	27,376/-			
18/12/19	received from store	01		
Transfer out to stock register volume 1				
Page No - 10				



✓
 MEDICAL OFFICER
 CHC, KEEZHAPALLY
 KANNUR DT. - 670704