


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KASARGOD
2	Name of Hospital	PHC PADNA
3	Name of Equipment with Make, Model and Serial Number	SEMI AUTO BIOCHEMISTRY ANALYZER MAKE: AGD BIOMEDICALS MODEL: AGD 2020 S/N: A0161214
4	Equipment ID & Barcode	124562 / 1440715
5	Date of purchase / Year of manufacture / Installation Date	Ticket master installation date is 10-02-2017 As per stock book installation is 19-01-2017 (Stock book attached for reference)
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	22-07-2024
9	Action taken	Checked the machine, found Display with touch and main board are defective. Enquired OEM for quotation.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No Recommendation
12	Cost of spares (specify parts and cost)	EA 7 INCH DISPLAY ASSEMBLY WITH TOUCH – 24780/- EA000019 PCB ASSEMBLY MCB 2020-20060/- TOTAL- Rs.44840/-

MEDICAL OFFICER
PHC PADNA
KASARGOD DIST



13	Asset Value	₹ 65000
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	68.98%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report and OEM quotation attached.
16	Reasons for recommending the equipment as BER	Checked the machine, found machine main board and display with touch are defective. Ticket master installation date is 10-02-2017 As per stock book installation date is 19-01-2017 (Stock book attached for reference) aged up to 7 years and 4 months. Repairing cost of equipment is 68.98% with respect to asset value. Both criteria for RBER met. As per the tender clause 5.3.14.1 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	ABHINAV O <i>Abhinav O</i>

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

checked the machine and found main board and display with touch are defective. Equipment installed on 10/02/2017 (stock book details attached). Repairing cost is 68.98% of asset value. so recommending the equipment for BER. cyrix service report and oem quotation attached.

Signature of JC B.M. (NHM) *per*
**Biomedical Engineering
 Arogyakeralam (NHM
 Kasaragod at Kanhangad**

Date



Signature
 Signature of
 Superintendent / Medical Officer (i/c)
**MEDICAL OFFICER
 P H O PADNE
 PIN: 671 312
 KASARAGOD DIST**



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 209219

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <i>PHC Padna</i>	Call Registration Date : <i>22.07.2024</i>
Address <i>Padna</i>	Caller ID : <i>124562</i>
<i>Kasaragod</i>	Date of Visit : <i>23.07.2024</i>
Ph : <i>8137933401</i>	Asset No. : <i>1440715</i>
	EQPT Name : <i>Biochemistry Analyzer</i>
	Manufacture <i>AGD Biomedical</i> Model : <i>AGD 2020</i>
	S. No. <i>A061214</i> Dept. <i>Laboratory</i>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : *Touch not working, machine stuck (hanging)*

Action Taken : *Checked the machine and found problem with machine mainboard. Need to replace the mainboard for further servicing of the equipment.*


Completed Date : *23.07.24* Time : *12.30pm* Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<i>Abhinav. O</i>	<i>23.07.24</i>	<i>10.40 AM</i>	<i>12.30 PM</i>

Customer Remark _____ Completed Pending



Service Engineer Name : *Abhinav. O*
Signature : *[Signature]*
Date : *23.07.2024*
Contact Number : *8943392989*

Customer Name : *Vipina. K V*
Signature : *[Signature]*
Date : *23-7-2024*
Contact Number : *81 37 933401*
Designation : *LI*
Hospital Seal : *[Seal]*

MEDICAL OFFICER
P H O PADNE
PIN: 671 312
KASARAGOD DIST

Quotation

AGD BIOMEDICALS PVT. LTD.

MEHTA TRADE CENTRE, No. 1 SHIVAJI COLONY, SIR M V ROAD, -, ANDHERI
EAST, MUMBAI - 400099 (MAHARASHTRA) INDIA.
Phone. : 9833628002

GST No. :27AABCA2897B1ZU



Lab Name :
P. H. C. PADNE HOSPITAL.
Address : ,PADANNA,,KASARAGOD,
Pin Code.: 671 312.
State.: KERALA
Mo.:
E-Mail.:
GST NO.:

Quotation No.: QO/AGD/0309

Quotation Date : 23/08/2024

Client PO No:

Client PO Date : 23/08/2024

Representative Name :

AMAL MOHAN K

Mo: 9633389673

SR.No	Description of Goods	Quantity	Rate(₹)	Amount(₹)	CGST%	CGST Amount(₹)	SGST%	SGST Amount(₹)	IGST%	IGST Amount(₹)	Net Amount(₹)
1	EA 7 INCH DISPLAY ASSEMBLY WITH TOUCH	1 NOS	21000.00	21,000.00	0.00	0.00	0.00	0.00	18.00	0.00	24,780.00
2	EA000019 PCB ASSEMBLY MCB 2020	1 NOS	17000.00	17,000.00	0.00	0.00	0.00	0.00	18.00	0.00	20,060.00
	PA000015										
Sub Total(₹)										38,000.00	
Total IGST										₹ 6,840.00	
Total(₹)										44,840.00	

Amount Chargeable (in words)

Rupees Forty-Four Thousand Eight Hundred and Forty Point zero Only.

Remarks :

Authorised Sign



Jayesh Krishnan <jayeshcyrix@gmail.com>

PRICE ENQUIRY 1440715

2 messages

Jayesh Krishnan <jayeshcyrix@gmail.com>

Mon, Sep 9, 2024 at 9:01 PM

To: kerala Branch <keralabranh@agdbio.com>

Cc: Amal.Mohan@agdbio.com, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear Sir/madam,

Kindly share the price and availability of main board and touch for biochemistry analyzer.

Make: AGD

model: AGD 2020

S/N: A0161214

--

Thanks & Regards

JAYESH KRISHNAN,
Sr. Biomedical Engineer,
District In-charge
Kasargod.
PH:- 7593847156

jayeshcyrix@gmail.comwww.cyrixhealthcare.com**Cyrix Healthcare Pvt Ltd****| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India****KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH****Simi K S** <keralabranh@agdbio.com>

Tue, Sep 10, 2024 at 10:47 AM

To: Jayesh Krishnan <jayeshcyrix@gmail.com>

Cc: Amal Mohan <Amal.Mohan@agdbio.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear Jayesh,

Please find the attached Quotation.

**Simi K. S**

Office Admin

AGD Biomedicals Pvt. Ltd.

Near Rubber Estate (V.R. Puram)

Potta Ashramam Road, Chalakudy.

Thrissur, Kerala . Pin : 680722, India

Mob No. +91 8356912368 / 8593030603



keralabranh@agdbio.com | www.agdbio.com**From:** Jayesh Krishnan <jayeshcyrix@gmail.com>**Sent:** Monday, September 9, 2024 8:31 AM**To:** Simi K S <keralabranh@agdbio.com>

Cc: Amal Mohan <Amal.Mohan@agdbio.com>; Blesson Jose <zm2.klbemp@cyrix.in>; DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Subject: PRICE ENQUIRY 1440715

[Quoted text hidden]

 **Quotation (24).pdf**
435K



X *[Handwritten Signature]*

MEDICAL OFFICER
PHC PADNE
PIN: 671 312
KASARAGOD DIST