

**PROFORMA  
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)  
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: TH PANATHADY	Name of District: KASARGOD
Name of Equipment: INFRARED LAMP	Equipment ID & Barcode: #138675- 1420506
Make: BEURER	Model: N/A
Serial Number: N/A	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 24/09/2013	Present status of the equipment: Fully damaged
Date of breakdown: 04/10/2024	*AMC/ CAMC Period agreed
Date of registration of complaint through email/Toll free): Toll free	at the time of purchase: No AMC/CAMC
Action taken: Checked the IR lamp, found lamp and cable defective. Enquired vendor for Lamp quotation.	

Recommendations for repair (required service details): Not recommending for repair

Cost of spares (specify parts and cost): IR BULB- 850/-

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 171.71%	Asset Value:495/-
--	-------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service report and OEM quotation attached

Reasons for recommending the equipment as BER: Checked the equipment, found problem with IR lamp and cable. Machine installed on 24-09-2013 and aged up to 11 years. Repairing cost is above 171% of asset value. Both criteria for RBER met. Considering the tender clause 5.3.14.1, We recommending the equipment for condemnation

ABHINAV O *[Signature]*  
19.10.2024  
Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM: Checked the machine and found problem with IR lamp and cable. Equipment installed on 24/09/2013. Repairing cost is above 171% of asset value. Attached cyrix service report and OEM quotation.

**JR CONSULTANT**  
**Biomedical Engineering**  
**Arogyakeralam (NHM)**  
**Kasaragod at Kanhangad**

Recommended for BER (Yes/ No):  Yes

Date: 19-10-2024

*[Signature]*  
Signature of JC BM (NHM)

Date: 24-10-24 Seal

*[Signature]*  
Signature of Superintendent / Medical Officer (i/c)

\*Not mandatory #Based on the period of life and value as per the BER guidelines \* Attach Photograph



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

No : 1015143

**CYRIX**<sup>®</sup>  
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

<p>Health Facility..... <u>Taluk Hospital</u>.....</p> <p>Address..... <u>Panathady</u>.....</p> <p style="padding-left: 40px;"><u>Kasaragod</u>.....</p> <p>Ph:..... <u>7795757907</u>.....</p>	<p>Call Registration Date :..... <u>04.10.2024</u>.....</p> <p>Caller ID :..... <u>138675</u>.....</p> <p>Date of Visit :..... <u>04.10.2024</u>.....</p> <p>Asset No :..... <u>1420506</u>.....</p> <p>EQPT Name :..... <u>Infrared lamp</u>.....</p> <p>Manufacture..... <u>Beurer</u>..... Model :..... <u>NA</u>.....</p> <p>S. No..... <u>NA</u>..... Dept..... <u>physiotherapy</u>.....</p>
--	--

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : Not switching on

Action Taken : Checked the machine and found that machine lamp and wire are defective. Need to change these spares for further servicing of the equipment.

Completed  Date : 04.10.24 Time : 2:00pm Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Abhinav O</u>	<u>04.10.24</u>	<u>1.00pm</u>	<u>2.00pm</u>

Customer Remark Completed  Pending

Service Engineer Name : Abhinav O

Signature: [Signature]

Date: 04.10.2024

Contact Number: 8943392989

Customer Name : Soumya


Signature: [Signature]

Date: 04.10.2024

Contact Number: 9847299500

Designation : Physiotherapist

Hospital Seal





# Sreeraj Surgicals

GST No: 32ABYF53674H12L

DI No : 13/023/20B/2011

13/024/21B/2011

Dt.: 29-01-2011

Aiswarya Commercial Complex, Nr. Ashirvad Hospital, Thavakkara, Kannur - 670 002

Ph: 0497 2706307. Mob: 9349846307, 9387529848

Email: sreerajsurgicals@gmail.com, rajeshsreerajsurgical@gmail.com

www.sreerajsurgical.com Follow us on:  www.facebook/ sreerajsurgicals

Date:

## QUOTATION

DATE: 16-10-2024

TO

CYRIX HEALTH CARE PVT LTD

SIR,

THANK YOU FOR YOUR ENQUIRY TO THE RATE OF THE FOLLOWING ITEMS.

SL.NO	ITEMS	RATE
1	INFRARED BULB ONLY	850.00

WE EXPECT YOUR VALUABLE ORDERS AND KIND CO-OPERATION WITH US

GST INCLUDED



THANKING YOU

Authorised Dealer:



Suppliers of Surgical instruments, Hospital Furnitures, X-ray Materials, Lab Equipments & Disposables



Jayesh Krishnan &lt;jayeshcyrix@gmail.com&gt;

---

**QUOTATION REQUEST FOR IR LAMP 1420506**

6 messages

---

**Jayesh Krishnan** <jayeshcyrix@gmail.com>

Thu, Oct 10, 2024 at 4:32 PM

To: "Sreerajsurgicals@gmail.com" &lt;Sreerajsurgicals@gmail.com&gt;

Cc: DM Sarang-KLBEMP &lt;dm1.klbemp@cyrix.in&gt;, Blesson Jose &lt;zm2.klbemp@cyrix.in&gt;

Dear sir,

Kindly send price and availability of IR lamp.

Requirement: Infrared Lalo

Make: Beurer

---

**Jayesh Krishnan** <jayeshcyrix@gmail.com>

Sat, Oct 12, 2024 at 10:44 AM

To: "Sreerajsurgicals@gmail.com" &lt;Sreerajsurgicals@gmail.com&gt;

Cc: DM Sarang-KLBEMP &lt;dm1.klbemp@cyrix.in&gt;, Blesson Jose &lt;zm2.klbemp@cyrix.in&gt;

Sir,

Kindly send the quotation.

[Quoted text hidden]

--

Thanks &amp; Regards

JAYESH KRISHNAN,  
Sr. Biomedical Engineer,  
District In-charge  
Kasargod.  
PH:- 7593847156

[jayeshcyrix@gmail.com](mailto:jayeshcyrix@gmail.com)[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)**Cyrix Healthcare Pvt Ltd****| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India****KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH**

---

**Jayesh Krishnan** <jayeshcyrix@gmail.com>

Mon, Oct 14, 2024 at 8:47 PM

To: "Sreerajsurgicals@gmail.com" &lt;Sreerajsurgicals@gmail.com&gt;

Cc: DM Sarang-KLBEMP &lt;dm1.klbemp@cyrix.in&gt;, Blesson Jose &lt;zm2.klbemp@cyrix.in&gt;

Sir,

Kindly update.

[Quoted text hidden]

---

**Jayesh Krishnan** <jayeshcyrix@gmail.com>

Wed, Oct 16, 2024 at 10:18 AM

To: "Sreerajsurgicals@gmail.com" &lt;Sreerajsurgicals@gmail.com&gt;

Cc: DM Sarang-KLBEMP &lt;dm1.klbemp@cyrix.in&gt;, Blesson Jose &lt;zm2.klbemp@cyrix.in&gt;

Sir,

Kindly send the quotation for Infrared bulb only.

[Quoted text hidden]

---

**sreerajsurgicals** <sreerajsurgicals@gmail.com>

Wed, Oct 16, 2024 at 10:22 AM

To: Jayesh Krishnan &lt;jayeshcyrix@gmail.com&gt;

Sent from Mail for Windows

[Quoted text hidden]

---

 **img905.pdf**  
137K

---

**sreerajsurgicals** <sreerajsurgicals@gmail.com>  
To: Jayesh Krishnan <jayeshcyrix@gmail.com>

Wed, Oct 16, 2024 at 12:31 PM

Sent from Mail for Windows

---

**From:** Jayesh Krishnan  
**Sent:** 16 October 2024 10:18 AM  
**To:** Sreerajsurgicals@gmail.com  
**Cc:** DM Sarang-KLBEMP; Blesson Jose  
**Subject:** Re: QUOTATION REQUEST FOR IR LAMP 1420506

Sir,

[Quoted text hidden]

[Quoted text hidden]

---

 **img907.pdf**  
137K



...  
 ...  
 ...  
 ...

# PHYSIOTHERAPY ITEMS

No. 078

Date	Issue Note No	MFG	EXP	Batch No	Receipt	Issue	Balance	Remarks
24.9.13	Received from NRMH Dept.					10000001	(14)	
Physiotherapy Dept	R.026	(1)	LR lamp	1 Nos		RS. 1100/-		
	R.027	(2)	CP stand	1 Nos		RS. 7400/-		
	R.028	(3)	CP chair	1 Nos		RS. 7400/-		
	R.029	(4)	Physio Ball (B)	1 Nos		RS. 1500/-		
	R.030	(5)	Physio Ball (S)	1 Nos		RS. 1240/-		
	R.031	(6)	Hand Ball	1 Nos		RS. 200/-		
	R.032	(7)	Shoulder wheel	1 Nos		RS. 7800/-		✓
22.11.13	Issued above (7) items to Physiotherapy Unit Nil							
	Ray				Received 7 items correctly Dinesh Rose Jeany Physiotherapist			
					✓			
27.1.18	Bill No: 215 dtd 22.01.2018 Sherath's Medical Systems, Chennai Ph: 044-24516904, 24516539							
	(1)	Traction - by Physiotherapist	Blow	(1)		RS. 18000.00		
	(2)	Traction - Table		(1)		RS. 11500.00		
	(3)	Wax Bath		(1)		RS. 7000.00		
	(4)	Hand Exerciser (Animal)		(1)		RS. 18000.00		
	(5)	Balance Board		(1)		RS. 2600.00		
					Total	RS. 63952.00		
					Cost	RS. 6852/-		
27.1.18	Pend above					Received 5 items		
	(5)	Hand to Physiotherapy Unit			Nil	(correctly)		
						Kinta. K. Mathan Physiotherapist.		

Contd Page 191



മെഡിക്കൽ ഓഫീസർ  
 സെക്രട്ടറി  
 സി. സെന്റർ കൗൺസിലർ ജിജി