

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: THQH KUNNAMKULAM	Name of District: Thrissur
Name of Equipment: Nebulizer	Equipment ID & Barcode: #146944- 0822779
Make: S Cure	Model: NA
Serial Number: NA	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: Ticket master installation date -25-02-2019 Stock book installation date- 10-03-2019	Present status of the equipment: Fully damaged
Date of breakdown: 15-11-2014 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: The machine checked and found motor and piston are defective. Need to replace these spares for further checking of equipment.	


Recommendations for repair (required service details): Not Recommending

Cost of spares (specify parts and cost): NA



#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 1353/-
---	---------------------


Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix report attached.

Reasons for recommending the equipment as BER: The machine checked and found motor and piston are defective. In ticket master installation date is 25/02/2019, As per the stock book installation date is 10/03/2019 aged up to 5 years 8 months. Quotation not submitted since the spares are not available in the market. So, we recommend the equipment for condemnation.



BIJO T JOY
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:
Piston and motor are defective

Recommended for BER (Yes/ No): <i>Yes</i>	  Signature of JC BM (NHM)
Date:	

Date: <i>25/11/2024</i>	Seal	 Signature of Superintendent / Medical Officer (i/c)
-------------------------	------	--

*Not mandatory //Based on the period of life and value as per the BER guidelines * Attach Photograph



SUPERINTENDENT
TALUK HOSPITAL
KUNNAMKULAM



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1006069



ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility: <u>J.HQH...Kunnampulam</u>	Call Registration Date: <u>15/11/24</u>
Address: <u>Kunnampulam</u>	Caller ID: <u>146944</u>
<u>Thriassur</u>	Date of Visit: <u>16/11/24</u>
Ph: <u>9037450541</u>	Asset No: <u>0822779</u>
	EQPT Name: <u>Nebulizer</u>
	Manufacture: <u>S:core</u> Model: <u>Piston Compressor</u>
	S. No: <u>NA</u> Dept: <u>Female ward</u>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Not working Nebulizer, Piston of motor damaged

Action Taken : When checked the machine met found Piston of motor broken. Need Piston type motor for nebulizer for further checking

Completed Date : 16/11/24 Time : 12:30pm Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>SINTO JOY</u>	<u>16/11/24</u>	<u>11:00 Am</u>	<u>12:30pm</u>

Customer Remark Completed Pending

Service Engineer Name : SINTO JOY
Signature: [Signature]
Date: 16/11/24
Contact Number: 7907406353



Customer Name : [Signature]
Signature: [Signature]
Date: 16/11/24
Contact Number:
Designation : Nursing officer
Hospital Seal

SUPERINTENDEN
TALUK HOSPITAL
KUNNAMPULAM

8943679676

NEBULISER

Brought forward from page no. 51 Vol. II

		R	I	B	
1/2/19	B.F. Balance (70080916) C.1/4/10 Rd from store)			1	do
10/3/19	Rd from store (70088072)	1		2	
10/3/19	Issued to FU (0822719)		2	Nil	
26/2/23	Rd from store (Local purchase) book page 166	1		3 (2+1)	
	Amount to US Book Page no. 46		1		do
5/4/2024	Tr. in from LP Account book Page no. 166 (Donatio no bill) Rd date 5/8/23, 19/12/23	2		5	do



