


PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: THQH CHALAKUDY	Name of District: Thrissur
Name of Equipment: Nebulizer	Equipment ID & Barcode: #146364- 0820691
Make: ACCUNEB JL	Model: ACCUNEB
Serial Number: 141201204	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 3-12-2015	Present status of the equipment: Fully damaged
Date of breakdown: 12-11-2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: The machine checked and found motor and piston are defective. Need to replace these spares for checking working condition of equipment.	
Recommendations for repair (required service details): Not Recommending	
Cost of spares (specify parts and cost): NA	


#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 1353/-
---	---------------------


Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): cyrix report attached.


Reasons for recommending the equipment as BER: The machine checked and found motor and piston are defective. The machine installed on 03-12-2015 and aged up to 7 years and 11 months. Quotation not submitted since the spares are not available in the market. So, we are recommended to condemnation.


BIJO T JOY
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:
Motor and piston are damaged.

Recommended for BER (Yes/ No): <i>Yes</i> Jr. Consultant (Biomedical) NHM, Thrissur  25/11/2024 Signature of JC BM (NHM)
Date:	

Date:	Seal	 Signature of Superintendent / Medical Officer (i/c)
-------	------	--



*Not mandatory #Based on the period of life and value as per the BER guidelines. Attach Photograph



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1017602

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility: <u>T.H.OM</u> Address: <u>Chelakkudy</u> <u>Thrisur</u> Ph: <u>7907039716</u>	Call Registration Date : <u>12/11/24</u> Caller ID : <u>146364</u> Date of Visit : <u>13/11/24</u> Asset No : <u>0820691</u> EQPT Name : <u>nebuliser</u> Manufacture: <u>ACU nebul</u> Model : <u>ACUWEB</u> S. No. <u>141201204</u> Dept. <u>Labour room</u>
---	--

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : not switching on motor

Action Taken : Checked the machine and found motor and piston are defective. need to replace these spares for checking working condition of equipment.

Completed Date : 13/11/24 Time : 3:30pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. <u>Motor</u>	1		
2. <u>piston</u>	1		
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Jinko</u>	<u>13/11/24</u>	<u>2:00pm</u>	<u>3:30pm</u>

Customer Remark Completed Pending

Service Engineer Name : Jinko Jay Customer Name : Ashale
 Signature: [Signature] Signature: [Signature]
 Date: 13/11/24 Date: 13/11/24
 Contact Number: 7593847125 Contact Number: 7907039716
 Designation : Nursing Officer Jr 2
 Hospital Seal



