

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: THQH OTTAPALAM	Name of District: PALAKKAD
Name of Equipment: DIALYSIS MACHINE	Equipment ID & Barcode: # 15015 - 67093022
Make: BBRAUN	Model: DIALOG PLUS
Serial Number: 106337	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: Ticket master Installation date=22/06/2012 (as per stock book equipment installed on 21/07/2013, stock book attached for reference)	Present status of the equipment: Fully damage
Date of breakdown: 15/06/2022 Date of registration of complaint through email/ Toll free): Toll Free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC

Action taken: Checked the Dialysis machine found low temperature and conductivity problem. Problem with Membrane for balance chamber, Analog board, Pressure monitoring module, Heater element, Conductivity cell, SAD version 4 with red detecto damaged. Enquire OEM for quotation.


Recommendations for repair (required service details) No recommendations

Cost of spares (specify parts and cost):
 Membrane for balance chamber = 9323/-
 Compact flash card 1GB = 12986/-
 Analog board = 60230/-
 Pupm body for piston pump = 40191/-
 Heater element = 36573/-
 Bicarb. Conductivity sensor v2 = 23218/-
 Pressure monitoring module = 50463/-
 End Conductivity Sensor v2 = 39131/-
 SAD version 4 with red detecto = 65994/-
 GST 18% = 60,870.42/-
TOTAL = 3,99,039.42/-/-

# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 78.7%	Asset Value:506667/-
---	----------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)
 Service report and Quotation attached

Reasons for recommending the equipment as BER: Checked the Dialysis machine found low temperature and conductivity problem. Problem with Membrane for balance chamber, Analog board, Digital board, Pressure monitoring module, Heater element, Conductivity cell, SAD version 4 with red detecto damaged. Enquire OEM for quotation. Ticket master installation date is 22/06/2012, As per stock book equipment installed on 21/07/2013 and aged up to 8 years 11 months. Repairing cost is 78.7% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending for condemnation.

AKHILESH P 
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:
 As per stock book equipment installed on 21/07/2013 and aged up to 11 years. Repairing cost is 78.7% of asset value. Hence recommended for BER.

Recommended for BER (Yes/ No): Yes.

Date: 21/11/24.

Date: _____ **Seal**  **Superintendent / Medical Officer (i/c)**  6/11/24

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph
Junior Consultant
 ജൂനിയർ കൺസൾട്ടന്റ്
 താലൂക്ക് ഹെഡ്ക്വാർട്ടേഴ്സ് ഹോസ്പിറ്റൽ
 മലപ്പുറം



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1021821

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility: <u>THQH</u>	Call Registration Date: <u>15/06/2022</u>
Address: <u>Ottapalam</u>	Caller ID: <u>15015</u>
<u>Palakkad</u>	Date of Visit: <u>20/09/2024</u>
Ph: <u>9061096021</u>	Asset No: <u>67093022</u>
	EQPT Name: <u>Dialysis machine</u>
	Manufacture: <u>B-BRAUN</u> Model: <u>Dialog +</u>
	S. No: <u>106337</u> Dept: <u>Dialysis</u>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Not working

Action Taken : checked the machine, found not working. DEM attended and Membrane for balance chamber, Compact flash and I.G.B, Analog board, Digital board, Pressure monitoring module, Heater element, Bicarbonate conductivity sensor, End conductivity sensor, S.A.D version all are defective, need to replace for further checking

Completed Date : 20/09/2024 Time : 10:00 am Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Shivakrishnan</u>	<u>20/09/24</u>	<u>10:00 am</u>	<u>11:00 am</u>

Customer Remark Pending



Service Engineer Name : Shivakrishnan

Customer Name : Manjusha M

Signature: [Signature]

Signature: [Signature]

Date: 20/09/2024

Date: 20/09/2024

Contact Number: 6238755132

Contact Number: 81 [unclear]
Designation: Staff [unclear]
Hospital Seal: [unclear]

To,
Cyrix Healthcare
C/O Ottapalam

Sub: - Quotation for Spare Parts**Our Ref.BB/BA-W-296/24 Date: 13.09.2024**

Dear Sir/ Madam,

Greetings from BBraun Medical India Pvt. Ltd.

Kindly find our most competitive rate for Spare part for Hemodialysis machine.

Art. No.	Description	Qty	Unit price (Excl. Tax)	Total Price (Excl. Tax)
3451615A	Membrane for balance chamber	1	9,323	9,323
3456088A	Compact flash card 1 GB	1	12,986	12,986
34560734	Analog board	1	60,230	60,230
34562303	Digital board, version 6	1	82,308	82,308
34561641	Pressure monitoring module	1	50,463	50,463
3456020B	Heater Element 240v/900w	1	36,573	36,573
3456102A	Bicarbonate conductivity sensor v2	1	23,218	23,218
3456103A	End Conductivity sensor v2	1	39,131	39,131
3456005E	SAD version 4 with red detecto	1	65,994	65,994
			Total:	3,80,226
			GST: 18%	68,440.68
			Total Amount:	4,48,666.68


Terms & Conditions:

1. Taxes : As per GST Rate w.r.t. HSN code of spare part.
2. Payment : 100% Advance Cheque/DD in Favour of M/s B.Braun Medical (I)
Pvt. Ltd.
3. Delivery : Within 2-3 weeks from PO.
4. Quotation : Validity 30 days

Thanking you,

Yours faithfully,

For B|Braun Medical (India) Pvt Ltd.




SUSMEETA PANDEY

Senior Executive - Technical Service

E Mail ~ susmeeta.mazumdar@bbraun.com

Phone No: 022 66682205



BABU MURUGESAN

Manager - Technical Service

E Mail ~ babu.murugesan@bbraun.com

Mob No: 9884450417



Purchase Order -KL 24 25 101570 -B. Braun Medical (India) Pvt. Ltd - Mumbai -67093022 -THQH OTTAPALAM

15 messages

Anjitha T L <purchaseryx2@gmail.com>

Wed, 7 Aug, 2024 at 5:10 pm

To: Anoop Karthikeyan <anoop.karthikeyan@bbraun.com>, Babu Murugesan <babu.murugesan@bbraun.com>

Cc: Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Izrath K K <purchaseryx5@gmail.com>, Akhilesh P <akhileshcyrix@gmail.com>

Dear Sir,

Please find the attached Purchase Order and do the needful. Kindly send a confirmation mail with details of material delivery (Docket details) along with the invoice Copy mentioning our PO No. (For reference) with Seal and Signature.

EDD as per payment term: 14/08/24

For Queries and Clarifications

1. Technical: dm1.klbemp@cyrix.in/dm2.klbemp@cyrix.in
2. Accounts: paymentcyrix@gmail.com
3. Logistics: logistics@cyrix.in
4. PO Clarification: ampurchase@cyrix.in/purchasemgr@cyrix.in
5. Delivery address: **Cyrix Healthcare PVT LTD**

**Methanam Road
Door No: 312/a Eloor North,
Kerala – 683501- India
Nikesh V
Mob: 9072522014**

@ Cyrix Accounts

Payment Term: 100% ADV

Thanks & Regards

Anjitha T L
Sr. Purchase Executive
PH:- 9947719111
purchsecyrix2@gmail.com
www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd
| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India
KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

Wed, 7 Aug, 2024 at 6:22 pm

Babu Murugesan <babu.murugesan@bbraun.com>
To: Anjitha T L <purchsecyrix2@gmail.com>, ajith@cyrix.in <ajith@cyrix.in>
Cc: Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Izrath K K <purchsecyrix5@gmail.com>, Akhilesh P <akhileshcyrix@gmail.com>, Anoop <anoop.karthikeyan@bbraun.com>

Dear Anjitha,

After attending every service call, we are sending quote for the required spare parts.

But till date we didn't get any Work Order from you for the quoted spare parts.

Please issue PO for the required spare parts to proceed further.

Thanks & Regards
Babu Murugesan

From: Anjitha T L <purchsecyrix2@gmail.com>
Sent: Wednesday, August 7, 2024 5:10 PM
To: Anoop <anoop.karthikeyan@bbraun.com>; Babu Murugesan <babu.murugesan@bbraun.com>
Cc: Cyrix Purchase Mgr. <purchasemgr@cyrix.in>; Project Manager KLBEMP <pm.bempkl@cyrix.in>; Assistant Manager Purchase <ampurchase@cyrix.in>; Fathima Service Coordinator <bempservicekl2@gmail.com>; Anagha <bempservicekl3@gmail.com>; Cyrix Payments <paymentcyrix@gmail.com>; Blesson Jose <zm2.klbemp@cyrix.in>; DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>; Cyrix Logistics <logistics@cyrix.in>; Kerala GRN <keralagr@cyrix.in>; Izrath K K <purchsecyrix5@gmail.com>; Akhilesh P <akhileshcyrix@gmail.com>
Subject: Purchase Order -KL 24 25 101570 -B. Braun Medical (India) Pvt. Ltd - Mumbai -67093022 -THQH OTTAPALAM

CAUTION: This email originated from outside the organization. Do not click links or open attachments unless you recognize the sender and know the content is safe.

[Quoted text hidden]

Babu Murugesan

Manager - Technical Services
+919884450417
babu.murugesan@bbraun.com

B. Braun Medical (India) Pvt. Ltd. (CIN : U33112MH1984PTC214514)
WorkeZ - Block-B, Door No.3, 2nd Floor, Hansa Building , RK Swamy Centre, 147, Pathari Rd, Thousand Lights, Chennai, 600006
Tamil Nadu, IN, India
Chennai, 600006
,India

<http://www.bbraun.co.in>
www.linkedin.com/company/bbraun-group

As a true partner
we develop smart solutions
and set standards
to improve health care.

B. Braun – Sharing Expertise

The information contained in this communication is confidential, may be attorney-client privileged, may constitute inside information, and is intended only for the use of the addressee. It is the property of the company of the sender of this e-mail. Unauthorized use, disclosure, or copying of this communication or any part thereof is strictly prohibited and may be unlawful. If you have received this communication in error, please notify us immediately by return e-mail and destroy this communication and all copies thereof, including all attachments.

Thu, 8 Aug, 2024 at 5:53 pm

Cyrix Payments <paymentcyrix@gmail.com>

To: Anjitha T L <purchasocyrix2@gmail.com>, Anoop Karthikeyan <anoop.karthikeyan@bbraun.com>

Cc: Babu Murugesan <babu.murugesan@bbraun.com>, Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Izrath K K <purchasocyrix5@gmail.com>, Akhilesh P <akhileshcyrix@gmail.com>

Dear Sir,

Please note that we had transferred Rs.37014 against Invoice

Cust./Vendor Ref. No.	Total Payment
9260162996	INR 9,440.00
9260163249	INR 27,574.00

07/08/2024	37,014.00	NEFT	B. Braun Medica	FBBT242217797368
------------	-----------	------	-----------------	------------------

[Quoted text hidden]

Thanks & Regards,

Tincy T Jose

Accounts Department.

Cyrix Healthcare Pvt Ltd

Office : 30/641 B, First Floor, Petta Junction, Poonithura, Cochin, Kerala - 682 038

Website:- www.cyrixhealthcare.com

Mob : 7593847123

Tue, 13 Aug, 2024 at 10:01 am

Cyrix Payments <paymentcyrix@gmail.com>

To: Anjitha T L <purchasocyrix2@gmail.com>, Anoop Karthikeyan <anoop.karthikeyan@bbraun.com>

Cc: Babu Murugesan <babu.murugesan@bbraun.com>, Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Izrath K K <purchasocyrix5@gmail.com>, Akhilesh P <akhileshcyrix@gmail.com>

Dear Sir/Mam,

Please note that we had transferred Rs.13920 /- as advance against PO101570 after deducting TDS Rs.240/-

Medical (India) Pvt. Ltd - Mumbai					
---	--	--	--	--	--

On Wed, Aug 7, 2024 at 5:10 PM Anjitha T L <purchasecyrix2@gmail.com> wrote:
[Quoted text hidden]

[Quoted text hidden]

Babu Murugesan <babu.murugesan@bbraun.com>

Tue, 13 Aug, 2024 at 2:26 pm

To: Purchase Cyrix <purchasecyrix2@gmail.com>, ajith@cyrix.in <ajith@cyrix.in>
Cc: Anoop <anoop.karthikeyan@bbraun.com>, purchasemgr@cyrix.in <purchasemgr@cyrix.in>, pm.bempkl@cyrix.in <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, BEMP Service 2 <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagrn@cyrix.in>, Izrath K K <purchasecyrix5@gmail.com>, Akhilesh P <akhileshcyrix@gmail.com>

Reminder 1

Thanks & Regards
Babu Murugesan

[Quoted text hidden]
[Quoted text hidden]

Babu Murugesan <babu.murugesan@bbraun.com>

Fri, 23 Aug, 2024 at 4:46 pm

To: Purchase Cyrix <purchasecyrix2@gmail.com>, ajith@cyrix.in <ajith@cyrix.in>
Cc: Anoop <anoop.karthikeyan@bbraun.com>, purchasemgr@cyrix.in <purchasemgr@cyrix.in>, pm.bempkl@cyrix.in <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, BEMP Service 2 <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagrn@cyrix.in>, Izrath K K <purchasecyrix5@gmail.com>, Akhilesh P <akhileshcyrix@gmail.com>

Reminder 2

Please issue PO for all the required spare parts to proceed further.

Thanks & Regards
Babu Murugesan

[Quoted text hidden]
[Quoted text hidden]

Project Manager KLBEMP <pm.bempkl@cyrix.in>

Thu, 29 Aug, 2024 at 4:02 pm

To: Babu Murugesan <babu.murugesan@bbraun.com>, Purchase Cyrix <purchasecyrix2@gmail.com>, Cyrix Ajith <ajith@cyrix.in>
Cc: Anoop <anoop.karthikeyan@bbraun.com>, Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, bempservicekl2 <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagrn@cyrix.in>, Izrath K K <purchasecyrix5@gmail.com>, Akhilesh P <akhileshcyrix@gmail.com>, Afsal Yunuze <zm1.klbemp@cyrix.in>

Dear Babu sir,

Hope you are doing well.

Sorry for the delay in getting back to you.

We already discussed with Mr. Anoop regarding the future services with BBRAUN. Following are the agreed terms;

1. Cyrix will initiate all services with BBRAUN only including one time service and spare parts.
2. Cyrix will process an essential stock order against mail received from BBRAUN for essential stock. As your terms for payment is 100%, we are awaiting our payment from KMSCL for further process. we expect to execute essential stock order on or before 10th of sept 2024.

Currently we are having one time service pending from BBRAUN at Pathanamthitta, Kottayam, Trivandrum, Thrissur, Ottapalam.

Requesting to kindly expedite this orders on high priority.

Thanks & Regards

Sarath K B

Project Manager

Bemp-Kerala

Cyrix Healthcare Pvt. Ltd.

30/641 B, Petta Junction | Poonithura

Kochi | Kerala-682038

KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH || MAHARASHTRA || RAJASTHAN || KUWAIT || UAE

Mob.:- 9744555622,9605603452

Email:- pm.bempkl@cyrix.in | Website:- www.cyrixhealthcare.com

"Genius in Healthcare"

From: Babu Murugesan <babu.murugesan@bbraun.com>

Sent: Friday, August 23, 2024 4:45 PM

To: Purchase Cyrix <purchasecyrix2@gmail.com>; Cyrix Ajith <ajith@cyrix.in>

Cc: Anoop <anoop.karthikeyan@bbraun.com>; Cyrix Purchase Mgr. <purchasemgr@cyrix.in>; Project Manager KLBEMP <pm.bempkl@cyrix.in>; Assistant Manager Purchase <ampurchase@cyrix.in>; bempservicekl2 <bempservicekl2@gmail.com>; Anagha <bempservicekl3@gmail.com>; Cyrix Payments <paymentcyrix@gmail.com>;

Blesson Jose <zm2.klbemp@cyrix.in>; DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>; Cyrix Logistics <logistics@cyrix.in>; Kerala GRN <keralagr@cyrix.in>; Izrath K K <purchasecyrix5@gmail.com>; Akhilesh P <akhileshcyrix@gmail.com>

[Quoted text hidden]

[Quoted text hidden]

[Quoted text hidden]

Babu Murugesan <babu.murugesan@bbraun.com>

Mon, 2 Sept, 2024 at 2:15 pm

To: Project Manager KLBEMP <pm.bempkl@cyrix.in>, Purchase Cyrix <purchasecyrix2@gmail.com>, Cyrix Ajith <ajith@cyrix.in>

Cc: Anoop <anoop.karthikeyan@bbraun.com>, Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, bempservicekl2 <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Izrath K K <purchasecyrix5@gmail.com>, Akhilesh P <akhileshcyrix@gmail.com>, Afsal Yunuze <zm1.klbemp@cyrix.in>

Dear Sarath,

Greetings from B. Braun!!!

Based on your mail we will process all your current service request. I have already instructed our service colleagues to visit, check and give reports.

Kindly make sure that you are keeping all the essential spares to keep the machines up and working.

Thanks & Regards
Babu Murugesan

[Quoted text hidden]
[Quoted text hidden]

Project Manager KLBEMP <pm.bempkl@cyrix.in> Tue, 3 Sept, 2024 at 6:59 am
To: Babu Murugesan <babu.murugesan@bbraun.com>, Purchase Cyrix <purchasecyrix2@gmail.com>, Cyrix Ajith <ajith@cyrix.in>
Cc: Anoop <anoop.karthikeyan@bbraun.com>, Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, bempservicekl2 <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Izrath K K <purchasecyrix5@gmail.com>, Akhilesh P <akhileshcyrix@gmail.com>, Afsal Yunuze <zm1.klbemp@cyrix.in>

Thank you, sir.
We will release the essential spare order asap.

Thanks & Regards

Sarath K B
Project Manager
Bemp-Kerala
Cyrix Healthcare Pvt. Ltd.

30/641 B, Petta Junction | Poonithura
Kochi | Kerala-682038

KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH || MAHARASHTRA || RAJASTHAN || KUWAIT || UAE

Mob.:- 9744555622,9605603452

Email:- pm.bempkl@cyrix.in | Website:- www.cyrixhealthcare.com

"Genius in Healthcare"

From: Babu Murugesan <babu.murugesan@bbraun.com>

Sent: Monday, September 2, 2024 2:14 PM

To: Project Manager KLBEMP <pm.bempkl@cyrix.in>; Purchase Cyrix <purchasecyrix2@gmail.com>; Cyrix Ajith <ajith@cyrix.in>

Cc: Anoop <anoop.karthikeyan@bbraun.com>; Cyrix Purchase Mgr. <purchasemgr@cyrix.in>; Assistant Manager Purchase <ampurchase@cyrix.in>; bempservicekl2 <bempservicekl2@gmail.com>; Anagha <bempservicekl3@gmail.com>; Cyrix Payments <paymentcyrix@gmail.com>; Blesson Jose <zm2.klbemp@cyrix.in>; DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>; Cyrix Logistics <logistics@cyrix.in>; Kerala GRN <keralagr@cyrix.in>; Izrath K K <purchasecyrix5@gmail.com>; Akhilesh P <akhileshcyrix@gmail.com>; Afsal Yunuze <zm1.klbemp@cyrix.in>

[Quoted text hidden]

[Quoted text hidden]

Babu Murugesan <babu.murugesan@bbraun.com>
To: Anjitha T L <purchasecyrix2@gmail.com>, Anoop <anoop.karthikeyan@bbraun.com>

Fri, 13 Sept, 2024 at 6:10 pm

Cc: Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Izrath K K <purchasecyrix5@gmail.com>, Akhilesh P <akhileshcyrix@gmail.com>

Dear Anjitha,

Please find Service report & Quote for the required parts in the attached file.

Thanks & Regards
Babu Murugesan

From: Anjitha T L <purchasecyrix2@gmail.com>

Sent: Wednesday, August 7, 2024 5:10 PM

To: Anoop <anoop.karthikeyan@bbraun.com>; Babu Murugesan <babu.murugesan@bbraun.com>

Cc: Cyrix Purchase Mgr. <purchasemgr@cyrix.in>; Project Manager KLBEMP <pm.bempkl@cyrix.in>; Assistant Manager Purchase <ampurchase@cyrix.in>; Fathima Service Coordinator <bempservicekl2@gmail.com>; Anagha <bempservicekl3@gmail.com>; Cyrix Payments <paymentcyrix@gmail.com>; Blesson Jose <zm2.klbemp@cyrix.in>; DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>; Cyrix Logistics <logistics@cyrix.in>; Kerala GRN <keralagr@cyrix.in>; Izrath K K <purchasecyrix5@gmail.com>; Akhilesh P <akhileshcyrix@gmail.com>

Subject: Purchase Order -KL 24 25 101570 -B. Braun Medical (India) Pvt. Ltd - Mumbai -67093022 -THQH OTTAPALAM

CAUTION: This email originated from outside the organization. Do not click links or open attachments unless you recognize the sender and know the content is safe.

Dear Sir,

[Quoted text hidden]

[Quoted text hidden]

Babu Murugesan <babu.murugesan@bbraun.com>

Mon, 16 Sept, 2024 at 5:59 pm

To: Project Manager KLBEMP <pm.bempkl@cyrix.in>, Purchase Cyrix <purchasecyrix2@gmail.com>, Cyrix Ajith <ajith@cyrix.in>

Cc: Anoop <anoop.karthikeyan@bbraun.com>, Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, bempservicekl2 <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Izrath K K <purchasecyrix5@gmail.com>, Akhilesh P <akhileshcyrix@gmail.com>, Afsal Yunuze <zm1.klbemp@cyrix.in>

Dear Sarath,

As agreed, you supposed to release PO for essential spares.

Kindly update the status.

[Quoted text hidden]

[Quoted text hidden]

Project Manager KLBEMP <pm.bempkl@cyrix.in>

Thu, 19 Sept, 2024 at 11:52 am

To: Babu Murugesan <babu.murugesan@bbraun.com>, Purchase Cyrix <purchasecyrix2@gmail.com>, Cyrix Ajith <ajith@cyrix.in>

Cc: Anoop <anoop.karthikeyan@bbraun.com>, Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, bempservicekl2 <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Izrath K K <purchasecyrix5@gmail.com>, Akhilesh P <akhileshcyrix@gmail.com>, Afsal Yunuze <zm1.klbemp@cyrix.in>

Dear Babu sir,

Sorry for the delay.

We are working on it and will release order in 2 days.
Due to festive season holidays, process got delayed.

Thanks & Regards

Sarath K B

Project Manager

Bemp-Kerala

Cyrix Healthcare Pvt. Ltd.

30/641 B, Petta Junction | Poonithura

Kochi | Kerala-682038

**KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH || MAHARASHTRA || RAJASTHAN || KUWAIT ||
UAE**

Mob.:- 9744555622,9605603452

Email:- pm.bempkl@cyrix.in | Website:- www.cyrixhealthcare.com

"Genius in Healthcare"

From: Babu Murugesan <babu.murugesan@bbraun.com>

Sent: Monday, September 16, 2024 5:58:53 PM

[Quoted text hidden]

[Quoted text hidden]

Babu Murugesan <babu.murugesan@bbraun.com>

Fri, 20 Sept, 2024 at 1:07 pm

To: Anjitha T L <purchasecyrix2@gmail.com>, Anoop <anoop.karthikeyan@bbraun.com>

Cc: Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Project Manager KLBEMP <pm.bempkl@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Fathima Service Coordinator <bempservicekl2@gmail.com>, Anagha <bempservicekl3@gmail.com>, Cyrix Payments <paymentcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Kerala GRN <keralagr@cyrix.in>, Izrath K K <purchasecyrix5@gmail.com>, Akhilesh P <akhileshcyrix@gmail.com>

Dear Anjitha,

Please find invoice in the attached file.

Thanks & Regards

Babu Murugesan

[Quoted text hidden]

[Quoted text hidden]

Izrath K K <purchasecyrix5@gmail.com>

Mon, 23 Sept, 2024 at 2:35 pm

To: Kerala GRN <keralagr@cyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Logistics Cyrix <logisticscyrix@gmail.com>, Akhilesh P

<akhileshcyrix@gmail.com>

Cc: Anjitha T L <purchasecyrix2@gmail.com>, Cyrix Purchase Mgr. <purchasemgr@cyrix.in>

Hi,

Kindly take GRN

po:101570

@Akhilesh P kindly give confirmation

Thanks & Regards
Izrath KK
Jr Purchase Executive
PH: 8848002238

On Wed, Aug 7, 2024 at 5:10 PM Anjitha T L <purchsecyrix2@gmail.com> wrote:
[Quoted text hidden]

Akhilesh P <akhileshcyrix@gmail.com> Mon, 23 Sept, 2024 at 4:32 pm
To: Izrath K K <purchsecyrix5@gmail.com>
Cc: Kerala GRN <keralagrncyrix.in>, Cyrix Logistics <logistics@cyrix.in>, Logistics Cyrix <logisticscyrix@gmail.com>, Anjitha T L <purchsecyrix2@gmail.com>, Cyrix Purchase Mgr. <purchasemgr@cyrix.in>

Dear mam
PO 101570 one time visit completed.

Thanks & Regards


AKHILESH P
DISTRICT INCHARGE - PALAKKAD
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com
[Quoted text hidden]

HAEMO DIALYSIS MACHINE.

11

	Tr in from Old Reg P. No.	RD	Used	Balance	m. Stock
2/7/2013	RD from Store (B Braun)	5		5	
	(man stock - B6-P.N.O. 48) @ 28,19,250/-				
4/11/2013	RD from Store (B. Braun)	2		7	7
	(Install dt 15/7/2013)				
5/11/2013	Used to unit		7	nil	7
3/8/2017	RD from Store	2		2	9
8/9/2017	Used to unit		2	nil	9
	(man stock - B6-10) dt 3/18/2017				
	Nipro 2 machine 4,74,600 x 2				
	Install dt on 8/9/17.				
7/5/2018	RD from Store	10		10	19
	(m. Stock B6-10) @ 4,04,9920 +				
	10,1248 (2 machine) 8 machine				
	NIPRO model 55Plus				
	SL NO: 186 x 9247, 9867, 176 x 9045				
	176 x 9041, 8947, 9043, 8957,				
	9037, 9136, 9039.				
	Warranty from 6-4-2018 to 5/4/2021				
8/5/2018	Used to unit		10	nil	19
30/5/2019	RD from Str. (Freshwag)	5		5	24
	Warranty - 2019-2020.				
	(entered Old Stock Register P No. 30)				
30/5/19	Used to unit.		5	nil	
30/12/2020	Rel from Str	5			29
	(Cost @ 563,300 x 5)				
	Total Bill 4,50,000				
5/1/2021	Used to unit		5	nil	29
21/7/22	Tr out to Unserviceable Register				
	B. No. 7, Page No. 48, B. Braun (2) machines				
	(67093026, 67093022) RD on 4/11/13				
	Inst. dt 15/7/13, 0921442				
	B. Braun - 106699				
	" - 106339				



Superintendent
Tata Headmaster Hospital
Ottobram

3+2
(5)

24

24

Dialog +

Aparelho para Hemodiálise

REF 710200N

SN 106337

Bj 2012

VA 2500

V 230~

Hz 50/60

IP 21 CE 0123



Manufacturer/Fabricado por:

B.Braun Avitum AG

34209 Melsungen, Germany

Made in Germany

Importado e Distribuído por: B. Braun
Laboratórios B. Braun S.A.
Brasil Registro ANVISA n°:8.01369.90498

Not to be used in presence of
flammable anaesthetic gases





**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1021816

CYRIX[®]
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility: THQH	Call Registration Date: 15/06/2022
Address: Ottapalam	Caller ID: 15015
Palakkad	Date of Visit: 10/09/2024
Ph: 9061096021	Asset No: 67093022
	EQPT Name: Dialysis Machine
	Manufacture: B. BRAUN Model: Dialog +
	S.No: 106337 Dept: Dialysis

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Not working, part missing case.

Action Taken : checked the dialysis machine, found part missing case, now we replaced Mother board, Acid pump with motor, UF pump with motor, bicarb pump with motor, power board for valve, Digital board, Flow pump input and output, Degassie pump PE sensor, PBA sensor, SMPS, But still complaint exist, Required DEM Support

Completed Date: 10/09/2024 Time: 11: AM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer: Shivakrishnan.K

Customer Remark: Machine not working, replaced components, machine not working, technician replace 10/09/2024



Junior Consultant
Start Time: 11:30 AM
Date: 10/09/2024

Service Engineer Name : Shivakrishnan.K
Signature: [Signature]
Date: 10/09/24
Contact Number: 6238755132

Customer Name : Manjusha.M
Signature: [Signature]
Date: 10/09/2024
Contact Number: 8157943029
Designation : Staff Nurse.
Hospital Seal

[Signature]
K. Manjusha

EQUIPMENT STATUS VERIFICATION REPORT		
Hospital Name : <i>TIRU Ottapalam</i>		District : <i>palakkad</i>
Equipment Name : <i>Dialysis Machine</i>	Barcode : <i>67093022</i>	Department : <i>Dialysis</i>
Warranty Status: <i>Yes / No</i>	Name of Supplier : <i>B. BRAUN</i>	
Installation date:		
S/N : <i>106337</i>	Model : <i>Dialog +</i>	Manufacturer : <i>B BRAUN</i>
Breakdown reported date : <i>15-06-2022</i>	Complaint id : <i>15015</i>	
Present working status of machine: Not working <input type="checkbox"/> Partially working <input checked="" type="checkbox"/>		
Nature of complaint :		
<i>part missing</i>		
Remarks of CYRIX executives :		
<i>checked and found that following are missing Mother board Acid pump with motor, UF pump with motor, bicarb pump with motor, power board for valve, digital board, flow pump in and out, Degasis pump, PE sensor, PDA sensor</i>		
Remarks of Junior Consulting Biomedical Engineer (NHM) :		

Remarks of Hospital Authorities / User :

Any other details:


Damaged part : Electric Board Cables Tubings Probes Display
Transducer Others part missing.....(specify)

Physical damage Rodent attack Natural calamities

Hereby we certify that the repair is due to the above mentioned factor
Signature with date and seal

[Handwritten Signature]
Signature of CYRIX
Representative

Signature of JCBME's


[Handwritten Signature]
Signature of Superintendent
Quarters Hospital
Tamil Nadu
Medical Officer (i/c)



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 23747

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : THQH	Call Registration Date : 15-06-2022
Address : Ottapalam palakkad.	Caller ID : 15015
Ph : 9061096621	Date of Visit : 15-06-2022
	Asset No. : 67093022
	EQPT Name : Dialysis Machine
	Manufacture : B. BRAUN Model : Dialyxt
	S. No. : Dept. : Dialysis

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified :
Not working

Action Taken : Checked and found that following parts are missing
Motor board, Acid pump with motor, UF pump with motor, bicarb pump
with motor, power board for valve, Digital board, Flow pump is out of
Degasin pump, PE sensor, PPA sensor, S.M.P.S.

Completed Date : 15-6-2022 Time : Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Midhun rajk			

Customer Remark Completed Pending

Service Engineer Name : Midhun rajk
Signature : *[Signature]*
Date : 15-06-2022
Contact Number : 9072522081



Name : REESHNA SOM
Designation : NURSING OFFICER
Contact Number : 9061096621
Hospital Seal :

[Signature]
Superintendent
Petta Head Quarters Hospital
Ottapalam



ESV SERVICE REPORT - PART MISSING 67093022.

1 message

Blessen Jose <bleseencyrix@gmail.com>
To: Biomedical Engineer Palakkad <bmeprd2020@gmail.com>
Cc: Project Manager KLBEMP <pm.bempkl@cyrix.in>

Wed, 15 Jun 2022 at 4:06 pm

Dear mam,

Please find the attached ESV REPORT, SERVICE REPORT AND IMAGES, for the complaint reported from THQH OTTAPALAM, For the complaint reported against Dialysis machiene - 67093022.

Thanks & Regards

Blessen Jose
Zonal Manager I North Zone
Bemp-Kerala
Mobile : 7593847153
Email : bleseencyrix@gmail.com

CYRIX HEATHCARE (P)LTD
30/641B, Petta, Punithura, Ernakulam- 682038
www.cyrixhealthcare.com