




**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: THQH CHITTUR	Name of District: PALAKKAD
Name of Equipment: NEBULIZER	Equipment ID & Barcode: #146688 - 0920019
Make: MEDIGOLD	Model: NA
Serial Number: NA	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: In ticket master Installation date 24/01/2022, As per stock 26/11/2018 stock book attached for reference.	Present status of the equipment: Fully damaged
Date of breakdown: 14/11/2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked the nebulizer found motor stuck. Problem with motor and piston, need to be replace these spares for further checking of machine.	
Recommendations for repair (required service details): No recommendations	
Cost of spares (specify parts and cost): NA	
#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 1353/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix Service report attached.	
Reasons for recommending the equipment as BER: Checked the nebulizer found motor stuck. Problem with motor and piston, need to be replace these spares for further checking of machine. In ticket master installation date is 24/01/2022, As per the stock book Equipment installed on 26/11/2018 aged up to 6 years .Quotation not submitted since spare are not available from market. So we are recommended the equipment for condemnation	
AKHILESH P  Name & Signature of CYRIX Authority with date	
Remarks of Junior Consultant (Biomedical) NHM: <i>As per the stock book equipment installed on 26/11/2018 and aged up to 6 years. quotation not submitted.</i>	
Recommended for BER (Yes/ No): <i>Yes.</i>	 Consultant (Biomedical Engg) NHM (AGM, Palakkad)
Date:	
Date:	 Signature of Superintendent / Medical Officer (i/c)



Attached photograph
SUPERINTENDENT
 Taluk Head Quarters Hospital
 Chittur-678 101, Palakkad

*Not mandatory

##Based on the period of life and value of the BER guidelines

Nebulizer

142.

Date	Rd From where	Rd	Iss	Bal	MSA	S
	BF from VOLA Pg 57			5	5	0
3/02/2018	Rd from Store (Nebulizer LP) Hmc Pg 96	02	2	7	7	0
	ISSUED for use					
26/11/18	Rd from Store Rs 2000/-	2	2	9	9	
	ISSUED for use					
15/3/19	Rd from Store	01	1	10	10	
	ISSUED for use					
15/6/2021	Condensation done as per order No. A72398/2020 dated 15/8/2021 of Supt THQH Chittur		05		05	0
24/1/2022	Rd from Store	03	3	08	08	0
	Rd for ISSUED for use		3	5	5	
	Rd from Store	8		16	16	0
	Post covid fund NAC No KL01074/ NAC/00161/23-24 661312v 30000/-					
8/4/24	Te out to PW (Pg No 30 PW Stock Regt)		1		15	
8/4/24	Te out to CHW (Pg No 251 Stock Regt)		1		14	



A.P. Jagan
24/4/24

SUPERINTENDENT
Taluk Head Quarters Hospital
Chittur-678 101, Palakkad

