

**PROFORMA  
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)  
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: WOMEN AND CHILD HOSPITAL		Name of District: PALAKKAD	
Name of Equipment: OPHTHALMOSCOPE		Equipment ID & Barcode: #142997-0912721	
Make: APPASAMI		Model: NA	
Serial Number: NA		Warranty details: No Warranty	
Date of purchase/Year of manufacture/ Installation Date: In ticket master Installation date - 12/04/2014, As per stock 26/11/2014 stock book attached for reference.		Present status of the equipment: Fully damaged	
Date of breakdown: 26/10/2024 Date of registration of complaint through email/Toll free: Toll free		*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC	
Action taken: Checked the Ophthalmoscope found clarity issue. Problem with power supply assembly, Head band assembly, Prism damaged. need to be replace these spares for further checking of machine.			
Recommendations for repair (required service details): No recommendations			
<b>Cost of spares (specify parts and cost):</b> Power supply assembly = 5000/- Head band assembly = 5000/- Prism indirect ophthalmoscope = 1000/- 12% GST = 1320/- <b>Total = 12320/-</b>			
#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 82.13%		Asset Value: 15000/-	
Abstract of Service Report provided by the OEM/Authorized Service Provider/CYRIX (Attached or Not): Cyrix Service report attached and quotation attached.			
Reasons for recommending the equipment as BER: Checked the Ophthalmoscope found clarity issue. Problem with power supply assembly, Head band assembly, Prism damaged. need to be replace these spares for further checking of machine. In ticket master installation date is 12/04/2014, As per the stock book Equipment installed on 26/11/2014 aged up to 9 years 11 months. Repairing cost is 82.13% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending for condemnation.			
		AKHILESH P Name & Signature of CYRIX Authority with date 13/11/24	
Remarks of Junior Consultant (Biomedical) NHM: Equipment installed on 26/11/2014 aged up to 9 years 11 months. Repairing cost is 82.13% of asset value. Criteria of RBER met.			
Recommended for BER (Yes/ No): <i>yes</i>		Signature of JCBM (NHM) Consultant (Biomedical Engg) NHM (Arogyakeralam), Palakkad 13/11/24	
Date: 13/11/24		Signature of Superintendent/Medical Officer (i/c) Consultant (Biomedical Engg) NHM (Arogyakeralam), Palakkad 19/11/24	
Date:		Seal	

\*Not mandatory

#Based on the period of life and value as per the BER guidelines of Hospital Palakkad \* Attach Photograph







# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

No : 126310

# CYRIX<sup>®</sup>

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>W&amp;C Hospital</u> ..... Address..... <u>Palakkad</u> ..... ..... Ph:..... <u>9562063687</u> .....	Call Registration Date : ..... <u>26.10.2024</u> ..... Caller ID : ..... <u>142997</u> ..... Date of Visit : ..... <u>28.10.2024</u> ..... Asset No : ..... <u>0912721</u> ..... EQPT Name : <u>Ophthalmoscope</u> ..... Manufacture <u>Appasami</u> Model : <u>NA</u> ..... S. No..... <u>NA</u> ..... Dept..... <u>DEIC</u> .....
---------------------------------------------------------------------------------------------------------------------------------------	-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : ..... Not working.....  
 .....  
 .....

Action Taken : ..... checked the equipment found problem with power supply assembly, head band assembly, and prism damaged. Need to be replace these spare for further checking.....  
 .....

Completed  Date : 28.10.2024 Time : 10:00 AM Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Libil Lal</u>	<u>28.10.24</u>	<u>9.00AM</u>	<u>10.00AM</u>

Customer Remark Completed  Pending

Service Engineer Name : Libil Lal  
 Signature: [Signature]  
 Date: 28.10.2024  
 Contact Number: 8891382686

Customer Name :  
 Signature: [Signature]  
 Date: 19/11/2024  
 Contact Number:  
 Designation : **SUPERINTENDENT**  
 Hospital Seal: **Govt. W&C Hospital Palakkad**

## QUOTATION



Appasamy Associates Private Limited  
No 20 SBI Officers Colony  
1st Street Arumbakkam  
Area & City : Arumbakkam, Chennai  
Dist. : Chennai , 600106  
Tamil Nadu, TN, India  
Contact : 9381216555, 9381216555  
GSTIN No.: 33AASCA3660R1Z0

Quotation Number SQN/HO-24-22587 / SPARES	Date 05/11/2024 15:54:42	Representative Name Thrissur-2-Thirumalai Kumar Rep Mobile: 9349211172
----------------------------------------------	-----------------------------	------------------------------------------------------------------------------

<b>Customer:</b> M/S. CYRIX HEALTHCARE PVT LTD No. 30/641B, First Floor, Petta Junction, Area & City : Poonithura, Ernakulam Dist. : Ernakulam , 682038 Kerala, KL, India Contact : , 9744555073 GSTIN No: 32AAFCC2499H2ZM	<b>Ship To</b> M/S. CYRIX HEALTHCARE PVT LTD No. 30/641B, First Floor, Petta Junction, Area & City : Poonithura, Ernakulam Dist. : Ernakulam , 682038 Kerala, KL, India Contact : , 9744555073 GSTIN No: 32AAFCC2499H2ZM
----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------	--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

S.No	Description	Qty	Rate/Unit	Taxes	Total
1	[IOPS-001-S] INDIRECT POWER SUPPLY ASSEMBLY-INDIRECT OPHTHALMOSCOPE	1 Nos	5,000.00	IGST 12%	5,000.00 ₹
2	[IOHB-H-S] HEAD BAND ASSEMBLY (H)-INDIRECT OPHTHALMOSCOPE	1 Nos	5,000.00	IGST 12%	5,000.00 ₹
3	PRISM INDIRECT OPHTHALMOSCOPE	1 Nos	1,000.00	IGST 12%	1,000.00 ₹
<b>Total</b>		<b>3</b>	<b>Subtotal</b>		<b>11,000.00 ₹</b>
			IGST 12% on 11,000.00 ₹		<b>1,320.00 ₹</b>
			<b>Total</b>		<b>12,320.00 ₹</b>
Amount in words: Twelve Thousand, Three Hundred Twenty Rupees					<b>12,320.00 ₹</b>

**Terms & Condition**

For Appasamy Associates Private Limited

(Authorized Signatory)

Head Office: #20, SBI Officers Colony, Arumbakkam, Chennai - 600 106.

Computer generated order and requires no signature



**Re: REQUIRED QUOTATION 0912721 - 142997 FOR INDIRECT OPHTHALMOSCOPE @ WOMEN AND CHILD HOSPITAL PALAKKAD , PALAKKAD**

6 messages

**Akhilesh P** <akhileshcyrix@gmail.com>

Fri, 1 Nov, 2024 at 8:59 am

To: M.Vettri Selvan M.Selvam <sivamselap@gmail.com>, radhakrishnan.thrissur@appasamy.com, info@appasamy.com

Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <Zm2.klbemp@cyrix.in>

Dear sir

Kindly update the status.

Thanks & Regards

AKHILESH P  
DISTRICT INCHARGE - PALAKKAD  
Bemp-Kerala  
Mobile : +91 7593847154  
Email : [akhileshcyrix@gmail.com](mailto:akhileshcyrix@gmail.com)

CYRIX HEALTHCARE (P) LTD  
30/641B, Petta, Punithura  
Ernakulam- 682038  
[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

On Wed, 30 Oct, 2024, 11:37 am Akhilesh P, <[akhileshcyrix@gmail.com](mailto:akhileshcyrix@gmail.com)> wrote:

Dear sir  
Please share the price and stock availability of prism, adapter box , lamp head for ophthalmoscope at women and child hospital palakkad

Details are furnished below

Hospital Name : WOMEN AND CHILD HOSPITAL PALAKKAD

Equipment Name : INDIRECT OPHTHALMOSCOPE

Make : APPASWAMMY

Model : CLASS 1 TYPE B

SN : IOPSH01140447

Thanks & Regards

AKHILESH P  
DISTRICT INCHARGE - PALAKKAD  
Bemp-Kerala  
Mobile : +91 7593847154  
Email : [akhileshcyrix@gmail.com](mailto:akhileshcyrix@gmail.com)

CYRIX HEALTHCARE (P) LTD  
30/641B, Petta, Punithura  
Ernakulam- 682038  
[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

**Akhilesh P** <akhileshcyrix@gmail.com>

Sat, 2 Nov, 2024 at 7:09 am

To: M.Vettri Selvan M.Selvam <sivamselap@gmail.com>, radhakrishnan.thrissur@appasamy.com, info@appasamy.com

Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <Zm2.klbemp@cyrix.in>



Dear sir  
Kindly update the status.

Thanks & Regards

AKHILESH P  
DISTRICT INCHARGE - PALAKKAD  
Bemp-Kerala  
Mobile : +91 7593847154  
Email : [akhileshcyrix@gmail.com](mailto:akhileshcyrix@gmail.com)

CYRIX HEALTHCARE (P) LTD  
30/641B, Petta, Punithura  
Ernakulam- 682038  
[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

[Quoted text hidden]

---

**Akhilesh P** <[akhileshcyrix@gmail.com](mailto:akhileshcyrix@gmail.com)> Mon, 4 Nov, 2024 at 12:35 pm  
To: M.Vettri Selvan M.Selvam <[sivamselap@gmail.com](mailto:sivamselap@gmail.com)>, radhakrishnan.thrissur@appasamy.com, info@appasamy.com  
Cc: DM Sarang-KLBEMP <[dm1.klbemp@cyrix.in](mailto:dm1.klbemp@cyrix.in)>, Blesson Jose <[Zm2.klbemp@cyrix.in](mailto:Zm2.klbemp@cyrix.in)>

Dear sir  
Kindly update the status.

Thanks & Regards

AKHILESH P  
DISTRICT INCHARGE - PALAKKAD  
Bemp-Kerala  
Mobile : +91 7593847154  
Email : [akhileshcyrix@gmail.com](mailto:akhileshcyrix@gmail.com)

CYRIX HEALTHCARE (P) LTD  
30/641B, Petta, Punithura  
Ernakulam- 682038  
[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

[Quoted text hidden]

---

**Akhilesh P** <[akhileshcyrix@gmail.com](mailto:akhileshcyrix@gmail.com)> Mon, 4 Nov, 2024 at 1:44 pm  
To: [info@appasamy.com](mailto:info@appasamy.com), [mariyaraj1220@gmail.com](mailto:mariyaraj1220@gmail.com)  
Cc: DM Sarang-KLBEMP <[dm1.klbemp@cyrix.in](mailto:dm1.klbemp@cyrix.in)>, Blesson Jose <[Zm2.klbemp@cyrix.in](mailto:Zm2.klbemp@cyrix.in)>

Dear sir  
Gentle reminder.

Thanks & Regards

AKHILESH P  
DISTRICT INCHARGE - PALAKKAD  
Bemp-Kerala  
Mobile : +91 7593847154  
Email : [akhileshcyrix@gmail.com](mailto:akhileshcyrix@gmail.com)

CYRIX HEALTHCARE (P) LTD  
30/641B, Petta, Punithura  
Ernakulam- 682038  
[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

[Quoted text hidden]

---

**Akhilesh P** <[akhileshcyrix@gmail.com](mailto:akhileshcyrix@gmail.com)> Tue, 5 Nov, 2024 at 4:42 pm  
To: [sakthivel.sub@appasamy.com](mailto:sakthivel.sub@appasamy.com)

++ looping

Thanks & Regards

AKHILESH P  
DISTRICT INCHARGE - PALAKKAD  
Bemp-Kerala  
Mobile : +91 7593847154  
Email : [akhileshcyrix@gmail.com](mailto:akhileshcyrix@gmail.com)

CYRIX HEALTHCARE (P) LTD  
30/641B, Petta, Punithura  
Ernakulam- 682038  
[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)  
[Quoted text hidden]

Wed, 6 Nov, 2024 at 9:37 am

**Sakthivel** <[sakthivel.sub@appasamy.com](mailto:sakthivel.sub@appasamy.com)>  
To: Akhilesh P <[akhileshcyrix@gmail.com](mailto:akhileshcyrix@gmail.com)>  
Cc: Mariyaraj <[mariyaraj1220@gmail.com](mailto:mariyaraj1220@gmail.com)>, Thirumalaikumar <[thirumalai.kumar@appasamy.com](mailto:thirumalai.kumar@appasamy.com)>

Dear Mr. Akhilesh,

As discussed, please find the attached quotation.  
Thank you.

Regards,  
Sakthivel . S  
Area Manager - Technical Services  
Kerala.

Mob: 7200343793

 **APPASAMY  
ASSOCIATES**  
Empowering Vision\*

---

**From:** Akhilesh P <[akhileshcyrix@gmail.com](mailto:akhileshcyrix@gmail.com)>


**Sent:** Tuesday, November 5, 2024 4:42 PM

**To:** Sakthivel <[sakthivel.sub@appasamy.com](mailto:sakthivel.sub@appasamy.com)>

**Subject:** Fwd: REQUIRED QUOTATION 0912721 - 142997 FOR INDIRECT OPHTHALMOSCOPE @ WOMEN AND CHILD HOSPITAL PALAKKAD , PALAKKAD

[Quoted text hidden]





Product : Indirect Ophthalmoscope  
 Power Supply  
 Model No : **IOPS-H01140447**  
 Input : 220V, 50Hz ~ AC  
 Output : 6V/2A  
 Usage : 1 A Slow Blow  
 Mode Of : Intermediate  
 Operation : (Refer User Manual)  
**CLASS I**  
 'TYPE'B'  
 Manufactured & Marketed By  
**APPASAMY ASSOCIATES**  
 11A, Valluvar Neeruppalai, Rani Anna Nagar  
 Arundakkam, Chennai-600 108




19/11/2024  
 SUPERINTENDENT  
 Govt. W & C Hospital  
 Palakkad



# INDEX

NAMES	Page No.	NAMES	Page No.
① Illuminated vision testing drum	61		
As per the order no DPMSU - PLKD/1102/E1/2011 (DPMSU II) dated on 11.12.11			
Global Traders, 11/350 opp Post office, Kumbakonam, Kollayam - 16, Kerala.			
Illuminated vision testing drum		Invoice No : 632 Rate: 2000/-	
		Received from the store	
		Rachmath CA	
		Deborah L	
		1717114 (Coplane Dist)	
Infanto meter.	80		
② Indirect Ophthalmoscope	109		
As per order no DPMSU - PLKD/809/E1/2013. dated 19/12/14			
Appasamy Associates 20 SBI office Colony, Ambakram, Chennai - 600106.			
Indirect ophthalmoscope, Invoice no: 110/201			
Rate 36,615/- Received from store.			
		<i>(Signature)</i>	
		19/11/2014	
		Kiosk Manager	



  
 19/11/2014  
 SUPERINTENDENT  
 Govt. W&C Hospital  
 Palakkad