



**PROFORMA**


**RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)**

BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: PHC CHELORA	Name of District: KANNUR
Name of Equipment: BP APPARATUS	Equipment ID & Barcode: 139851 #1344753
Make: ELKO	Model: EL510
Serial Number: NA	Warranty details: NO Warranty
Date of purchase/ Year of manufacture/ Installation Date: 24.08.2019	Present status of the equipment: Fully damaged
Date of breakdown: 10.10.2024 Date of registration of complaint through email/ Toll free): Toll free	*AMC/CAMC Period agreed at the time of purchase: NO AMC/ CAMC
Action taken: Checked the Equipment found Equipment is not switching ON, found that Main Board, Pressure sensor and Display Defective. Need to replace these spare for further checking.	
Recommendations for repair (required service details): No recommendations	
Cost of spares (specify parts and cost): N/A	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: N/A	Asset Value: 2200/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Service report attached	
Reasons for recommending the equipment as BER: Checked the Equipment found Equipment is not switching ON, found that Main Board, Pressure sensor and Display Defective. Need to replace these spare for further checking. Equipment installed on 24.08.2019. aged up to 5 Years. Quotation not submitted since spares are not available in the market. So we recommending for condemnation.	
Avinash  Name & Signature of CYRIX Authority with date 17/10/24	
Remarks of Junior Consultant (Biomedical) NHM: checked & identified that machine not working due to main board, pressure sensor and display defective. Aged upto 5 years. quotation not submitted since spare not available, so recommended for BER.	
Recommended for BER (Yes/ No):	 Signature of JC BM (NHM) CONSULTANT BIOMEDICAL NATIONAL HEALTH INSTITUTION KANNUR - 670002
Date: 17/10/2024.	
Date: 17/10/24	Signature of Superintendent / Medical Officer (i/c)

\*Not mandatory

\*\* Based on the period of life and value as per the BER guidelines



MEDICAL OFFICER  
 PRIMARY HEALTH CENTRAL  
 CHELORA, P.O. VARAM  
 KANNUR DISTRICT  
 PIN-670594

\*Attach Photograph







144

## STOCK REGISTER

24.8.19


Q17435

Name of Articles KARUNYA KIT Cost Rate

Date	PARTICULARS	V.No. or Bill No	RECEIPT		ISSUED		BALANCE	
			Quantity		Quantity		Quantity	
24.8.19.	BP Apparatus Digital		1		-		1	
"	Dust bin Black		1		0		1	
"	Dust bin Red		1		0		1	
"	Dust bin Yellow		1		0		1	
"	Examination Tubch.		1		0		1	
"	STETHESCOPE		1		0		1	
"	Foetoscope		1		0		1	
"	Gloves Sterile? (SURGICARE)		25		25		25	Page no. 13
"	Glucometre Strips		300		300		Nil	Page no. 11
"	Glucometre (on calls plus)		1	1	-		1	Page no. 11
"	Haemometre Test Strips (50 test/pack-trials)		1		1		211	Page no. 11
"	Haemometre (True Hb)		1		-		1	Page no. 11
"	Hub Cutter (Manual Syringe and needle) cuffs		1		-		1	N/O
"	Infantometre		1		-		1	
"	KIT FOR TESTING RESIDUAL CHLORINE IN DRINKING WATER		100		100		Nil	Page no. 132
"	LANCET'S		100		100		Nil	Page no. 132
"	MEASURE TARE		1		-		1	
"	NEBULISER		1		-		1	
"	Oxygen Cylinder 100FT B TYPE		1		-		1	(PAC)







OFFICE OF THE  
DIGITAL OFFICER  
CHIEF OF POLICE  
Date: \_\_\_\_\_  
\* PRIMARY HEALTH CARE \*