

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: PHC MOUCODE	Name of District: KASARGOD
Name of Equipment: STERILIZER (In ticket master equipment name is sterilizer actual equipment is AUTOCLAVE)	Equipment ID & Barcode: 143262- 1440420
Make: SREERAJ SURGICALS	Model: N/A
Serial Number: N/A	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 25/11/2017	Present status of the equipment: Fully damaged
Date of breakdown: 28/10/2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked the unit, found coil, gasket, pressure gauge and whistle are damaged. Enquired vendor for quotation.	

Recommendations for repair (required service details): Not recommending for repair

Cost of spares (specify parts and cost):
 GASKET-1450/-
 COIL-2000/-
 PRESSURE GAUGE-1450/-
 AUTOCLAVE WHISTLE-1100/-
 TOTAL- 6000/-

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 78.22%	Asset Value: 49491/- (In ticket master asset value is 49491, As per stock book asset value is 7670/-)
---	---

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service report and OEM quotation attached

Reasons for recommending the equipment as BER: Checked the equipment, found problem with coil, gasket, pressure gauge and whistle. Machine installed on 25-11-2017 and aged up to 6 years and 11 months. Repairing cost is above 78% of asset value (In ticket master asset value is 49491/-, as per stock book asset value is 7670/-). Both criteria for RBER met. Considering the tender clause 5.3.14.1, We recommending the equipment for condemnation.

ABHINAV. O
Abhinav
 Name & Signature of CYRIX Authority with date 09.11.2024


Remarks of Junior Consultant (Biomedical) NHM: checked the machine and found problem with coil, gasket, pressure gauge and whistle. Machine installed on 25/11/2017. In ticket master asset value is 49491/- as per stock book asset value is 7670/-) Repairing cost is above 78% of asset value.

Recommended for BER (Yes/ No): <i>Yes</i>	Jr. CONSULTANT Biomedical Engineering Arogyakeralam (NHM) Kasargod at Kanha Signature of JC BM (NHM)
Date: <i>15/11/2024</i>	

Date: <i>18-11-24</i>	Seal	Signature of Superintendent / Medical Officer (i/c)
-----------------------	------	---

*Not mandatory #Based on the period of life and value as per the BER guidelines * Attach Photograph

Dr. SOORYA RAGHAVAN
 MBBS, MD (Pulmonary Medicine)
 Reg. No: 56919
 Medical Officer
 F H C Moucode





BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1014970

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility... <u>PHC Moucode</u> Address... <u>Moucode</u> <u>Kasaragod.</u> Ph: <u>9747476376</u>	Call Registration Date: <u>28.10.2024</u> Caller ID: <u>143262</u> Date of Visit: <u>30.10.2024</u> Asset No: <u>1440420</u> EQPT Name: <u>Sterilizer (Autoclave)</u> <u>Sreejaj Surgical</u> Manufacture: <u>Sreejaj Surgical</u> Model: <u>NA</u> S. No. <u>NA</u> Dept. <u>OP</u>
--	---

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Not heating

Action Taken : Checked the machine and found that machine heater coil, gasket, pressure gauge and whistle are damaged. Enquired vendor for quotation.


Completed Date : 30.10.2024 Time : 12:30pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer <u>Abhinav D</u>	Date <u>30.10.24</u>	Start Time <u>11:00Am</u>	End Time <u>12:30pm</u>
------------------------------------	-------------------------	------------------------------	----------------------------

Customer Remark



SCORVA RA
MBBS, MD (Pulmonology)
Reg. No: 5199

Service Engineer Name : Abhinav D
 Signature: [Signature]
 Date: 30.10.2024
 Contact Number: 8942392989

Customer Name: Shyama Aa
 Signature: [Signature]
 Date: 30/10/24
 Contact Number: 9747476376
 Designation: Nurses Officer in Charge
 Hospital Seal



Sreeraj Surgicals

GST No: 32ABYFS3674H1ZL

DI No : 13/023/20B/2011

13/024/21B/2011

DI.: 29-01-2011

Aiswarya Commercial Complex, Nr. Ashirvad Hospital, Thavakkara, Kannur - 670 002
Ph: 0497 2706307. Mob: 9349846307, 9387529848
Email: sreerajsurgicals@gmail.com, rajeshsreerajsurgical@gmail.com
www.sreerajsurgical.com  www.facebook/ sreerajsurgicals

Date:

QUOTATION

DATE: 07-11-2024

TO
CYRIX HEALTHCARE (P) LTD

SIR,

THANK YOU FOR YOUR ENQUIRY TO THE RATE OF THE FOLLOWING ITEMS.

SL.NO	ITEMS	AMOUNT
1	AUTOCLAVE GASKET	1450.00
2	AUTOCLAVE COIL	2000.00
3	PRESSURE GAUGE	1450.00
4	AUTOCLAVE WHISTLE	1100.00
TOTAL		6000.00

WE EXPECT YOUR VALUABLE ORDERS AND KIND CO-OPERATION WITH US
GST INCLUDED



Authorised Dealer:



Suppliers of Surgical Instruments, Hospital Furnitures, X-ray Materials, Lab Equipments & Disposables



Jayesh Krishnan <jayeshcyrix@gmail.com>

QUOTATION REQUEST FOR AUTOCLAVE SERVICE 1440420

4 messages

Jayesh Krishnan <jayeshcyrix@gmail.com>

Tue, Nov 5, 2024 at 11:25 AM

To: Sreerajsurgical@gmail.com

Cc: Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear sir,

Autoclave installed in PHC Moucode having problem, which the unit supplied by Sreeraj surgicals..
Complaint: Coil, gasket, whistles and pressure gauge damaged. Autoclave body having crack. Photo of the unit is attached for your reference. Kindly do the needful.

—
Thanks & Regards

JAYESH KRISHNAN,
Sr. Biomedical Engineer,
District In-charge
Kasargod.
PH:- 7593847156

jayeshcyrix@gmail.comwww.cyrixhealthcare.com**Cyrix Healthcare Pvt Ltd****| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India****KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH**autoclave.jpg
624K

Jayesh Krishnan <jayeshcyrix@gmail.com>

Thu, Nov 7, 2024 at 8:44 AM

To: Sreerajsurgical@gmail.com

Cc: Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Sir,

Kindly update.

[Quoted text hidden]

sreerajsurgical <sreerajsurgical@gmail.com>

Thu, Nov 7, 2024 at 10:30 AM


To: Jayesh Krishnan <jayeshcyrix@gmail.com>

Sent from Mail for Windows

19/11/2024, 20:36

Gmail - QUOTATION REQUEST FOR AUTOCLAVE SERVICE 1440420

[Quoted text hidden]

 **img933.pdf**
145K

sreerajsurgicals <sreerajsurgicals@gmail.com>
To: Jayesh Krishnan <jayeshcyrix@gmail.com>


Thu, Nov 7, 2024 at 3:33 PM

Sent from Mail for Windows

From: Jayesh Krishnan
Sent: 07 November 2024 08:44 AM
To: Sreerajsurgicals@gmail.com
Cc: Blesson Jose; DM Sarang-KLBEMP

[Quoted text hidden]


[Quoted text hidden]


 **img936.pdf**
147K



FAMILY HEALTH CENTRE MOUCODE STOCK REGISTER


Name DP AUTOCLAVE.

Date	Received from or Issued to	Invo No. or Issue	Rate		Receipts	Issues	Balance	Remarks
			Rs.	P.				
	<u>B/FN from</u>	<u>stock</u>	<u>Register</u>		<u>page No: 86</u>		<u>Vol I</u>	
<u>25/11/2017</u>	<u>Received from</u>	<u>B/F</u>						
	<u>the store</u>	<u>1501</u>	<u>7670/-</u>		<u>1</u>		<u>1</u>	
	<u>Kanbur sreeraj surgicals</u>		<u>HSN CODE 84192</u>					
			<u>Put to use</u>		<u>1</u>			


Dr. SURIYA RAGHAVAN
 MBBS, MD (Pulmonary Medicine)
 Reg. No: 56919
Medical Officer
 F.H.C. Moucode






Dr. SOORYA RAGHAVAN
 MBBS, MD (Pulmonary Medicine)
 Reg. No: 56919
 Medical Officer
 F.H.C Moudode