

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

| | |
|--|---|
| Name of Hospital: PHC CHITTARIKKAL | Name of District: KASARGOD |
| Name of Equipment: INFRARED LAMP | Equipment ID & Barcode: 141516- 1440287 |
| Make: NIL | Model: N/A |
| Serial Number: N/A | Warranty details: No Warranty |
| Date of purchase/ Year of manufacture/ Installation Date: 07/08/2014 | Present status of the equipment: Fully damaged |
| Date of breakdown: 18/10/2024 Date of registration of complaint through email/Toll free): Toll free | *AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC |
| Action taken: Checked the IR lamp, found lamp defective. Enquired vendor for Lamp quotation. | |

Recommendations for repair (required service details): Not recommending for repair

Cost of spares (specify parts and cost): IR Lamp: 850/-

| | |
|--|--------------------|
| #Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 171.71% | Asset Value: 495/- |
|--|--------------------|

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service report and Vendor quotation attached

Reasons for recommending the equipment as BER: Checked the equipment, found problem with IR lamp. Machine installed on 07-08-2014 and aged up to 10 years and 2 months. Repairing cost is more than 100% of asset value. Both criteria for RBER met. Considering the tender clause 5.3.14.1, We recommending the equipment for condemnation.

ABHINAV . O
(Signature)
 Date: 08.11.2024
 Name & Signature of CYRIX Authority with date


Remarks of Junior Consultant (Biomedical) NHM:
 Checked the machine and found problem with IR Lamp. Equipment installed on 07-08-2014. Repairing cost is more than 100% of asset value.

| | |
|---|--|
| Recommended for BER (Yes/ No): Yes | Jr. CONSULTANT Biomedical Engineering Arogya Kendram (NHM) Kasargod <i>(Signature)</i> Signature of JC-BM (NHM) |
| Date: 15/11/24 | |

| | | |
|------------------|------|---|
| Date: 19/11/2024 | Seal | <i>(Signature)</i> Signature of Superintendent / Medical Officer (i/c) |
|------------------|------|---|

*Not mandatory //Based on the period of life time value as per the BER guidelines * Attach Photograph

മെഡിക്കൽ ഓഫീസർ
 മുഖ്യമന്ത്രിയുടെ കാര്യാലയം
 വിറ്റാരിക്കാട്





BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1002615

CYRIX[®]
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

| | |
|--|---|
| Health Facility: <u>PAC Chittarikal.</u> | Call Registration Date: <u>18.10.2024</u> |
| Address: <u>Chittarikal.</u> | Caller ID: <u>141516.</u> |
| <u>Kasaragod.</u> | Date of Visit: <u>19.10.2024</u> |
| Ph: <u>6238415312.</u> | Asset No: <u>1440287.</u> |
| | EQPT Name: <u>Infrared Lamp</u> |
| | Manufacture: <u>NIL</u> Model: <u>NA</u> |
| | S. No: <u>NA</u> Dept: <u>Physiotherapy</u> |

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Not Switching ON.

Action Taken : Checked the machine and found that machine IR lamp defective and 2 pin code wire damaged. Need to change these spares for further servicing of the equipment.

Completed Date : 19.10.2024 Time : 2:30pm. Spare Required

Spare Replaced Requested

| Description | Qty. | Part Number | PR Number |
|-------------------------|----------|----------------|----------------|
| 1. <u>IR lamp</u> | <u>1</u> | <u>1440287</u> | <u>1002615</u> |
| 2. <u>Pin code wire</u> | <u>2</u> | <u></u> | <u></u> |
| 3. <u></u> | <u></u> | <u></u> | <u></u> |

| Cyrix Engineer | Date | Start Time | End Time |
|------------------|-----------------|---------------|----------------|
| <u>Abhinav O</u> | <u>19.10.24</u> | <u>2:00pm</u> | <u>2:30pm.</u> |

Customer Remark: Completed Pending

Service Engineer Name : Abhinav O
Signature: [Signature]
Date: 19.10.2024
Contact Number: 8943392989

Customer Name : Linku K Lal,
Signature: [Signature]
Date: 19/10/24
Contact Number: 6238415312
Designation: Nursing officer
Hospital Seal: [Seal]



Sreeraj Surgicals

GST No: 32ABYF53674H12L
DI No : 13/023/208/2011
13/024/218/2011
DL: 29-01-2011

Aiswarya Commercial Complex, Nr. Ashirvad Hospital, Thavakkara, Kannur - 670 002
Ph: 0497 2706307. Mob: 9349846307, 9387529848
Email: sreerajsurgicals@gmail.com, rajeshsreerajsurgical@gmail.com
www.sreerajsurgical.com Follow us on  www.facebook/ sreerajsurgicals

Date:

QUOTATION

DATE: 16-10-2024

TO


CYRIX HEALTH CARE PVT LTD

SIR,

THANK YOU FOR YOUR ENQUIRY TO THE RATE OF THE FOLLOWING ITEMS.

| SL.NO | ITEMS | RATE |
|-------|--------------------|--------|
| 1 | INFRARED BULB ONLY | 850.00 |

WE EXPECT YOUR VALUABLE ORDERS AND KIND CO-OPERATION WITH US
GST INCLUDED



THANKING YOU

Authorised Dealer:



Suppliers of Surgical Instruments, Hospital Furnitures, X-ray Materials, Lab Equipments & Disposables



Jayesh Krishnan <jayeshcyrix@gmail.com>

QUOTATION REQUEST FOR IR LAMP 1420506

6 messages

Jayesh Krishnan <jayeshcyrix@gmail.com>

Thu, Oct 10, 2024 at 4:32 PM

To: "Sreerajsurgicalsg@gmail.com" <Sreerajsurgicalsg@gmail.com>

Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

Dear sir,

Kindly send price and availability of IR lamp.

Requirement: Infrared Lalo

Make: Beurer

Jayesh Krishnan <jayeshcyrix@gmail.com>

Sat, Oct 12, 2024 at 10:44 AM

To: "Sreerajsurgicalsg@gmail.com" <Sreerajsurgicalsg@gmail.com>

Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

Sir,

Kindly send the quotation.

[Quoted text hidden]

—
Thanks & Regards

JAYESH KRISHNAN,
Sr. Biomedical Engineer,
District In-charge
Kasargod.
PH:- 7593847156

jayeshcyrix@gmail.comwww.cyrixhealthcare.com**Cyrix Healthcare Pvt Ltd****| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India****KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH**

Jayesh Krishnan <jayeshcyrix@gmail.com>

Mon, Oct 14, 2024 at 8:47 PM

To: "Sreerajsurgicalsg@gmail.com" <Sreerajsurgicalsg@gmail.com>

Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

Sir,

Kindly update.

[Quoted text hidden]

Jayesh Krishnan <jayeshcyrix@gmail.com>

Wed, Oct 16, 2024 at 10:18 AM

To: "Sreerajsurgicalsg@gmail.com" <Sreerajsurgicalsg@gmail.com>

Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

Sir,

Kindly send the quotation for Infrared bulb only.

[Quoted text hidden]

sreerajsurgicalsg <sreerajsurgicalsg@gmail.com>

Wed, Oct 16, 2024 at 10:22 AM

To: Jayesh Krishnan <jayeshcyrix@gmail.com>

25/10/2024, 21:54

Gmail - QUOTATION REQUEST FOR IR LAMP 1420506

Sent from Mail for Windows

[Quoted text hidden]

 **img905.pdf**
137K

sreerajsurgicals <sreerajsurgicals@gmail.com>
To: Jayesh Krishnan <jayeshcyrix@gmail.com>

Wed, Oct 16, 2024 at 12:31 PM


Sent from Mail for Windows

From: Jayesh Krishnan
Sent: 16 October 2024 10:18 AM
To: Sreerajsurgicals@gmail.com
Cc: DM Sarang-KLBEMP; Blesson Jose
Subject: Re: QUOTATION REQUEST FOR IR LAMP 1420506

Sir,

[Quoted text hidden]

[Quoted text hidden]

 **img907.pdf**
137K

1 R Lamp


Received from PMS Medi

7.5 Services Invoice no: 42
2014 dated 6/8/2014




1

1

Transferred to the PMS Palliative
Medicine Stock Register vol III Page no: 96



Handwritten signature and date: 7/11/2014
മെഡിക്കൽ ഓഫീസർ
പാലിയേറ്റീവ് മെഡിസിൻ
എറണാകുളം



Handwritten signature and date: 19/11/2024.
Handwritten text: 'പ്രസിദ്ധപ്പെടുത്തി' (Released) and 'സംസ്ഥാന സർക്കാർ' (State Government).