

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: TH MANGALPADY	Name of District: KASARGOD
Name of Equipment: HAEMATOLOGY ANALYZER	Equipment ID & Barcode: 143052- 1420353
Make: RAPID	Model: PE 6000
Serial Number: A6006160300062E11	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 25/01/2017	Present status of the equipment: Fully damaged
Date of breakdown: 26/10/2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked the machine, found CCU board, 2-way valve and 3 way valves are defective. Enquired OEM for quotation.	

Recommendations for repair (required service details): Not recommending for repair

Cost of spares (specify parts and cost): 3- Way valve- 2Nos: 36900/-
 2-Way valve- 2Nos: 36900/-
 CCU Board: 59500/-
 GST (18%): 23994/-
TOTAL: 157294/-

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 79% Asset Value: 199083/-

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service report and OEM quotation attached

Reasons for recommending the equipment as BER: Checked the machine, found machine CCU board, 3-way valve and 2 way valves are defective. Machine installed on 25-01-2017 and aged up to 7 years and 9 months. Repairing cost of equipment is 79% with respect to asset value. Both criteria for RBER met. Considering the tender clause 5.3.14.1, We recommending the equipment for condemnation.

Ajit Baburk *[Signature]*
 14-11-24
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM: Checked the machine and found their machine CCU Board, 3 way valve, & 2 way Valves are defective. Equipment installed on 25/01/2017, repairing cost of the equipment is 79% of the asset value.

Recommended for BER (Yes/ No): **Yes**
 Date: 15/11/24

[Signature]
 Signature of Superintendent / Medical Officer (i/c)

Date: 20-11-24 Seal **TALUK HEAD QUARTERS HOSPITAL MANGALPADY**
 *Not mandatory *Based on the period of life and value as per the BER guidelines * Attach Photograph

P.O. UPPALA
 KASARAGOD DIST.
 PIN: 671 322
 PH: 04998 242098

* Attach Photograph



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1022234

CYRIX[®]
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... THQH..... mangalpadu	Call Registration Date : 26-10-24
Address..... kasaragod	Caller ID : 143052
..... kerala	Date of Visit : 28-10-24
Ph : 9446354342	Asset No : 1420353
	EQPT Name : Haematology Analyser
	Manufacture... Pookan. Model : PE-6000
	S. No. A6006160300064 Dept..... LAB

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified :
Reading issues

Action Taken : checked the machine and found problem with CU board, 2 way valve and 3 way valve. need to replace those spark for another working condition of the equipment.

Completed Date : 28-10-24.. Time : 11:00 AM..... Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Ajith Babu	28-10-24	10:00 AM	11:00 AM

Customer Remark	Completed	Pending
	<input checked="" type="checkbox"/>	<input type="checkbox"/>

Service Engineer Name : Ajith Babu
Signature: *[Signature]*
Date: 28-10-24
Contact Number: 8848029288

Customer Name : Sreelakshmi
Signature: *[Signature]*
Date: 28/10/24
Contact Number: 9446354342
Designation :
Hospital Seal

TALUK HEAD QUARTERS HOSPITAL, MANGALPADY
P.O. UPPALA
KASARAGOD DIST.
PIN: 671322
PH: 04998 242098

SUPERINTENDENT, HEAD QUARTERS HOSPITAL, MANGALPADY, UPPALA, KASARAGOD.



Aspen Diagnostics Pvt. Ltd.

(Rapid Diagnostic Group Of Companies)



ISO 9001:2015
CERTIFIED COMPANY

Proforma Invoice No.: ADPL/S/RD/20241104/24-25				Date :04.11.2024		
To, Cyrix Healthcare Pvt. Ltd. Ernakulam, First Floor, Petta Junction 30/641B, Poonithura, Ernakulam-682038				Ref. : By Mail.		
				Date :04.11.2024		
				TH Mangalpady		
Sr. No	Description	Part No.	Price Each Ex- Work	Qty	Total (In Rs.)	
A	(Instrument Model: Hematology Analyzer) (Model : PE-6000) (S/No. A6006160300062EH10)					
	<u>Spares Procurement Details:</u>					
	(i)	3-Way Valve	PE00-006	18,450.00	2	36,900.00
	(ii)	2-Way Valve	PE00-005	18,450.00	2	36,900.00
	(iii)	CCU Board	PE60-010	59,500.00	1	59,500.00
Total					1,33,300.00	
(+ GST @ 18%					23,994.00	
Total: One Lakh Fifty Seven Thousand Two Hundred Ninety Four Only.					1,57,294.00	

Terms & Conditions :

(1) Warranty is not applicable for spare parts , (2) Delivery : 3-4 Weeks , (3) Payment Terms : 100% Advance)

Bank Account Details :-

Name of the Firm	: Aspen Diagnostic Pvt. Ltd.	Aspen Diagnostics Pvt. Ltd.
Bank Name	: ICICI Bank Limited	
Account No.	: 016005003239	
Account Type	: Current Account	
MICR Code	: 110229029	
IFSC Code	: ICIC0000160	



Jayesh Krishnan <jayeshcyrix@gmail.com>

REQUIRED SPARE QUOTATION FOR HAEMATOLOGY ANALYZER 1420353

3 messages

Jayesh Krishnan <jayeshcyrix@gmail.com>

Sat, Nov 2, 2024 at 1:16 PM

To: aspenkochi@gmail.com

Cc: Binto Joseph <binto@rdgc.in>, rapiddiagnosticgroup@gmail.com, Blesson Jose <zm2.klbemp@cyrix.in>, dmpl@rdgc.in, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear sir,

Haematolgy analyzer installed in TH Mangalpaday having problem and required following spare parts.

Machine details:

Make: Rapid

model: PE 6000

S/N: A6006160300062EH

Spares required:

1. 3-WAY VALVE- 2 Nos.

2. CCU BOARD

3. 2-WAY VALVE- 2 Nos.

—
Thanks & RegardsJAYESH KRISHNAN,
Sr. Biomedical Engineer,
District In-charge
Kasargod.
PH:- 7593847156jayeshcyrix@gmail.com
www.cyrixhealthcare.com**Cyrix Healthcare Pvt Ltd**
| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India
KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

Binto Joseph <binto@rdgc.in>

Sat, Nov 2, 2024 at 1:21 PM

To: service@rdgc.in

Cc: jayeshcyrix@gmail.com

Sir

Please share spare quotation

Regards

Binto

[Quoted text hidden]

service support <servicesupport@rdgc.in>

Mon, Nov 4, 2024 at 5:46 PM

To: Jayesh Krishnan <jayeshcyrix@gmail.com>

Cc: BINTO JOSEPH <binto@rdgc.in>, service@rdgc.in, "Rajaram.rdg" <rajaram.k@rdgc.in>, rustam@rdgc.in, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Jayesh Krishnan <jayeshcyrix@gmail.com>

Dear sir,

Please find the attached P.I. & share the P.O. along with payment details for billing.


20/11/2024, 23:30

Gmail - REQUIRED SPARE QUOTATION FOR HAEMATOLOGY ANALYZER 1420353

Thanks & Regards

Hari Ram

[Quoted text hidden]

 **IMG_0006.pdf**
580K


STOCK BOOK OF.....

Name of article..... HAEMATOLOGY ANALYSER.....

Date	No. and date of voucher or invoice	From whom received or to whom issued	Receipt	Issued	Balance after each transaction	Initials of receiver	Remarks
		Brought forward ..					
23.5.2012		BC 3000 PLUS					
	EQ103 16-5-2012	KMSCL TUN. CODE 21801300 212418/90278090	1	1	1		bl
		with (reagents) Diluent	20 litre x 4				
		Rinse	20 litre x 1				
		lyse.	2x500	1			
		Cleaner	Ax50	1			
		prob cleaners	Ax50	1			
23/5/12		Put in use		1			
25.1.2014		KMSCL Haematology analysers with UPS	1	1	2		bl
		Diluent 20 litre. : (2)	1	(From			bl
		Diluent 1 litre. : (3)					
		Rinse 10 l.	1				
		Rinse 1 litre.	1				
		lyse 500 ml	1				
		lyse 200ml	1	2			
		prob. clean 100ml	1				
		E. x cleaner. 100ml	1				
25-1-2012		Put in use		1			
16.6.2014		(Mouse) for haematology analyser					
		C'Dball. (style 36)	1	1	1		bl
16.6.2012		put in use		1	(adding)		bl

Carried over ..
 QUARTERS HOSPITAL
 P.O. UPPALA
 KASARAGOD DIST.
 PIN: 671 322
 PH: 04998-242098

SUPERINTENDENT
 MANGALPADY
 UPPALA, KASARAGOD - 671 322



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)
TOLL FREE NO: 1800 - 425-7669
BAR CODE - (8004-890615225) 1420353

FULLY AUTO HEMATOLOGY ANALYZER
MODEL: PE-6000
SERIAL NUMBER: A6006160300062EH 10
100-240V~ 50/60Hz 5150VA

SHENZHEN PROKAN ELECTRONICS INC.
No.1002,Bld. 4,Fantasia MIC Plaza,Pengli
Pioneering Park,Nanshan,Shenzhen,518067 P.R.China
Factory Add.:The 2nd floor, Yanda tech-park, FengHuangGang
XIXiang, Bao'An district, Shenzhen,518128 P.R.China
Shanghai International Holding Corp GmbH (Europe)
Eiffelstraße 80, 20537 Hamburg, Germany



Handwritten signature

SUPERINTENDENT
TALUK HEAD QUARTERS HOSPITAL
MANGALPADY
UPPALA, KASARAGOD - 671 322
KASARAGOD DIST. PIN: 671 322
PH: 04998 242098
MANGALPADY

71 322