


**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: CHC CHERPULASSERY	Name of District: PALAKKAD
Name of Equipment: NEBULIZER	Equipment ID & Barcode: #145896 - 0950357
Make: AMKAY PRODUCT PVT LTD	Model: AMKAY
Serial Number: NA	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: In ticket master Installation date 27/03/2023, As per stock 15/10/2018 stock book attached for reference.	Present status of the equipment: Fully damaged
Date of breakdown: 11/11/2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked the nebulizer found motor stuck. Found Problem with motor and piston, need to be replace these spares for further checking of machine.	
Recommendations for repair (required service details): No recommendations	
Cost of spares (specify parts and cost): NA	
#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 1600/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix Service report attached.	
Reasons for recommending the equipment as BER: Checked the nebulizer found motor stuck. Found Problem with motor and piston, need to be replace these spares for further checking of machine. In ticket master installation date is 27/03/2023 As per the stock book Equipment installed on 15/10/2018 aged up to 6 years .Quotation not submitted since spare are not available from market. So we are recommended the equipment for condemnation	
<p align="right">AKHILESH P Name & Signature of CYRIX Authority with date <i>21/11/24</i></p>	
Remarks of Junior Consultant (Biomedical) NHM: <i>Nebulizer installed on 15/10/2018 aged up to 6 years quotation not submitted.</i>	
Recommended for BER (Yes/ No): <i>Yes</i>	<p align="right">Signature of Consultant (Biomedical Engg) NHM (Arogyakeralam), Palakkad <i>21/11/24</i></p>
Date: <i>21/11/24</i>	
Date: <i>22-11-2024</i>	<p align="center">Seal</p> <p align="right">Signature of Superintendent/ Medical Officer (i/c)</p>




*Not mandatory

#Based on the period of life and value as per the BER guidelines


Attach Photograph

Nebulizer

Date	Item Received from Store	Received	Issued	Balance		Sign	Comments
				Physical	New Balance		
01/01/08	Received from Store	1			1		
02/01/08	Issued to OP		1	Nil			
07/01/10	Received from Store (CHC)	2			5		
07/01/10	Issued to MW		1	Nil			
07/01/10	Issued to FW		1				
08/01/12	Received from Store (Cosca diagnostic Centre)	2			5		
08/01/12	Issued to CH (Tr. out to CH Fi P-7)		1	Nil			
	Issued to OP (out to OP Fi P-7)		1				
15/06/15	Received from Store [NRHS]	1			6.4		
16/06/15	Rs [2000] Issued to MW		1	Nil			
09/01/15	Received from Store [Donation]	1			7		
09/01/15	Issued to OP		1		9.7		
03/11/16	Received from Store	2			2		
03/11/16	Condemned as per O: NO: 145/18		2				
15/10/18	Received from Store (HCC)	3			7.5		
15/10/18	Issued to OP (Tr. out to OP Fi Substock)		2		5.5		
15/10/18	Issued to OP		1		4.5		
<p>Transfer over to new register From MW, CH, OP from 25/11/2018 MW = Fi P-7 From OP = Fi P-7</p>							
15/10/18	Tr. out to OP Fi P-7		1		4		
15/10/18	Tr. out to FW Fi P-7		3		1		
15/10/18	Tr. out to CH Fi P-7		1		Nil		



(Name)
 (Designation)
 (Address)



AO
 03-13-2013
 09:50:37
 03-13-2013 09:50:37
 03-13-2013 09:50:37