

**PROFORMA**  
**RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)**  
**BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**


Name of Hospital: THQH KOTHUPARAMBA	Name of District: KANNUR
Name of Equipment: NEBULIZER	Equipment ID & Barcode: 146373-1321562
Make: NOREN	Model: READY MIST
Serial Number: NA	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 16.04.2012	Present status of the equipment: Fully damaged
Date of breakdown: 12.11.2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked the equipment found that equipment is not switching ON. Found that Motor, Piston are defective. Need to replace these spares for further checking.	
Recommendations for repair (required service details): No recommendations	

Cost of spares (specify parts and cost): NA

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 1353/-
---	---------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service Report attached

Reasons for recommending the equipment as BER Checked the equipment found that equipment is not switching ON. Found that Motor, Piston are defective. Need to replace these spares for further checking. Need to replace these spares for further checking. Equipment installed on 16.04.2012 aged up to 12 years 7 months. Quotation not submitted since spares are not available in the market. So we recommend the equipment for condemnation.

Avinash T   
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM: checked & identified that not working due to motor, piston are defective. Aged up to 12 years & above. Quotation not submitted. But the spare cost is economically not feasible so recommended to BER as per BER Protocol.

Recommended for BER (Yes/ No):

Date: 23/11/24

Date:  Signature of Superintendent / Medical Officer (i/c)

\*Not mandatory

#Based on the period of life and value as per the BER guidelines

\* Attach Photograph

SUPERINTENDENT  
 GOVT. TALUK HEADQUARTERS  
 HOSPITAL  
 KUTHUPARAMBA, PIN - 670 643





**BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAMME  
UNDER  
NATIONAL HEALTH MISSION**



SERVICE PROVIDED AT  
Tender No. WO-37/2021/2022/698

No : 133089

SUPERINTENDENT  
GOVT TALUK HEADQUARTERS  
KUTHUPARAMBA, PIN-686043

**CYRIX**<sup>®</sup>  
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report** 30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website :www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility... T.H.Q.H. Koothuparamba...	Call Registration Date : ... 12-11-24 .....
Address..... Koothuparamba .....	Caller ID : ..... 146373 .....
..... Kanner, Kerala .....	Date of Visit : ..... 13-11-24 .....
Ph: ..... 956 2847869 .....	Asset No : ..... 1321562 .....
	EQPT Name : ..... Nebuliser .....
	Manufacture..... No.ven..... Model: Reedy mist.
	S. No..... NA..... Dept..... OP.....

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : ..... Motor & piston are defective .....

Action Taken : ..... checked the equipment found that equipment is not switching on. Found that motor, piston are defective. Need to replace these spare for further checking .....

Completed  Date : 13-11-24 ..... Time : 11:00 AM ..... Spare Required

Spare Replaced  Requested


Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Shyamkrishnan	13-11-24	10:30 AM	11:00 AM

Customer Remark Completed  Pending

Service Engineer Name : Shyamkrishnan  
Signature: [Signature]  
Date: 13-11-24  
Contact Number: 8281188514

Customer Name : Renu K.G.  
Signature: [Signature]  
Date: 13-11-24  
Contact Number: 9946644326  
Designation : Senior Nursing Officer  
Hospital Seal





L. P. Rao  
3/11/20

058

NEBULIZER


STOCK REGISTER

Date	Particulars	Received from Store	Issued	Balance	Signature
15/1/2020	Tr. in from old book Vol 2 Pm. 167			04 ✓	[Signature]
	Part Store 1/1/18 BPL medical Technologies (E)				
	Kivluskannu	70134021 70134022			
	Exp date - 31/3/2024				
	BPL - 70134021 Hi Advance - 67131976 - 70134022 Compac - 67137081 (Cynix ID - 1321665)				
	put in use		04		[Signature]
13/06/2024	Received from store INNO 00062/24-25				
	(From OLIVE HEALTH CARE AND SURGICAL) P. 1/20/01			05	[Signature]
	put in use		01		[Signature]
6/9/21	Tr. out to unseparable book on 6/9/2021				
	for condensation (Rel date 11/9/18)		1	14	
	(BPL breath. ezee)				
	(Cynix ID - 1321666)				
16.4.12	Physically found on verification		1		
	(Koren - Cynix ID - 1321562)				


Dr. K.G. S. Rao  
[Signature]

K. R. S. Rao  
[Signature]

SUPERINTENDENT  
GOVT. TALUK HEADQUARTERS  
HOSPITAL  
KUTHUPARAMBA, PIN - 670 643







*Kudhans  
Rajun*  
BME.

*[Signature]*  
Rema K.G  
SNO.



*[Signature]*  
SUPERINTENDENT  
GOVT. TALUK HEADQUARTERS  
HOSPITAL  
KUTHUPARAMBA, PIN - 670 643