

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: PHC MUNDERI	Name of District: KANNUR
Name of Equipment: WEIGHING MACHINE	Equipment ID & Barcode: #139523-1340075
Make: SAMSO	Model: SKINNY
Serial Number: NA	Warranty details: NO Warranty
Date of purchase/ Year of manufacture/ Installation Date: 24-08-2019	Present status of the equipment: Fully damaged
Date of breakdown: 08-10-2024	*AMC/CAMC Period agreed
Date of registration of complaint through email/ Toll free): Toll free	at the time of purchase: NO AMC/ CAMC
Action taken: Checked the machine and found that Machine is not switching on, Found main board, Load cells damaged. Need to replace these spares for further checking.	
Recommendations for repair (required service details): No recommendations	

Cost of spares (specify parts and cost): N/A


# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: N/A	Asset Value: 850/-
---	--------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Service report attached

Reasons for recommending the equipment as BER: Checked the machine and found that machine is not switching ON. Found that main board and load cells damaged. Equipment installed on 24-08-2019 Aged up to 5 years 2 months. Quotation not submitted since spares are not available in the market so we recommend the equipment for condemnation



Avinash T 
 Name & Signature of CYRIX Authority with date 15/10/24

Remarks of Junior Consultant (Biomedical) NHM: checked & identified that machine not working due to the main board and load cells are damaged and aged up to 5 years & above. So recommended for BER as per BER guidelines.

Recommended for BER (Yes/ No):	 16/10/24 OFFICE OF THE JUNIOR CONSULTANT BIOMEDICAL NATIONAL HEALTH MISSION KANNUR Signature of JC BM (NHM)
--------------------------------	---

Date: 16/10/24 Shejina A Shef Jr PHN Girji	 16/10/24 Signature of Superintendent / Medical Officer (i/c)
--	---

*Not mandatory #Based on the period of life and value as per the BER guidelines Attach Photograph





**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1004000

CYRIX[®]
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website :www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility... <u>PHC... Munderi</u>	Call Registration Date : ... <u>08/10/24</u>
Address..... <u>Munderi</u>	Caller ID : ... <u>139523</u>
..... <u>Kannur</u>	Date of Visit : ... <u>09/10/24</u>
Ph: <u>9496 453180</u>	Asset No : ... <u>1340075</u>
	EQPT Name : ... <u>Weighing Machine</u>
	Manufacture... <u>Samsa</u> Model : <u>Slippy</u>
	S. No..... <u>NA</u> Dept. <u>Palliative</u>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Equipment not turning on.

Action Taken : Checked the equipment found that equipment is switching on, Found main board load cells damaged need to replace these spares for further checking.

Completed Date : 09/10/24 Time : 11:00am Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Avinash</u>	<u>09/10/24</u>	<u>10:30AM</u>	<u>11:00am</u>

Customer Remark Complete Pending

Service Engineer Name : <u>Avinash</u>	Customer Name : <u>Shejina A</u>
Signature: <u>[Signature]</u>	Signature: <u>[Signature]</u>
Date: <u>09/10/24</u>	Date: <u>9/10/2024</u>
Contact Number: <u>63640 64948</u>	Contact Number: <u>9496 453180</u>
	Designation : <u>Jr PHN Gr I</u>
	Hospital Seal




MEDICAL OFFICER
PRIMARY HEALTH CENTRE
MUNDERI - 670 591

151

STOCK REGISTER OF




Name of Article

Date	No. and date of voucher or invoice	From whom received or to whom issued	Receipt	Issued	Balance after each transaction	Initials of Receiver	Remarks
		Brought forward				53.25	1/2022
18	18	Oxygen cylinder	1	-	-	166.25	12/2023
24/8/19	19	O ₂ cylinder	1	-	-	1413.35	12/2024
24/8/19	20	O ₂ cylinder	1	-	-	1179.23	12/2024
24/8/19	21	O ₂ flow meter	1	-	-	21.80	10/2024
24/8/19	22	Salt testing kit	5	-	-	467.80	12/2024
24/8/19	23	Stetho meter	1	-	-	308.04	12/2024
24/8/19	24	Stethoscope	1	-	-	67.76	1/2024
"	25	Thermometer	1	-	-	102.68	3/2024
"	26	Urine strip	9	-	-	1658.04	1/2024
"	27	weighing machine	1	-	-	4239.49	12/2024
"	28	weighing machine	1	-	-	58.00	
"	29	Black disposal bag	1	-	-	58.00	
"	30	Red disposal bag	1	-	-	58.00	
"	31	Yellow disposal bag	1	-	-	58.00	
"	32	Hand towel	1	-	-	36.40	
"	33	Bed sheet	1	-	-	320.00	
		Issued to					
		Ravi Kanady					above items
		Carried over					



GPC. 25/267/2007/5,000x100. SI. © Govt. of Kerala.

Beena Mathew
PHN



PHC
PHC