

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: CHC MULLASSERY	Name of District: Thrissur
Name of Equipment: BP apparatus	Equipment ID & Barcode: #142453 - 0832699
Make: LIFELINE	Model: Lifecare
Serial Number: NA	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 28-11-2019	Present status of the equipment: Fully damaged
Date of breakdown: 23-10-2024	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Date of registration of complaint through email/Toll free): Toll free	
Action taken: checked the machine and found pump and motor are defective. Need to replace these spares for checking working condition of equipment.	
Recommendations for repair (required service details): Not Recommending	
Cost of spares (specify parts and cost): NA	

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 2200/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix report attached.	

Reasons for recommending the equipment as BER: Checked the machine and found pump and motor are defective. The machine installed on 28-11-2019 and aged up to 4 years and 11 months. Quotation not submitted since the spares are not available in the market. So, we are recommending for condemnation.

[Signature]
BIJO T JOY
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:
spare not available

Recommended for BER (Yes/ No): <i>Yes</i>	<i>[Signature]</i> Signature of JC BM (NHM)
Date:	

Date:	Seal	<i>[Signature]</i> Signature of Superintendent Medical Officer (i/c)
-------	------	---



COMMUNITY HEALTH CENTRE
MULLASSERY
 Attach Photograph

*Not mandatory

##Based on the period of life and value as per the RBER guidelines



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No 1006871



ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website :www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>CHC</u>	Call Registration Date : <u>23/10/24</u>
Address..... <u>Mullassery</u>	Caller ID : <u>142453</u>
..... <u>Thrissur</u>	Date of Visit : <u>24/10/24</u>
Ph : <u>9539242234</u>	Asset No : <u>0832699</u>
	EQPT Name : <u>BP apparatus</u>
	Manufacture <u>LifeLine</u> Model : <u>lifecare</u>
	S. No. <u>NA</u> Dept. <u>PP unit.</u>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : equipment not going on

Action Taken : Checked the machine and found pump and motor are defective. need to replace these spares for checking working condition of equipment

Completed Date : 24/10/24 Time : 10:00am Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Christy</u>	<u>24/10/24</u>	<u>9:00am</u>	<u>10:00am</u>

Customer Remark Completed Pending

Service Engineer Name : Christy Customer Name : Sandhya . C.A
Signature: [Signature] Signature: [Signature]
Date: 24/10/24 Date: 24/10/24
Contact Number: 6238844347 Contact Number: 9539242234
Designation : JPH SUPERINTENDENT
Hospital Seal: COMMUNITY HEALTH CENTRE MULLASSEERY

10.19
05

88

Page No 17

Duggal B. P. Apparatus

Date	Item	Quantity	Iss	Value
9/2/19	Duggal B P Apparatus	1 No		1
28/11/19	B P apparatus Digital.	1		



