

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: PHC KAIPPAMANGALAM	Name of District: Thrissur
Name of Equipment: Centrifuge	Equipment ID & Barcode: 0844091 & 142422
Make: Rotek	Model: RCH-8
Serial Number: 1908	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: Ticket master installation date – NA Stock book installation date – 10-01-2019	Present status of the equipment: Fully damaged
Date of breakdown: 23-10-2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: The machine checked and found transformer, head with tubes, knob and motor are defective. Need to replace these spares for further checking working condition of equipment.	

Recommendations for repair (required service details): Not Recommending

Cost of spares (specify parts and cost):	Centrifuge motor	-3000/-
	Knob	- 50/-
	Transformer set with switch	- 500/-
	8 tube head with tubes	- 1000/-
	GST	- 819/-
	TOTAL	- 5369/-

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 88%	Asset Value: Ticket master asset value 9651/- In stock book asset value 6100/- (stock book attached)
--	---

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix report attached and OEM quotation attached

Reasons for recommending the equipment as BER: The machine checked and found transformer, head with tubes, knob and motor are defective. In ticket master installation date is Nil, As per stock book machine installed on 10-01-2019 and aged up to 5 years and 9 months. In ticket master asset value is 9651/- but As per the stock book asset value is 6100/-. The spare cost is 88% of asset value. Both criteria met for RBER. As per tender clause 5.3.14.1 we recommending the equipment for condemnation.


BIJO T JOY
Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:
transformer damaged.

Recommended for BER (Yes/ No): *Yes*

Date: _____ Signature of JC BM (NHM)

Date: _____ Signature of Superintendent / Medical Officer (i/c)



*Not mandatory

//Based on the original life and value as per the BER guidelines

* Attach Photograph



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1017734

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>PHC</u> Address..... <u>KALIPPAMANGALAM</u> <u>CHIRASSUR</u> Ph:..... <u>7909634211</u>	Call Registration Date:..... <u>23/10/2024</u> Caller ID:..... <u>142422</u> Date of Visit :..... <u>23/10/2024</u> Asset No:..... <u>0844091</u> EQPT Name :..... <u>CENTRIFUGE</u> Manufacture..... <u>ROTEK</u> Model: <u>NA</u> S. No..... <u>NA</u> Dept..... <u>LAB</u>
--	---

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified :

Action Taken : THE MACHINE CHECKED & FOUND TRANSFORMER, HEAD WITH TUBES, KNOB & MOTORS ARE DEFECTIVE. NEED TO REPLACE THESE SPARES FOR FURTHER CHECKING WORKING CONDITION OF EQUIPMENT

Completed Date : 23/10/2024 Time : 10:00AM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>AMBROO.V.S</u>	<u>23/10/24</u>	<u>9:00AM</u>	<u>10:00AM</u>

Customer Remark Completed Pending

Service Engineer Name : <u>AMBROO.V.S</u> Signature: <u>[Signature]</u> Date: <u>23/10/24</u> Contact Number: <u>79099650</u>	Customer Name : <u>[Signature]</u> Signature: <u>[Signature]</u> Date: <u>23/10/24</u> Contact Number: <u>[Signature]</u>
--	--



B&C INDUSTRIES, West Vengola P.O, Perumbavoor ,

Ernakulam , Kerala - 683556 Mail ID – mail@rotekinstruments.com

Ph: 9446139563, 8593077000, 8593000606

GST : 32AAEFB4140N1ZO

QTN- B&C /298/24-25

19/11/2024

To

Cyrix Health Care

Ernakulam

Sub: Quotation for centrifuge repair

Sir

We are sending the rate of centrifuge spares. The rate is given below for your kind consideration

Centrifuge - RCH- 8

Centrifuge motor	- Rs 3000 -00
Knob	- Rs 50 -00
Transformer set with switch	- Rs 500 -00
8 tube head with tubes	- Rs 1000 -00

4550 -00

GST @ 18% 819 -00

5369 -00
=====

Thanks & Regards

For B&C Industries



quotation enquiry - phc kaipamangalam - 0844091-142422

2 messages

BiJo T Joy <bijocyrix@gmail.com>
To: mail@rotekinstruments.com
Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

Tue, 19 Nov, 2024 at 10:39 am

Dear sir,
There is a centrifuge complaint reported from phc kaipamngalam, Thrissur, Kerala, Regarding the complaint with , Rotek, RCH-8 centrifuge machine ,Please share the price and availability of the below mentioned part, as soon as possible.
Equipment - centrifuge
Make : ROTEK
Model: RCH-8
quotation for - motor, knob, transformer and head with tube
Serial No: 1903
Hospital- phc kaipamangalam
District - Thrissur
Barcode - 0844091
Caller id - 142422
With Regards,
BIJO T JOY
DISTRICT INCHARGE THRISSUR,
CYRIX HEATHCARE PRIVATE LIMITED
30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia
M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

Rotek Instruments <mail@rotekinstruments.com>
To: BiJo T Joy <bijocyrix@gmail.com>

Tue, 19 Nov, 2024 at 3:39 pm

Quotation

Sir
PFA

Thanks & Regards



ROTEK INSTRUMENTS
B&C Industries, West Vengola P.O.,Perumbavoor- 683556, Ernakulum Dist., Kerala,
PH: 8593070000/859300606 / 9446139563 web: www.rotekinstruments.com



[Quoted text hidden]

CENTRIFUGES

10/1/19

sponsored by Chemtura Diagnostics
Chalakkudy Amount 6100/-
Mfr: Band C Industries.

Issued to lab. |
pharmacy | NIL

29/3/19

purchased from Kochin Inlab equipment
84211999 Rate: 17,660 Amt: 17,800.

Issued to pp. lab | NIL.

1) Remi swing out rotor for R&C plus
Rate: 5800/-

Issued to lab | NIL



Andy

