

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: CHC PAPPINISSERY	Name of District: KANNUR
Name of Equipment: ECG MACHINE	Equipment ID & Barcode: 143690-1330853
Make: EDAN	Model: SE-1200
Serial Number: 106759M12C03880001	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 11.03.2013	Present status of the equipment: Fully damaged
Date of breakdown: 30.10.2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked the equipment found that equipment is not switching ON. Found main board defective. Enquired OEM for quotation.	

Recommendations for repair (required service details): No recommendations

Cost of spares (specify parts and cost): NA

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value:

Asset Value: 62000/-

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service Report & OEM End of life letter attached


Reasons for recommending the equipment as BER: Checked the equipment found that equipment is not switching ON. Found main board defective. Enquired OEM for quotation. Equipment installed on 11.03.2013 aged up to 11 years. OEM informed that machine is very old and spare parts are not available, end of life letter issued. As per tender clause 5.3.14.2 we recommend the equipment for condemnation.

Avinash 
Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM: *Checked and identified that machine not working due to the main board defective and also aged upto 11 years EOL issued by OEM. So recommended for BER as per BER protocol.*

Recommended for BER (Yes/ No): Yes

Date: 11/11/24



Bindumole CS

SNO:
Date: 8/11/2024

Binda.
Seal


Signature of JC BM (NHM)

NATIONAL HEALTH MISSION
KANNUR - 70002

Signature of Superintendent / Medical Officer (i/c)

*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

for
30/11/2024
Medical Officer
Pappinisseri



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 230615

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>CHC PAPPINISSERY</u></p> <p>Address <u>PAPPINISSERY</u> <u>ICANNUR</u></p> <p>Ph : <u>8848665721</u></p>	<p>Call Registration Date : <u>30/10/24</u></p> <p>Caller ID : <u>143690</u></p> <p>Date of Visit : <u>31/10/24</u></p> <p>Asset No. : <u>1330853</u></p> <p>EQPT Name : <u>ECG machine</u></p> <p>Manufacture <u>Edan</u> Model : <u>SE 1200</u></p> <p>S. No. <u>106759M120388⁰⁰¹</u> Dept. <u>OP</u></p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Equipment is not switching ON

Action Taken : checked the equipment found that equipment is not switching ON. Found main board defective. Enquired OEM for quotation.

Completed Date : 31/10/24 Time : 12:40pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Aswin</u>	<u>31/10/24</u>	<u>12:20pm</u>	<u>12:40pm</u>

Customer Remark Completed Pending



Service Engineer Name : Aswin
Signature : [Signature]
Date : 31/10/24
Contact Number : 8301012709

Customer Name : Bindumole C S
Signature : [Signature]
Date : 31/10/2024
Contact Number : 8848665721
Designation : SNO
Hospital Seal :

12.11.2024

To,

M/s Cyrix Healthcare

First Floor, Petta Junction 301647 B

Poonithura, Ernakulam- 682038

Sub: Reg. End of Life of ECG Machine 12 Channel

Dear Sir/Madam,

We would like to inform you that the equipment as per the below details is very old and the spare parts are not available. Hence we are unable to repair the equipment

Equipment: ECG Machine 12 Channel


Model: SEI200

SN: 106759M12C03880001

Make: Edan Instruments Inc

Thanking You

Anoop N



Manger

Sree Gokulam Pvt .Ltd.





ONE TIME SERVICE REQUIRED ECG MACHINE 1330853#143690@ CHC PAPPINISSERI,KANNUR

2 messages

avinash t <avinashtcyrix@gmail.com>

Tue, 12 Nov, 2024 at 11:46 am

To: service@gokulamhealthcare.com

Dear Sir,

There is a complaint reported from CHC PAPPINISSERI, KANNUR, KERALA Regarding the complaint with ECG MACHINE. We required your one time service support for rectifying the problem.

ECG MACHINE

barcode: 1330853

make: EDAN

Model:SE-1200

SL NO- 106759M12C03880001

complaint- NOT WORKING

Thanks & Regards

Avinash T

District charge, kannur

bemp kerala

mobile: +91 6364064948

E mail: avinashtcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD

30/641B | Petta | Punithura | Kochi | Kerala- 682038

www.cyrixhealthcare.com

service gokulamhealthcare <service@gokulamhealthcare.com>

Tue, 12 Nov, 2024 at 2:36 pm

To: avinash t <avinashtcyrix@gmail.com>


Dear Sir,

Kindly see the attached EOL.

[Quoted text hidden]

ECG Machine

Date	Items	Rd	Issued	Batch	Cont	Sig HDC
30/6/05	Received from the store	1		1		✓
8/13/08	Received from the store (Automat)	1		2		✓
11/3/13	Received from the store (Store stock book HDC page no: 154)	1		3	76755	81/✓
11/3/13	In use			3		✓
	Condemned as per order no GO (MS) NO 389 / 2009 GO (MS) NO 66 / 2014 H.R. P.W. dt 3-1-2014 ma (640116 dt 25.7.16) 30.6.05, — \$.3.08 IN USE			2	#1	✓
30-10-2020	Received from the store (BPL Machine) (From Moragha. FHC) NHM FUND. (IN USE)	1		2		✓



For
 Medical Officer
 Community Health Centre
 Pappinisseri
 30/08/24

