

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: CHC KUZALMANNAM	Name of District: PALAKKAD
Name of Equipment: HEAMATOLOGY ANALYZER	Equipment ID & Barcode: #142686 - 0930173
Make: PROKAN	Model: PE6000
Serial Number: A6006160200018EH10	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: In ticket master Installation date - NIL, As per stock 26/03/2016 stock book attached for reference.	Present status of the equipment: Fully damaged
Date of breakdown: 24/10/2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked the hematology analyzer found variation. Problem with ccu board, wbc chamber with ruby, dilution block, 3 way valve and 2 way valve damaged. need to be replace these spares for further checking of machine.	
Recommendations for repair (required service details): No recommendations	
Cost of spares (specify parts and cost): 3 way valve = 36900/- 2 way valve = 36900/- Dilution block = 12000/- Wbc chamber = 35250/- Ccu board = 59500/- 18% gst = 32499/- Total = 213049/-	
#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 107%	Asset Value: 199083/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix Service report attached and quotation attached.	
Reasons for recommending the equipment as BER: Checked the hematology analyzer found variation. Problem with ccu board, wbc chamber with ruby, dilution block, 3 way valve and 2 way valve damaged. need to be replace for further checking of machine. In ticket master installation date is Nil, As per the stock book Equipment installed on 26/03/2016 aged up to 8 years 7 months. Repairing cost is more than 100% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending for condemnation.	
Name & Signature of CYRIX Authority with date AKHILESH P <i>[Signature]</i> 5/11/24	
Remarks of Junior Consultant (Biomedical) NHM: <i>As per stock book Haematology analyzer installed on 26/03/2016 and aged up to 8 years spare cost is more the usef value. Hence both criteria of RBER met.</i>	
Recommended for BER (Yes/ No): <i>Yes</i>	Consultant (Biomedical Engg) NHM (Arogyakeralam), Palakkad <i>[Signature]</i> Signature of JC BM (NHM)
Date: <i>5/11/24</i>	

Date:

Seal



Signature of Superintendent / Medical Officer (i/c)

Superintendent
Community Health Centre
Kuzhalmannam- 678 702

*Not mandatory

#Based on the period of life span value as per the BER guidelines

* Attach Photograph



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1021994

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility.....C.H.C...Kuzhalmannam	Call Registration Date : ...24.10.2024.....
Address.....Palakkad.....	Caller ID :142686.....
.....	Date of Visit :24.10.2024.....
Ph:.....04922 274350.....	Asset No :0930173.....
	EQPT Name : ...Haematology Analyzer
	Manufacture: Trokan..... Model: PE 6000
	S. No: Abo061602000..... Dept: Laboratory
	18EH10

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Result variation.....

Action Taken : ...checked the machine found variation. Problem with CCU board, wbc chamber with ruby, dilution block, 3 way valve, 2 way valve damaged. need to replace these spare's for further checking of machine.....

Completed Date : 28.10.24..... Time : Spare Required

Spare Replaced Requested


Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
	28.10.24		

Customer Remark Completed Pending

Service Engineer Name : Dileep A.D
 Signature: *Dileep A.D*
 Date: 28.10.2024
 Contact Number: 904875199

Customer Name : DR. SINDHU M.A
 Signature: *Sindhu*
 Date: 28-10-24
 Contact Number:
 Designation : Superintendent
 Hospital Seal: Community Health Centre Kuzhalmannam- 678 702





Aspen Diagnostics Pvt. Ltd.

(Rapid Diagnostic Group Of Companies)



Proforma Invoice No.: ADPL/S/RD/20241101/24-25				Date :01.11.2024	
To, Cyrix Healthcare Pvt. Ltd. Ernakulam, First Floor, Petta Junction 30/641B, Poonithura, Ernakulam-682038				Ref. : By Mail.	
				Date :01.11.2024	
				CHC Kuzhalmannam	
Sr. No	Description	Part No.	Price Each Ex- Work	Qty	Total (In Rs.)
	(Instrument Model: Hematology Analyzer) (Model : PE-6000) (S/No. A6006160200018EH10)				
A	Spares Procurement Details:				
(i)	3-Way Valve	PE00-006	18,450.00	2	36,900.00
(ii)	2-Way Valve	PE00-005	18,450.00	2	36,900.00
(iii)	Dilution Block	PE00-022	12,000.00	1	12,000.00
(iv)	WBC Chamber	PE60-003	35,250.00	1	35,250.00
(vi)	CCU Board	PE60-010	59,500.00	1	59,500.00
Total					1,80,550.00
(+) GST @ 18%					32,499.00
Total: Two Lakh Thirteen Thousand Fourty Nine Only.					2,13,049.00

Terms & Conditions :

(1) Warranty is not applicable for spare parts (2) Delivery : 3-4 Weeks (3) Payment Terms : 100% Advance

Bank Account Details :- *dm*

Name of the Firm : Aspen Diagnostic Pvt. Ltd.
Bank Name : ICICI Bank Limited
Account No. : 016005003239
Account Type : Current Account
MICR Code : 110229029
IFSC Code : ICIC0000160

Aspen Diagnostics Pvt. Ltd.

Authorized Signatory





REQUIRED QUOTATION # 0930173 - 142686 HEMATOLOGY ANALYZER @ CHC KUZHALMANNAM , PALAKKAD

2 messages

Akhilesh P <akhileshcyrix@gmail.com>

Fri, 1 Nov, 2024 at 8:55 am

To: service support <servicesupport@rdgc.in>, Service.RDG <service@rdgc.in>

Cc: binto@rdgc.in, aspenkochi@gmail.com, Blesson Jose <Zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear sir,

Please share the price and stock availability of

1. CCU BOARD
2. WBC CHAMBER WITH RUBY
3. DILUTION BLOCK
4. 3 WAY VALVE 2 NOS
5. 2 WAY VALVE 2 NOS

Deatils are furnished below

Hospital Name : CHC KUZHALMANNAM

Department : LABORATORY

Equipment Name : HEMATOLOGY ANALYZER

Make : RAPID

Model : PE 6000

SN : A6006160200018EH10

Thanks & Regards

AKHILESH P
DISTRICT INCHARGE - PALAKKAD
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com

service support <servicesupport@rdgc.in>

Fri, 1 Nov, 2024 at 12:36 pm

To: Akhilesh P <akhileshcyrix@gmail.com>

Cc: binto@rdgc.in, aspenkochi@gmail.com, Blesson Jose <Zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>.

Service.RDG <service@rdgc.in>

Dear sir,

Please find the attached P.I. & share the P.O. along with payment details for billing.

Thanks & Regards

Hari Ram


[Quoted text hidden]


Haematology Analyser - 1

25

Date	Particulars	Receipt	Issue	Balance	Main Stock	Remarks
26/3/16	Rapid PE-6000	1 ✓	1	NIL	1	✓
	UPS online (KMSCL)	1	1		1	
	SI: NO. 1602000018EH 26/3/16					
01.4.23 (CP)	Haematology analyser 1. 3-part (Gramaparrubayath aysard)			2.		
	Accessories					
	Diluent (DX3) 20L - 4					
	Lyse (LX3) 200ml - 4					
	Probe cleaner (DX3) - 2					
	Depositeruses (4x30ml) (3x3) - 2					
	cleaner (CX3) (4x50ml) 1000ml - 7					




 Superintendent
 Community Health Centre
 Kuzhalmannam - 678 702



Superintendent
Community Health Centre
Kuzhalmannam- 678 702

Cr. Chik
 pr: 11/09/20
 10: 09:30
 Cal: 12-
 Cal: 11-
 TOLL FREE NO: 1800-425-7669

CYRIX HEALTH BEMP SERVICE
 PREVENTIVE MAINTENANCE
 THIS DONE ON: _____
 NEXT DUE: _____
 PERFORMED BY: _____