



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1017366

CYRIX[®]
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility: <u>Wkch Alappuzha</u>	Call Registration Date: <u>7/10/24</u>
Address:	Caller ID: <u>139135</u>
<u>DIST: ALAPPUZHA</u>	Date of Visit: <u>8/10/24</u>
Ph: <u>9249759730</u>	Asset No: <u>0412253</u>
	EQPT Name: <u>Incubator</u>
	Manufacture: <u>Universal</u> Model: <u>Biochem</u>
	S. No. <u>1318</u> Biochemical Dept: <u>Blood Bank</u>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Not functioning

Action Taken : checked the machine and found that heating elements and thermostat are defective and internal chamber is fully rusted condition, machine is at not repairable condition.

Completed Date : 28/10/24 Time : 11:00 AM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Midhurya K</u>	<u>28/10/24</u>	<u>10:00 AM</u>	<u>11:00 AM</u>

Customer Remark Completed Pending

Equipment is not working condition.

Service Engineer Name : <u>Midhurya K</u>	Customer Name : <u>Redhikal. V 11)</u>
Signature: <u>[Signature]</u>	Signature: <u>[Signature]</u>
Date: <u>28/10/24</u>	Date: <u>28/10/24</u>
Contact Number: <u>9072522</u>	Contact Number: <u>28110204</u>
	Designation : <u>BBT</u>
	Hospital Seal: <u>BBT</u>



SUPERINTENDENT
W & C HOSPITAL, ALAPPUZHA

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: W&C HOSPITAL	Name of District: ALAPPUZHA
Name of Equipment: INCUBATOR	Equipment ID & Barcode: 139135 & 0412253
Make: UNIVERSAL BIOCHEMICALS	Model: BIOCHEM
Serial Number: 1318	Warranty details: NO WARRANTY
Date of purchase/ Year of manufacture/ Installation Date: 10/06/1997	Present status of the equipment: FULLY DAMAGED
Date of breakdown: 07/10/2024 (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: NO AMC/CAMC

Action taken:
CHECKED THE MACHINE AND FOUND THAT HEATING ELEMENT, THERMOSTAT ARE DEFECTIVE AND INTERNAL CHAMBER IS FULLY RUSTED CONDITION.

Recommendations for repair (required service details) :- NOT RECOMMENDED FOR REPAIR


Cost of spares (specify parts and cost): NA

Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA

Asset Value: 14625/-

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)
CYRIX SERVICE REPORT ATTACHED

Reasons for recommending the equipment as BER: CHECKED THE MACHINE AND FOUND THAT THE HEATING ELEMENT, THERMOSTAT ARE DEFECTIVE, AND THE INTERNAL CHAMBER IS FULLY RUSTED. THIS MACHINE WAS INSTALLED ON 10/06/1997 AND IS OVER 27 YEARS OLD. MACHINE IS FULLY RUSTED, AND WERE AND TEAR HAVE AFFECTED THE INTERNAL CHAMBER, TO IRREPARABLE EXTENT. THEREFORE WE RECOMMENDING THE EQUIPMENT AS BER.

Name & Signature of CYRIX Authority with date: MIDHUN RAJ R 

Remarks of Junior Consultant (Biomedical) NHM:

Recommended for BER (Yes/ No): YES

Date: 02.11.24


Signature of JC BM (NHM)

Date:
Medical Officer (i/c)

Seal


Signature of Superintendent /



SUPERINTENDENT
W & C HOSPITAL, ALAPPUZHA

Page 21

✓ 19 Bechem Incubator
Received
from the store
DHA 00/10/197
Thermometer Janyal 011

2 N.

2

Continued as per order no. B 206/13 dtd
25/5/13 and continued as per order no B
3046/14 W & L dtd. 21-7-2014

1 NOS

1 NOS

[Signature]



10/06/97

Stock entry verified Retrikkal V D

~~Patil A J~~
01/11/24
BB5



[Signature]
SUPERINTENDENT
W & C HOSPITAL, ALAPPUZHA



Rehikali AD *[Signature]*

BBT

