

PROFORMA

RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)  
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: PHC ATHIRAMPUZHA	Name of District: KOTTAYAM
Name of Equipment: Sterilizer	Equipment ID & Barcode: 139602 & 0541623
Make: Nill	Model: Autoclave
Serial Number: Nill	Warranty details: No Warranty
Date of purchase: Year of manufacture/ Installation Date: 17/06/2014	Present status of the equipment: Fully Damaged
Date of breakdown: 09/10/2024 (Toll Free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CMC

Action taken: Checked and found out the problem with top door , gasket, safety valve ,coil and power code need to replace those spare to further checking and working condition of the machine . Enquired the spare with Oem

Recommendations for repair (required service details) : Not recommending for repair.

Cost of spares (specify parts and cost): Gasket:649/-,Coil:1298/-,power code with connector :531/-,Steam valve :413/-,Top door : 3835/- TOTAL-6726/-

# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 76%      Asset Value:8820/-

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service report attached and vendor quotation attached

Reasons for recommending the equipment as RBER: Checked and found out the problem with top door , gasket, safety valve ,coil and power code need to replace those spare to further checking and working condition of the machine. Machine installed in 17/06/2014 and covered up to 10 +year. The repairing cost 76 %, both criteria met. So recommending the machine for condemnation as per the tender clause 5.3.14.1

Sanish Sivan *[Signature]* 25/10/2024  
Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM: *Asset value-8820 cost of spare-6726(76%)  
biller point 10 years - you informed that the spare cost is higher - the Auto-  
clave lid burst on 9/10/24 - and no damage to staff - The machine is old  
and cannot be usable.*

Recommended for BER (Yes/ No): *Yes*  
Date: *28/11/24*      *[Signature]* Signature of JC BM (NHM)

Date:  
Medical Officer (i/c)



*[Signature]* Signature of Superintendent  
MEDICAL OFFICER INCHARGE  
PRIMARY HEALTH CENTRE  
ATHIRAMPUZHA



**BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAMME  
UNDER  
NATIONAL HEALTH MISSION**



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

**CYRIX** (R)  
HEALTH CARE PVT LTD

No: 1007469

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>PHC</u>	Call Registration Date : <u>9-10-2024</u>
Address..... <u>PHC Athirampuzha</u>	Caller ID : ..... <u>139602</u>
..... <u>Icottayam</u>	Date of Visit : <u>17-10-2024</u>
Ph:..... <u>9695480922</u>	Asset No : ..... <u>0541623</u>
	EQPT Name : <u>sterilizer</u>
	Manufacture..... <u>nil</u> ..... Model : <u>autoclave</u>
	S. No..... <u>nil</u> ..... Dept..... <u>OP</u>

**Service Classification :** Breakdown call  PMS  Calibration  Cust. Training

**Problem Identified :** ..... sterilizer is popped up.....

**Action Taken :** checked and find out the sterilizer is popped up with blasting sound. may it happened due to improper usage. User can't use the equipment.

**Completed**  **Date :** 17-10-2024 **Time :** 12:30 pm **Spare Required**

**Spare Replaced**  **Requested**

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>AZIYA ASHRAF</u>	<u>17-10-2024</u>	<u>12:00 PM</u>	<u>12:30 PM</u>

**Customer Remark**  **Completed**  **Pending**

can't used. Mivi Joseph. Miv

**Service Engineer Name :** AZIYA ASHRAF

**Signature:** Aziya

**Date:** 17-10-2024

**Contact Number:** 9188046065

**Customer Name :** Mivi Joseph

**Signature:** Miv

**Date:** 17-10-24

**Contact Number:** 9847276040

**Designation :** NSQ. officer gr I.

**Hospital Seal**



*(Handwritten signature and stamp)*  
MEDICAL OFFICER INCHARGE  
PRIMARY HEALTH CENTRE  
ATHIRAMPUZHA

11/12/13 from Broke

Auto CLAKE

15/12/13 Condensed II NO on 100V  
order NO A-654/2013/MOPHCO/  
dt 15/12/13

(Auto Clack)

25/12/13  
De. from  
Voh...  
P.No. 35

12/4/14 Rd from store

Electric cooker type - B 84004  
H20

~~himu~~  
himu Mathew  
Nsg Officer GSI



Take

2

I

Cherian



**CLOUD99**

Office : 97444 73043

Service : 97444 73243

*Sales and service of all kinds of medical equipments*  
Room No: 22, First Floor, CSI Golden Jubilee Complex,  
Behind Shastri Road Waiting Shed, Kottayam-686001  
Email: cloud99biotech@gmail.com

**QUOTATION**

Ref No: C99B/0156/2023-24

Date: 22/10/2024

To,

PHC ATHIRAMPUZHA  
KOTTAYAM

Dear Sir

We are pleased to submit a quotation for PHC ATHIRAMPUZHA for Cooker Type Autoclave

SL No	ITEM NAME	PRICE PER UNIT	GST	PRICE INCLUDING GST
1	GASKET	550	12%	649
2	COIL 2000 W	1100	18%	1298
3	POWER CODE WITH COIL CONNECTOR	450	18%	531
4	STEAM RELEASE VALVE	350	18%	413
5	TOP DOOR	3250	18%	3835
	TOTAL			6726

**Terms & Conditions**

1. Taxes: GST @ 18 % inclusive
2. Payment Terms: 100% advance payment
3. Quotation validity: 15 days from date of quotation

We hope you will find our rates are competitive and wait for your valued order.

Thanking you,

Yours Faithfully

FOR CLOUD99 BIOTECH

AUTHORIZED SIGNATORY



We deal with sales and service of all kinds of hospital equipments, neonatal equipments, laboratory equipments, dental equipments, suction, hospital furniture, radiology equipments, physiotherapy equipments, all type of autoclaves, ICU equipments, etc.



*B*  
8W0  
FB002LMOL



PRIMARY HEALTH CENTRE  
PIN: 686 562  
ATHIRAMPUZHA

8ms JESSILMOL