


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

<b>Sl. No</b>	<b>Particulars</b>	<b>Details</b>
1	Name of District	THRISSUR
2	Name of Hospital	THQH PUTHUKKAD
3	Name of Equipment with Make, Model and Serial Number	Pulse Oximeter Make: CONTEC Model: CMS 5100 SN : 17090200084
4	Equipment ID & Barcode	#120876 – 0822184
5	Date of purchase / Year of manufacture /Installation Date	Ticket master installation date -18-11-2017 As per stock book installation date – 12-11-2018 (stock book attached for reference)
6	Warranty details (Yes/No)	Nil
7	*AMC/ CAMC Period agreed at the time of purchase	Nil
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	03-07-2024
9	Action taken	Checked the machine main board and spo2 module are complaint. Need to replace these spares for further checking and spare enquired to oem.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending
12	Cost of spares (specify parts and cost)	Main board – 12500/- Spo2 module board – 8500/- TOTAL- 21000/-

13	Asset Value	24700/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	85%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report, and oem quotation attached
16	Reasons for recommending the equipment as BER	Checked the machine and found main board and spo2 module are complaint. As per stock book Equipment installed on 12/11/2018 and aged up to 5 years 6 months. Spare cost is 85% of asset value and both criteria for RBER met. As per tender clause 5.3.14.1 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	BLIO T JOY 


\*Not mandatory

#Based on the period of life and value as per the BER guidelines

\* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Spare cost is 21,000/- Actual asset value is 24700/- only. Higher cost of spare.

  
 Consultant (Biomedical)  
 NHM, Tirissur

Signature of JC BM (NHM)



  
 സൂപ്രണ്ട്  
 മെഡിക്കൽ ഓഫീസർ (I/C)

Superintendent / Medical Officer (I/C)

Signature of seal

Date



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

**No. :**  
224517

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report** 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrix.in

<p><b>Health Facility</b> ..... THQH .....</p> <p><b>Address</b> ..... Puthukkad .....</p> <p><b>Ph :</b> ..... 9847643670 .....</p>	<p><b>Call Registration Date :</b> ..... 03-07-2024 .....</p> <p><b>Caller ID :</b> ..... 120876 .....</p> <p><b>Date of Visit :</b> ..... 04-07-2024 .....</p> <p><b>Asset No. :</b> ..... 0822184 .....</p> <p><b>EQPT Name :</b> ..... Pulse oximeter .....</p> <p><b>Manufacture</b> ..... Consec ..... <b>Model :</b> ..... CMS.5.100 .....</p> <p><b>S. No.</b> ..... 1709020084 ..... <b>Dept.</b> ..... Community .....</p>
--	---

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** ..... Spox Not sewing .....

**Action Taken :** ..... checked the machine main board and spox module esp complaint. need to replace these spox for further service .....

**Completed**  **Date :** ..... 04-07-2024 ..... **Time :** ..... 11:00 am ..... **Spare Required**

**Spare Replaced**  **Requested**

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Byo	04-7-2024	10:00 am	11:00 am

**Customer Remark** Completed  Pending

<p><b>Service Engineer Name :</b> Byo</p> <p><b>Signature :</b> </p> <p><b>Date :</b> 04/07/2024</p> <p><b>Contact Number :</b> 8921870010</p>	<p><b>Customer Name :</b> Alisha Raj</p> <p><b>Signature :</b> </p> <p><b>Date :</b> 4/7/24</p> <p><b>Contact Number :</b> [Handwritten]</p> <p><b>Designation :</b> Nursing office</p> <p><b>Hospital Seal :</b> </p>
--	--

**BCS Healthcare Pvt.Ltd**  
3136, IInd door, Mathewsons square  
MetroPiller No:464 NearChangapuzha park  
Cochin, Kerala - 682024,  
Email : info@bcshealthcare.in  
web: www.bcshealthcare.in



GST NO: 32AAKCB2766M1Z

DATE: 29/07/2024

Ref: BCS/96/24-25

**QUOTATION**

TO,

**CYRIX HEALTHCARE PVT LTD, THRISSUR**

SL.NO	PRODUCT NAME	QTY	UNIT RATE	TAX 12%	TOTAL
01	Pulse Oximeter Make : Contec Model: CMS 5100 SN :17090200084				
	1. Main Board	1	11,161.00	1,339.00	<b>12,500.00</b>
	2. Spo2 Module Board	1	7,589.00	911.00	<b>8,500.00</b>
<b>Grand Total</b>					<b>21,000.00</b>

**Terms & Conditions**

- Barcode - 0822184
- Caller Id - 120876
- Delivery Within 8 Days After Confirmation Of Purchase Order
- Payment 50% Advance Balance 50% After Delivery Time



**For, BCS HEALTH CARE PRIVATE LIMITED**

barcode - 0822184 caller id - 120876 thqh puthukad

7 messages

Thu, 25 Jul, 2024 at 6:22

Bijo T Joy <bijocyrix@gmail.com>  
To: info@bcshealthcare.in

dear sir,  
kindly provide the quotation for pulse oximeter's spares  
hospital - thqh puthukkad  
barcode - 0822184  
caller id - 120876  
equipment - pulse oximeter  
quotation for - spo2 module  
display  
main board

With Regards,  
BIJO T JOY

DISTRICT INCHARGE THRISSUR,



**CYRIX HEALTHCARE PRIVATE LIMITED**

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia  
M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

Mon, 29 Jul, 2024 at 11:20

BCS Healthcare <info@bcshealthcare.in>  
To: Bijo T Joy <bijocyrix@gmail.com>

[Quoted text hidden]



(Cms) 9mm

R.F.S Issue Name

Supl

12/1/18	R.F.S	2	2	<u>den</u>
---------	-------	---	---	------------

12/1/18	T & out to CLTY (Page 154)	1	1	<u>den</u>
---------	-------------------------------	---	---	------------

5/5/24	T & out to Substocks Registers page No 134 (Cms 5100)	1	N/A	<u>den</u>
--------	--	---	-----	------------

