


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	GH THRISSUR
3	Name of Equipment with Make, Model and Serial Number	VENTILATOR Make: GE Model: VERSAMED IVENT 201 SN: IV4061
4	Equipment ID/ Barcode	#123060- 0810725
5	Date of purchase/ Year of manufacture/Installation Date	30-05-2005
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free- 15-07-2024
9	Action taken	Checked the machine and found main board and o2 sensor are faulty. Need to replace these spares for further checking working condition of equipment. Spare enquired to OEM.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	400000/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Service report attached and OEM EOL attached
16	Reasons for recommending the equipment as BER	Checked the machine and found main board and o2 sensor are defective. Equipment installed on 30-05-2005 aged up to 19 years 2 months. This equipment has obsolete model OEM EOL letter attached. So as per tender clause 5.3.14.2 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	BIJO T JOY 


\*Not mandatory

#Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


*Very old and obsolete model*

  
 Signature of (Biomedical) NHM  
 Thrissur



Date:

**SUPERINTENDENT (Dy.DHS)  
 GENERAL HOSPITAL  
 THRISSUR - PIN: 690 001**

  
 Signature of  
 Superintendent/Medical Officer (i/c)



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

# CYRIX

HEALTHCARE PVT LTD

**No. :**  
224595

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report** 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>GH</u></p> <p>Address <u>THRISSUR</u> <u>Keralam.</u></p> <p>Ph : <u>8304960419</u></p>	<p>Call Registration Date : <u>15-7-2024</u></p> <p>Caller ID : <u>123060</u></p> <p>Date of Visit : <u>16-7-2024</u></p> <p>Asset No. : <u>0810725</u></p> <p>EQPT Name : <u>Ventilator</u></p> <p>Manufacture <u>CSE</u> Model : <u>Verismed</u> <u>1uat 201</u></p> <p>S. No. <u>104061</u> Dept. <u>OT</u></p>
---	--

**Service Classification :** Breakdown Call  PMS  Calibration  Cust. Training

**Problem Identified :** NOT working ventilator.

**Action Taken :** Checked the machine. and found main board and O2 sensor are faulty. need to replace these spares. For further checking work. Condition of equipment. Spare enquired to OEM.

Completed  Date : 16-7-24 Time : 2:00PM Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Absil</u>	<u>16-7-24</u>	<u>1:00PM</u>	<u>2:00PM</u>

**Customer Remark** Completed  Pending

**Service Engineer Name :** Absil Tony  
**Signature :**   
**Date :** 16-7-24.  
**Contact Number :** 9845328470

**SUPERINTENDENT (Dy. DHS)**  
**GENERAL HOSPITAL**  
**THRISSUR - PIN: 680 001**

**Customer Name :** \_\_\_\_\_  
**Signature :**   
**Date :** \_\_\_\_\_  
**Contact Number :** \_\_\_\_\_  
**Designation :** \_\_\_\_\_  
**Hospital Seal :** \_\_\_\_\_

Document Number: DOC2201514

To,

Chaitanya Sarawate  
Global Product Manager  
Anesthesia & Respiratory Care

GE Healthcare  
John F Welch Technology Center  
Odyssey, #122, EPIP Phase II  
Whitefield, Bangalore -560066, India

T +91 80 4088 2270  
M +91 974 010 4592  
[Chaitanya.sarawate@ge.com](mailto:Chaitanya.sarawate@ge.com)

December 4<sup>th</sup>, 2018

**Subject: Notice of End of Service Support from 30<sup>th</sup> June, 2019 for iVent 101, iVent 201, iVent 201E and other iVent product lines.**

Dear Valued Customer:

GE Healthcare's iVent product line has been a valued portfolio of devices in our ventilator portfolio. As the technology within our products change over time, we continually evaluate our ability to deliver high quality and timely service levels that you expect from GE Healthcare.


This letter is to inform you of GE Healthcare's intention to fully discontinue service, service contracts (parts, labor or comprehensive), and part support for the iVent product line as of 30<sup>th</sup> June, 2019 (End of Service Date). Until End of Service date, please continue to contact your local GE Healthcare service center for service support or spare parts.

This notice has no impact on your ability to continually use your product as per instructions in the product user manual beyond End of Service Date. If you have chosen to scrap iVent equipment or it is no longer in use, please do inform us so we can update our records.


We have newer devices that provide additional benefits to you and your patients. For information on the potential trade towards purchase of our new technology, please contact your local GE Healthcare sales representative.

Thank you for your trust and confidence in our products and solutions. At GE Healthcare, it is our pledge to serve evolving customer needs and to build intelligent devices that not only meet, but also exceed your expectations. Thank you for being a valued customer of GE Healthcare.

Sincerely,



Chaitanya Sarawate  
Global Product Manager



Prem Jayaraman  
Global Service Product Manager



## Need EOL letter

5 messages

Anjitha T L <purchasecyrix2@gmail.com>

Wed, 31 Jul, 2024 at 2:03 pm

To: Md, Sooryavarma (GE HealthCare, consultant) <sooryavarma.md@ge.com>, sooryavarma.md1@gehealthcare.com

Cc: Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Bijo T Joy <bijocyrix@gmail.com>

Dear sir,

Kindly provide the EOL letter for the ventilator machine.

Hospital - GH Thrissur  
Equipment - ventilator  
make - GE  
Model - VersaMed iVent 201  
SN - IV4061

--

Thanks & Regards

Anjitha T L  
Sr. Purchase Executive  
PH:- 9947719111  
[purchasecyrix2@gmail.com](mailto:purchasecyrix2@gmail.com)  
[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

**Cyrix Healthcare Pvt Ltd**  
| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India  
KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

MD, Sooryavarma <Sooryavarma.MD1@gehealthcare.com>

Wed, 31 Jul, 2024 at 2:15 pm

To: Anjitha T L <purchasecyrix2@gmail.com>, Md, Sooryavarma (GE HealthCare, consultant) <sooryavarma.md@ge.com>, P A, Anver

<Anver.PA@gehealthcare.com>, J, Meethu <meethu.j@gehealthcare.com>

Cc: Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Bijo T Joy <bijocyrix@gmail.com>

@P A, Anver/@J, Meethu, Pls support us with the EOL letter.

Regards,

**Sooryavarma MD**

**Service Sales Manager**

**Wipro GE Healthcare Pvt. Ltd.**

**Cochin, Kerala.**

**PH: +91-9497152800**



From: Anjitha T L <purchasecyrix2@gmail.com>

Sent: Wednesday, July 31, 2024 2:03 PM

**To:** Md, Sooryavarma (GE HealthCare, consultant) <sooryavarma.md@ge.com>; MD, Sooryavarma <Sooryavarma.MD1@gehealthcare.com>  
**Cc:** Cyrix Purchase Mgr. <purchasemgr@cyrix.in>; Assistant Manager Purchase <ampurchase@cyrix.in>; Bijo T Joy <bijocyrix@gmail.com>  
**Subject:** Need EOL letter

**WARNING:** This email originated from outside of GE HealthCare. Please validate the sender's email address before clicking on links or attachments as they may not be safe.

[Quoted text hidden]

---

**Anjitha T L** <purchasocyrix2@gmail.com> Wed, 14 Aug, 2024 at 9:45 am  
**To:** MD, Sooryavarma <Sooryavarma.MD1@gehealthcare.com>  
**Cc:** Md, Sooryavarma (GE HealthCare, consultant) <sooryavarma.md@ge.com>, P A, Anver <Anver.PA@gehealthcare.com>, J, Meethu <meethu.j@gehealthcare.com>, Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Bijo T Joy <bijocyrix@gmail.com>

Dear sir,

Please update the status of the EOL Letter.

[Quoted text hidden]

---

**Anjitha T L** <purchasocyrix2@gmail.com> Tue, 20 Aug, 2024 at 9:35 am  
**To:** MD, Sooryavarma <Sooryavarma.MD1@gehealthcare.com>  
**Cc:** Md, Sooryavarma (GE HealthCare, consultant) <sooryavarma.md@ge.com>, P A, Anver <Anver.PA@gehealthcare.com>, J, Meethu <meethu.j@gehealthcare.com>, Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Bijo T Joy <bijocyrix@gmail.com>

Dear Sir,

As per the last discussion, please share mail regarding the EOL.

[Quoted text hidden]

---

**MD, Sooryavarma** <Sooryavarma.MD1@gehealthcare.com> Tue, 20 Aug, 2024 at 1:12 pm  
**To:** Anjitha T L <purchasocyrix2@gmail.com>  
**Cc:** Md, Sooryavarma (GE HealthCare, consultant) <sooryavarma.md@ge.com>, P A, Anver <Anver.PA@gehealthcare.com>, J, Meethu <meethu.j@gehealthcare.com>, Cyrix Purchase Mgr. <purchasemgr@cyrix.in>, Assistant Manager Purchase <ampurchase@cyrix.in>, Bijo T Joy <bijocyrix@gmail.com>

Dear Anjitha,

Pls find the attached EOSL communication regarding the below mentioned model.

[Quoted text hidden]

# J.C.U. VENTILATOR

Rcd    Issued    55 Pulce    Qty.

~~30/5/22 Received from store~~

15

1

Store page  
No. 206

Rs. 5,84,000.

MILI ELECTRONICS. PVT. LTD

H - 7/3, 4th Avenue

JINDIRANAGAR. CHENNAI.

PIN. 600020.

PHONE. 044 2445518/1

24401554.

FAX. 044 - 21,125654.

E MAIL medmili@vsnl.net.

Model No. VERSAMID IVENT

201.

I.V. 4061

Items.

(1) J.C.U. Ventilator (1)

Patent Circuits - (3)

Advt. & pack

nebulizer - (1)

Humidifier - (1)

O<sub>2</sub> pressure reg. - (1)

Hose for O<sub>2</sub> Connection

Smt - (1)

Test lung - (2)

Hose for Compressor air

Smt - (1)

WARRANTY CARD. (2 yrs)

A.M.C. 5 yrs.

20/12/22

Tr. out to unservisable Register

Page No. 226

1

NIL

1

Service Engineer

Balachandran Kumar

Cochin

NO: 9747590147

Balachandran - 9287522322

