


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KASARGOD
2	Name of Hospital	TH MANGALPADY
3	Name of Equipment with Make, Model and Serial Number	Name: - UPS Make: - APC Model: -SRC2000 SN:- B21631000536
4	Equipment ID & Barcode	#126503- 1450171
5	Date of purchase / Year of manufacture /Installation Date	05/12/2016
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	30/07/2024 (Toll free)
9	Action taken	Checked the machine and found that main board is defective. Enquired OEM for quotation.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No Recommendation
12	Cost of spares (specify parts and cost)	SRC2KUXI MAIN PCB ASSY: 18000/-

13	Asset Value	11000/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	163.63%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report and OEM quotation attached.
16	Reasons for recommending the equipment as BER	Checked the machine, found machine main board defective. Machine installed on 05-12-2016 and aged up to 7 years and 7 months. Spare cost of equipment is 163.63% with respect to asset value. Both criteria for RBER met. As per the tender clause 5.3.14.1, We recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	AJITH BABU K 


*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Checked the machine and found main board defective. Equipment installed on 07/05/2019 (Received on 05/12/2016), Stock book details attached). Spare cost is 163.63% of asset value. cyrix service report and OEM quotation attached. so recommending the equipment for BER.

Signature of JC


JUNIOR CONSULTANT
BIOMEDICAL ENGINEERING
Arogyakeralam (NHM)
Kasaragod Mangalpad

Date




SUPERINTENDENT
TALUK HEAD QUARTERS HOSPITAL
MANGALPADY
P.O. UPPALA, KASARAGOD - 671 322
Superintendent / Medical Officer (i/c)



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 208630

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>THH Mangalpadu</u></p> <p>Address <u>Kasaragod</u> <u>code</u></p> <p>Ph : <u>9995209434</u></p>	<p>Call Registration Date : <u>30-7-24</u></p> <p>Caller ID : <u>126503</u></p> <p>Date of Visit : <u>2-8-24</u></p> <p>Asset No. : <u>1450171</u></p> <p>EQPT Name : <u>UPS SKVA</u></p> <p>Manufacture <u>APC</u> Model : <u>SRC 2000</u></p> <p>S. No. <u>B21631000536</u> Dept. <u>LAB</u></p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : not switching on

Action Taken : checked the machine and found problem with the main pcb, need to replace main pcb and checked working condition of the equipment

Completed Date : 2-8-24 Time : 11:00AM Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Ajith Babu K</u>	<u>2-8-24</u>	<u>10:00AM</u>	<u>11:00AM</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Ajith Babu K</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>2-8-24</u></p> <p>Contact Number : <u>8848079288</u></p>	<p>Customer Name : <u>Binyan</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>28/24</u></p> <p>Contact Number : <u>9995209434</u></p> <p>Designation : <u>Lab-Technician</u></p> <p>Hospital Seal :</p>
--	---





APS MARKETING COMPANY

27/233A, GROUND FLOOR, APS ARCADE
NETHAJI ROAD, PUTHIYARA, CALICUT - 673004
PH: 0495 - 2741313, Mob: 9447184933, Mail: info@apsmarketing.in

QUOTATION

Quotation No : APS-1754
Date : 05-08-2024
Valid Upto : 20-08-2024

Customer :
CYRIX HEALTHCARE PVT LTD
1ST FLOOR,30641B,KOCHI DANUSHKODI ROAD,ERNAKULAM

#	DESCRIPTION OF ITEM	QTY	RATE	TAXABLE	GST %	TOTAL
1	SRC2KUXI MAIN PCB ASSY APC SRC2KUXI	1	15254.24	15254.24	18.00	18000.00
					G-Total	18000.00

Amount Chargeable (in words)
Eighteen Thousand Rupees only

E & O.E

Remarks

FOR APS MARKETING COMPANY

Authorised Signatory



Jayesh Krishnan <jayeshcyrix@gmail.com>

REQUIRED SPARE QUOTATION FOR UPS TH MANGALPADY 1450171

2 messages

Jayesh Krishnan <jayeshcyrix@gmail.com>

Tue, Sep 3, 2024 at 10:09 AM

To: lanmark systems <lanmarkservices@gmail.com>

Cc: Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear sir,

Kindly share the price and availability of main board for UPS.

Make : apc

Model: src 2000

Sl.no: B21631000536

--

Thanks & Regards

JAYESH KRISHNAN,
Sr. Biomedical Engineer,
District In-charge
Kasargod.
PH:- 7593847156

jayeshcyrix@gmail.comwww.cyrixhealthcare.com**Cyrix Healthcare Pvt Ltd****| 30/S641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India****KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH****lanmark systems** <lanmarkservices@gmail.com>

Tue, Sep 3, 2024 at 10:43 AM

To: Jayesh Krishnan <jayeshcyrix@gmail.com>

Cc: Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Hi

quotation attached

THANKS & REGARDS**APS MARKETING COMPANY
PUTHIYARA, KOZHIKODE**


[Quoted text hidden]

**CYRIX SRC2KUXI-PCB-QUOTATION.pdf**

124K



for
SUPERINTENDENT
TALUK HEAD QUARTERS HOSPITAL
MANGALPADY
P.O. UPPALA, KASARAGOD - 671 322



STOCK BOOK OF.....

27

Name of article... FULLY AUTO ANALYSER EM-360

Date	No. and date of voucher or invoice	From whom received or to whom issued	Receipt	Issued	Balance after each transaction	Initials of receiver	Remarks
		Brought forward ..					
07/05/19	(11/16)	fully auto analyser EM-360		1			
		Invoice No/Date : 17360472 / 31/12/16					
		Sales order No/Date : 8003733 / 12/12/16					
		Sr.No: 60660					
		(Rs: 10,60,810.48/-)					
		System pack Dema 1st					
		(Lot No: 0061628 EXP - 31/Aug/17)					
		UPS 2 KVA					
		Sr.No: B21631000536					
		Installed on 03/04/19					

