





**PROFORMA  
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)  
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: CHC NEMMARA	Name of District: PALAKKAD
Name of Equipment: NEBULIZER	Equipment ID & Barcode: #144381 - 0951635
Make: OMRON	Model: NEC-C106
Serial Number: NA	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: In ticket master Installation date - 23/01/2024, As per stock 10/03/2020 stock book attached for reference.	Present status of the equipment: Fully damaged
Date of breakdown: 02/11/2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked the nebulizer found motor stuck. Found Problem with motor and piston, need to be replace these spares for further checking of machine.	
Recommendations for repair (required service details): No recommendations	
Cost of spares (specify parts and cost): NA	
#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: In ticket master asset value = Nil. As per stock book asset value 1562.65/- Reference Asset value taken form Chc Nemmara asset master 1353/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix Service report attached.	
Reasons for recommending the equipment as BER: Checked the nebulizer found motor stuck. Found Problem with motor and piston, need to be replace these spares for further checking of machine. In ticket master installation date 23/01/2024 As per the stock book Equipment installed on 10/03/2020 aged up to 4 years 7 months. Quotation not submitted since spare are not available from market. So we are recommended the equipment for condemnation	
AKHILESH P  13/11/24 Name & Signature of CYRIX Authority with date	
Remarks of Junior Consultant (Biomedical) NHM: Installation date of is 10/03/2020 aged up to 4 years 7 months. Quotation not attached.	
Recommended for BER (Yes/ No): Yes	 13/11/24 Signature of JC BM (NHM)
Date: 13/11/24	
Date: 13-11-24	 SUPERINTENDENT Community Health Centre Nemmara, Palakkad. Signature of Superintendent/ Medical Officer (i/c)



\*Not mandatory

\*\*Based on the period of life and value as per the BER guidelines

\* Attach Photograph





# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

No : 126504

# CYRIX®

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility: <u>LHC Nemmara</u> Address: <u>Palakkad</u> <u>Kerala</u> Ph: <u>09496991228</u>	Call Registration Date: <u>2-11-2024</u> Caller ID: <u>144381</u> Date of Visit: <u>4-11-2024</u> Asset No: <u>0951635</u> EQPT Name: <u>nebuliser</u> Manufacture: <u>omron</u> Model: <u>NE-C106</u> S. No: <u>202306024744K</u> Dept: <u>Casualty</u>
--	--

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : unit not working

Action Taken : checked the nebulizer found motor stuck. Found problem with motor and piston, need to be replace these spares for further checking of machine

Completed  Date : 4-11-2024 Time : 11:40 AM Spare Required


Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>ABIJITH B</u>	<u>4-11-2024</u>	<u>11:05 AM</u>	<u>11:40 AM</u>

Customer Remark Completed  Pending

Service Engineer Name : ABIJITH B Signature: [Signature]  
 Date: 4-11-2024 Date: 4.11.24  
 Contact Number: 9567164651 Contact Number: 9496991228  
 Designation: SNO







# NEBULISER.



		Rd	Used in ward	Bal. In Sub.	Balanc in Main stock
	B.F Old book page no.				
9.6.18 11/9/2014	Rd from store (L.P. page 35) 18301	2	2	-	2
10.3.2020	Rd from store Nil compressing nebuliser from Kamanya Medical Department R. 1562.65	1	1	Nil	3
10/2/2021	Drain from S.C.C Nemmara P.N. 57	2			5
5.1.22	Rd from store	2			7
6.1.22	Issued to ward		7	Nil	7
23.4.23	Tx out to inservice page no. 29		2 (over)		5
10/12/23	Tx out to inservice		2 (over)		3
23/1/24	Rd from the store (L.P)	2			5
24/1/24	Went to ward for use Rate R. (with, Rev, main) (1686.72 x 2 = 3373.44/-)		5	Nil	5



  
 SUPERINTENDENT  
 Community Health Centre  
 Nemmara, Palakkad.

  
 SWS





*Handwritten signature*

SUPERINTENDENT  
Community Health Centre  
Nemmara, Palakkad.