

PROFORMA

RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: CHC THOTTACKAD	Name of District: KOTTAYAM
Name of Equipment: Biochemistry Analyzer	Equipment ID & Barcode: 143940 & 0530231
Make: Robonik	Model: Prietest Touch
Serial Number: AT2330814RBK	Warranty details: No warranty
Date of purchase/ Year of manufacture/ Installation Date: 04-05-2015	Present status of the equipment: fully damaged
Date of breakdown: 31/10/2024 (Toll free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CMC

Action taken: Checked and found out the problem with primary board, display unit and filter defective. Enquired the spare with oem

Recommendations for repair (required service details) : Not recommending for repair.

Cost of spares (specify parts and cost): NA

# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA <i>EOL</i>	Asset Value: 67998 /-
---	-----------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not) : Cyrix service report and end of life letter attached

Reasons for recommending the equipment as RBER : Checked and found out the problem with primary board ,display unit and filter defective Machine installed in 04-05-2015 and covered up to 9 + years. Oem declared end of life for this machine So recommending the machine for condemnation as per the tender clause 5.3.14.2

Arun G C
Name & Signature of *CYRIX* Authority with date *19/11/2024*

Remarks of Junior Consultant (Biomedical) NHM: *Asset value 67998, cost of spare - End of life letter period - 9 years - The OEM submitted End of life certificate to the machine and (letter dated 5/11/24)*

Recommended for BER (Yes/ No): *Yes*

Date: _____

[Signature]
Signature of JC BM (NHM)
KOTTAYAM

Date: _____ Seal _____
Medical Officer (i/c) _____ Signature of Superintendent /



മെഡിക്കൽ ഓഫീസർമാർക്ക് മാത്രം
സാമൂഹിക ആരോഗ്യകേന്ദ്രം
തോട്ടക്കാട് - 686 539



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



No : 1007395

SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX[®]
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>CHC</u>	Call Registration Date : <u>31-10-2024</u>
Address..... <u>CHC Thottackad</u>	Caller ID : <u>143940</u>
..... <u>Kottayam</u>	Date of Visit : <u>5-11-2024</u>
Ph : <u>7356084838</u>	Asset No : <u>0530231</u>
	EQPT Name : <u>Biochemistry Analyser</u>
	Manufacture: <u>Robonik</u> Model: <u>protest Touch</u>
	S. No. <u>A.T.2.3308.1.4.R.B.</u> Dept..... <u>Lab</u>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Machine is not working

Action Taken : checked and found out the problem with primary board, display and filter defective. Enquired the spare with OEM.

Completed Date : 5-11-2024 Time : 12:00 P.m Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Arun G C</u>	<u>5-11-2024</u>	<u>11:00 Am</u>	<u>12:00 pm</u>

Customer Remark Completed Pending

Service Engineer Name : Arun G
Signature: Arun
Date: 5-11-2024
Contact Number: 8921992272



Customer Name: Advance - S Sij
Signature: [Signature]
Date: 5/11/24
Phone Number: 9495452655
Designation: Lab Technician
Hospital Seal

GOVERNMENT HOSPITAL, THOTTAKKAD
STOCK REGISTER OF MEDICINES, EQUIPMENTS ETC.

32

Name of Items: **AUTOMATIC BIOCHEMISTRY ANALYSER**

Date	Details of Transaction	Receipt	Issue	Balance	Remarks	Initials
4-5-15	Received from Store	1 Nos		1 Nos		AS
	Invoice no 099 Robonic India Pvt Ltd Dated 6-9-14 Model - Prietest touch Make - Robonic Sr no AT 2330810 RBK					
				Rs 67998		
5-5-15			1 Nos (In use)			
4-5-15	Received from Store	1 Nos		2 Nos		AV
	Invoice no 164 Robonic India Pvt Ltd Dated 13-8-13 Model : Prietest touch Sr no AT 0260713 RBK					
				Rs 84521		
	Warranty Expires on			28-8-16		
	Service engineer Sijith M			-9846463658		
	Manager Ranesh			9745810449		



Capt. Kdtty-61/15

ctd page no 33

Date: 08.11.2024

Dear Valued Customer,

Thank you for choosing **Robonik (India) Pvt. Ltd.** We appreciate the opportunity to support our business. This letter is to inform you regarding the EOL (End of Life) of the instrument Semi Automated Biochemistry Analyser (Model: Prietest Touch) having **Serial No. AT2330814RBK.**

As required, we stated herewith that the said instrument Semi Automated Biochemistry Analyser (Model : Prietest Touch) having **Serial No. AT2330814RBK** as **End of Life (EOL) Notified.** (During from installation is 10 years).

Thank you in advance for your understanding and your cooperation. If you have any questions, please contact our Customer Care Department at Toll Free No. 18005727977.

Yours faithfully,

For **ROBONIK (INDIA) PVT. LTD.**



Authorised Signatory



ടെസ്റ്റിംഗ് യൂണിറ്റിന്റെ ചാർജ്ജ്
 മാനിറ്ററിംഗ് യൂണിറ്റിന്മേൽ
 നോട്ടീഫിക്കേഷൻ - 686 530