

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: PHC NANNABRA	Name of District: MALAPPURAM
Name of Equipment: DIGITAL WEIGHING MACHINE	Equipment ID & Barcode: #145677 -1051901
Make: SKNOL	Model: NA
Serial Number: NA	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: In ticket master installation date is 12/02/2018, As per stock book installation date is 03/03/2023	Present status of the equipment: Fully damaged
Date of breakdown: 08/11/2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC

Action taken: Checked the machine and found Display, Mainboard & Load cell defective. Need to replace these spares for further checking.


Recommendations for repair (required service details): No Recommendation

Cost of spares (specify parts and cost): N/A


#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: N/A	Asset Value: In ticket master asset value Nil, Reference asset value taken from asset master is 5000/-
--	--

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix Service report attached

Reasons for recommending the equipment as BER: Checked the machine and found Display, main board & Load cell defective. In ticket master installation date is 12/02/2018, As per stock book installed on 03/03/2023. Aged up to 1 year and 8 months, Quotation not submitted since the spares are not available in the market. So, we recommending the equipment for condemnation.

Gokul KT  16/11/2024
Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:
checked with the damage of main board, display & load cell spares not available. Beyond repair.

Recommended for BER (Yes/ No):	 Signature of JC BM (NHM)
Date: 16/11/24	

Dr. Aswathi Sankar
19/11/24
For mo

Jr. Consultant Biomedical Engineer
Malappuram
Kodim Malappuram

Date: _____	Seal	Signature of Superintendent / Medical Officer (i/c)
-------------	------	---



*Not mandatory Based on the period of life and value as per the BER guidelines * Attach Photograph
F.H.C Nannabra
P.O. Kodinhi - 676 411



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1023280

CYRIX®

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrlix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>PHC</u> Address..... <u>Nannobra</u> <u>Malappuram</u> Ph:..... <u>9207112247</u>	Call Registration Date : <u>08/11/2024</u> Caller ID : <u>145677</u> Date of Visit : <u>11/11/2024</u> Asset No : <u>1051901</u> EQPT Name : <u>Digital Weighing Machine</u> Manufacture <u>Skol</u> Model : <u>NA</u> S. No. <u>NA</u> Dept. <u>OP</u>
--	---

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : machine not working

Action Taken : Checked the machine and found Display & Main Board & Load cell defective - need to replace these spares for further checking

Completed Date : 11/11/2024 Time : 12.10.00pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrlix Engineer <u>A. 201</u>	Date <u>11/11/24</u>	Start Time <u>12:10pm</u>	End Time <u>12:00pm</u>
-------------------------------	----------------------	---------------------------	-------------------------

Customer Remark to with Dr. Aswathi Sankar 19/11/24 for mu Completed Pending



Service Engineer Name : A. 201
 Signature: [Signature]
 Date: 11/11/24
 Contact Number: 9567362845

Customer Name : A. K. P
 Signature: [Signature]
 Date: 11/11/24
 Contact Number: 9207112247
 Designation : Nursing Officer
 Hospital Seal

Medical Office
F.H.C Nannambra
P.O. Kodinhi - 676 209



Dr. Aswathi Smt



Medical Officer -
F.H.C Nannambra
P.O. Kodinhi - 676 309

Dr. Aswathi
Smt
19/11/24
B r mo

