

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: CHC MELATTUR	Name of District: MALAPPURAM
Name of Equipment: ECG MACHINE	Equipment ID & Barcode: #147489- 1031643
Make: CONTEC	Model: ECG100G
Serial Number: 06AE001394	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: In ticket master installation date is Nil, As per stock book installation date is 14/08/2012	Present status of the equipment: Fully damaged
Date of breakdown: 09/11/2024	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Date of registration of complaint through email/Toll free): Toll free	
Action taken: Checked the machine and found Printer module and battery defective. Need to replace these spares for further checking.	

Recommendations for repair (required service details): No Recommendation

Cost of spares (specify parts and cost): Printer module: 9770/-
Battery: 2430/-
GST 18%:2196/-
TOTAL: 14396/-

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 67% Asset Value: 21200/-

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service report attached & Oem Quotation attached.

Reasons for recommending the equipment as BER: Checked the machine and found Printer module, Battery defective. In ticket master installation date is Nil, As per stock book installation date is 14/08/2012. Aged up to 12 years and 2 months, repairing cost is 67% of asset value, both criteria for RBER met. As per the tender clause 5.3.14.1 We recommending the equipment the equipment for condemnation.

23/11/24 GOKULK.T #
Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical), NHM:
Verified with the damage of printer module & Battery. High cost for spare. Beyond economic repair.

Recommended for BER (Yes/ No):

Date: 23/11/24

* Consultant Signature of JCBM (NHM)
NHM Malappuram

Date: 26/11/24

For MO
De Jyothi.cue
26/11/24
മുഖ്യ ഓഫീസർ
മാതൃകാശാസ്ത്ര കേന്ദ്രം
മലപ്പുറം

AROGYAKERALAM
NATIONAL HEALTH MISSION
MALAPPURAM-676

Signature of Superintendent / Medical Officer (i/c)

*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1023285

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>C.H.C</u>	Call Registration Date : ... <u>18/11/24</u>
Address..... <u>Melattur</u>	Caller ID : <u>147489</u>
..... <u>Malappuram</u>	Date of Visit : <u>20/11/24</u>
Ph : <u>9061999721</u>	Asset No : ... <u>1031643</u>
	EQPT Name : ... <u>ECG Machine</u>
	Manufacture <u>Contec</u> Model : <u>ECG 1600</u>
	S. No... <u>1394</u> Dept... <u>OP</u>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Printing issue.....

Action Taken : Checked the machine and found the printer module battery defective. Need to replace these spare parts. Further checking......

Completed Date 20/11/24..... Time : 12:40 PM..... Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer Abdul Rehman Date 20/11/24 Start Time 11:30 AM End Time 12:40 PM

Customer Remark മെഡിക്കൽ ഓഫീസർ സാമൂഹികാരോഗ്യ കേന്ദ്രം മലപ്പുറം. പിൻ: 679999 Completed Pending

Service Engineer Name : Abdul Rehman Customer Name : Remanick
 Signature: [Signature] Signature: [Signature]
 Date: 20/11/24 Date: 20/11/24
 Contact Number: 9567362845 Contact Number: 794...
 Designation : ENO
 Hospital Seal



Aspen Diagnostics Pvt. Ltd.

(Rapid Diagnostic Group Of Companies)



ISO 9001:2015
CERTIFIED COMPANY

Proforma Invoice No.: ADPL/S/RD/20241121/24-25			Date : 21-11-2024		
Cyrix Healthcare Pvt. Ltd. Malappuram			Ref. : Service Engineer Date : 21-11-2024 CHC Malattur		
Sr. No	Description	Part No.	Price Each	Qty	Total (In Rs.)
<u>Instrument Model: ECG Single Channel</u> (Model: ECG-3 Plus) (S/No.061394)					
A	Spares Procurement Details:				
(I)	Printer Module		9,770.00	1	9,770.00
(II)	Battery		2,430.00	1	2,430.00
Total					12,200.00
(+) GST @ 18%					2,196.00
Total: Fourteen Thousand Three Hundred Ninety Six Only.					14,396.00

Note: Visit Charge will be extra

This quote is only an indicative estimate, based on the initial inspection.
However if any part replaced during service , will be billed extra at actuals.

Terms & Conditions :

{Warranty is not applicable for spare parts } {Delivery : 3-4 Weeks} {Payment Terms : 100% Advance }

Bank Account Details :- @Mup

Name of the Firm : Aspen Diagnostic Pvt. Ltd.	Aspen Diagnostics Pvt. Ltd. Rajaram Krishnamurthy National Service Manager
Bank Name : ICICI Bank Limited	
Account No. : 016005003239	
Account Type : Current Account	
MICR Code : 110229029	
IFSC Code : ICIC0000160	

Rajaram Krishnamurthy
 National Service Manager
 21/11/24



PRICE ENQUIRY 1031643 ECG MACHINE @ CHC MELATTUR -MALAPURAM (Ticket id 147489)

2 messages

Wed, 20 Nov, 2024 at 10:03 pm

Gokul K T <gokulcyrix@gmail.com>
To: service support <servicesupport@rdgc.in>
Cc: Blesson Jose <Zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear sir,

There is a complaint reported from CHC MELATTUR - MALAPURAM, Regarding the complaint with ECG MACHINE make:Contec, model:Aspen ECG100G, serial Number - 061394 barcode number (1031643). Required price quotation for following spares, kindly share the estimate.

- printer module
- battery

GOKUL KT
CYIX HEALTHACARE
DISTRICT INCHARGE
MALAPPURAM
9809314280

service support <servicesupport@rdgc.in>
To: Gokul K T <gokulcyrix@gmail.com>
Cc: Blesson Jose <Zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Thu, 21 Nov, 2024 at 1:18 pm

Dear sir,

Please find the attached P.I. & share the P.O. along with payment details for billing.

Thanks & Regards

Hari Ram

[Quoted text hidden]



*Dr. JAWI-C.H.
26/11/24*

മെഡിക്കൽ ഓഫീസർ
സാമൂഹികാരോഗ്യ കേന്ദ്രം
മലപ്പുറം
മലപ്പുറം, പിൻ: 679 326

