

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: THQH BALUSSERY	Name of District: KOZHIKODE
Name of Equipment: BP APPARATUS	Equipment ID & Barcode: 148051 & 1121881
Make: BIO PLUS	Model: NA
Serial Number: NA	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: In ticket master installation date is 06/03/2018, As per stock book installation date is 25/11/2005	Present status of the equipment: Fully damaged
Date of breakdown: 21/11/2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked and found mercury insufficient so need to refill mercury for further checking of working condition of the bp apparatus	

Recommendations for repair (required service details): Not recommending for repair

Cost of spares (specify parts and cost): 4 ML MERCURY=935/-
GST 18%=168.30/-
TOTAL=1103.30/-

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 83.39%	Asset Value: 1323/-
---	---------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service report attached and vendor quotation attached

Reasons for recommending the equipment as BER: Checked and found mercury insufficient so need to refill mercury for further checking of working condition of the bp apparatus. In ticket master installation date is 06/03/2018, as per stock book installation date is 25/11/2005 (stock book attached for reference). Equipment out lived 19 years. Spare cost is 83.39% with respect to asset value, both criteria for RBER met, As per tender clause 5.3.14.1 we recommending the equipment for condemnation.

JINEESH PP
Name & Signature of CYRIX Authority with date *30/11/24*

Remarks of Junior Consultant (Biomedical) NHM:
checked and found that mercury insufficient. Machine installed on 25/11/2005. So recommended for condemnation

Recommended for BER (Yes/ No): Yes	Consultant Bio Medical Engineer National Health Mission Kozhikode
Date: 30/11/2024	Signature of JC BM (NHM)

Date: 3.12.24.	Signature of Superintendent / Medical Officer (i/c) THQH BALUSSERY
----------------	---

*Not mandatory

Based on the period of life and value as per the BER guidelines * Attach Photograph



മെഡിക്കൽ ഓഫീസർ
 നാലുക്ക് കളക്ടറുടെ ഓഫീസിൽ
 കൊച്ചി, പിൻ - 673612

MEDPLUS BIOMEDICALS



Athulraj M
Kumaraswamy raja road
Chelannur (P.O)
Kozhikode

Quotation

☎ 8848372192 ✉ medplusbiomedicals.clt@gmail.com
GSTIN 32CEYPA2140F1ZV

To,
Cyrix Healthcare
Manager
Kochi

Quotation# Quote-122

Date: 29/11/2024

Dear Sir/Mam,

Thank you for your valuable inquiry. We are pleased to quote as below

#	DESCRIPTION	QTY	PRICE	GST	TOTAL
1	4ML MERCURY TUBE	1	₹ 935.00	₹ 168.30 18.0%	₹ 1,103.30
SUB TOTAL					₹ 935.00
GST					₹ 168.30
GRAND TOTAL					₹ 1,103.30

We hope you find our offer to be in line with your requirement.

Payment Instructions

Account name : MEDPLUS
BIOMEDICALS

Acc no : 07791012000500

IFSC CODE : PUNB0077910

Bank name : PUNJAB NATIONAL
BANK

Branch : ERANJIPALAM .

For, MEDPLUS BIOMEDICALS

AUTHORIZED SIGNATURE

Terms & Conditions:

- Quotation validity - 6 months
- 100 % advance



Required quotation - Mercury for BP Apparatus

2 messages

Jineesh P P <jineeshcyrix@gmail.com>

Fri, 29 Nov, 2024 at 11:56 am

To: Athulraj m <medplusbiomedicals.clt@gmail.com>

Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

Dear Sir,

Please share price and availability of mercury for bp apparatus.

Mercury - 4ml

Thanks & Regards,

Jineesh PP
District Incharge - Kozhikode
KLBEMMP

Mob-7593847163

jineeshcyrix@gmail.com

www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd

| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India

KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH


Athulraj m <medplusbiomedicals.clt@gmail.com>

Fri, 29 Nov, 2024 at 3:26 pm

To: Jineesh P P <jineeshcyrix@gmail.com>

Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

[Quoted text hidden]



P. V. Manjula
 സാമ്പത്തിക വിഭാഗം
 താലൂക്ക് ഓഫീസ്, പാലക്കാട്, പിൻ - 673612