

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: CHC OLLUR	Name of District: Thrissur
Name of Equipment: oxygen concentrator	Equipment ID & Barcode: #147209- 0832438
Make: ACCUNEB JL	Model: ZY-10ZW
Serial Number: N/A	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: Ticket master installation date is Nil Stock book installation date - 22-02-2018	Present status of the equipment: Fully damaged
Date of breakdown: 16-11-2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked the machine and found that compressor motor, sieve bed and main board are defective, need to replace these spares for further checking.	


Recommendations for repair (required service details): Not Recommending

Cost of spares (specify parts and cost): NA

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 26500 /-
---	-----------------------


Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix report attached.

Reasons for recommending the equipment as BER: Upon inspecting the machine, it was discovered that the compressor motor, sieve bed, and main board are faulty. These spare parts need replacement to assess the equipment's operational status. As per hospital stock ledger equipment received on 22-02-2018. Unfortunately, the supplier information is unavailable from end user as the equipment was donated without manufacturing details. Due to the lack of spare part information, we advise condemning the equipment as a special case (5.3.14.2).


BIJOY JOY
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:
Donated equipment. No data available

Recommended for BER (Yes/ No): <i>Yes</i>	 Signature of JC BM (NHM)
Date:	

Date:	 Seal Signature of Superintendent / Medical Officer (i/c)
-------	--

*Not mandatory
 #Based on the period of life and value as per the BER guidelines
 *Attach Photograph



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1008548

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <i>Ch. c. allas</i> Address..... <i>Thississu</i> Ph:..... <i>7012136901</i>	Call Registration Date :..... <i>16/11/2024</i> Caller ID :..... <i>147209</i> Date of Visit :..... <i>17/11/2024</i> Asset No :..... <i>0832438</i> EQPT Name :..... <i>Oxygen concentrator</i> Manufacture..... <i>Arcomed</i> Model :..... <i>24-102W</i> S. No..... <i>NA</i> Dept.....
---	---

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified :..... *oxygen concentrator not working*

Action Taken :..... *checked the machine and found that compressor maker, sieve bed and main board are defective, need to replace these spares for further checking*

Completed Date :..... *17/11/24* Time :..... *3:40pm* Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer <i>Suhas</i>	Date <i>17/11/24</i>	End Time <i>3:40pm</i>
-----------------------------	----------------------	------------------------

Customer Remark Completed



Service Engineer Name : *Suhas. p. Rajan*
 Signature: *[Signature]*
 Date: *17/11/24*
 Contact Number: *7593847155*

Customer Name: *George*
 Signature: *[Signature]*
 Date: *17/11/24*
 Contact Number: *7012136901*
 Designation : *CHN*
 Hospital Seal



supply details. - 0832438-147209-

2 messages

BiJo T Joy <bijocyrix@gmail.com>

To: KMSCL <kmsclbmececentral1@gmail.com>, BIOMEDICAL WING THRISSUR <bmetsr2020@gmail.com>
Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

Tue, 26 Nov, 2024 at 8:33 am

Dear sir,
kindly provide below supply details of below mentioned oxygen concentrator.
hospital - CHC OLLUR
equipment - oxygen concentrator
make - NA
model- ZY-10ZW
barcode - 0832438
caller id - 147209
SN- na

With Regards,

BIJO T JOY

DISTRICT INCHARGE THRISSUR,

CYRIX HEATHCARE PRIVATE LIMITED

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia
M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

KMSCL <kmsclbmececentral1@gmail.com>

To: BiJo T Joy <bijocyrix@gmail.com>

Thu, 28 Nov, 2024 at 4:35 pm

Cc: BIOMEDICAL WING THRISSUR <bmetsr2020@gmail.com>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

The supply details for the above mentioned equipment were not available with us. Please refer to the hospital stock data for additional details.
[Quoted text hidden]

—

Regards,

SHALU U S

DISTRICT BIOMEDICAL ENGINEER

KMSCL

Contact No: +91 91889 10013

Oxygen Concentrator 5Ltr -

25
Remarks

		Quantity Received	Quantity Issued	Balance	rate	Remarks
22/2/18	store	5		5		
	store -	0832438			36999	
15/2/18	Zbrahim		1	4		
3.	store Thankam		1	3		given to pts -
4.	Memaka		1	2		
10/2/18	store -	1	1	2	rate - 36999	

Zbrahim






Oxygen Concentrator

Model: ZY-10ZW
Power: C220V / 50Hz
Power: 550W
Flow: 1-10 L/min
Oxygen Purity: 93% ± 3%
Other details can be found in the instructions
Date of production: fuselage bar code



Model No: ZY-10ZW
Date Of Manufacture: 24th 06, 2021
Serial No: 21 43411046



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)
TOLL FREE NO: 1800-425-7669
BAR CODE - (9004-990815225) 0832438

