

**PROFORMA  
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)  
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: UPHC ANAPUZHA	Name of District: Thrissur
Name of Equipment: Nebulizer	Equipment ID & Barcode: #145464- 0840327
Make: Omron	Model: NE-C101
Serial Number: NA	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: Ticket master installation date -13-08-2019, As per Stock book installation date is 22-02-2019	Present status of the equipment: Fully damaged
Date of breakdown:08-11-2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC

Action taken: Checked the machine and found motor and piston are defective. Need to replace these spares for further checking working condition of equipment.


Recommendations for repair (required service details): Not Recommending

Cost of spares (specify parts and cost): NA


#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 1875/-
---	---------------------


Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix report attached.


Reasons for recommending the equipment as BER: Checked the machine and found motor and piston are defective. In ticket master installation date is 13.08.2019, As per the stock book installation date is 22-02-2019 (Stock book attached for reference) and aged up to 5 years and 09 months. Quotation not submitted since the spares are not available in the market. So, we recommending the equipment for condemnation.

  
**BIJO T JOY**  
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:  
*Year met*



Recommended for BER (Yes/ No): <i>Yes</i>	 Signature of JC BM (NHM)
Date:	

**Medical Officer**  
**UFHC - Anapuzha**  
 Kodungallur, Thrissur.  
  
 Signature of Superintendent / Medical Officer (i/c)



Date: *3/12/2024*

\*Not mandatory      #Based on the period of life and value as per the BER guidelines      \* Attach Photograph







# STOCK REGISTER

59

Name of Articles NEBULISER

Cost Rate

Marshal

DATE	PARTICULARS	V.No. of Bill No.	RECEIPT		ISSUED		BALANCE
			Quantity		Quantity		Quantity
24/9/16	KERALA SURGICAL EQUIPMENT ERANJMKOLAM [LIFE LINE ALFA]	112	1	1930/-			1
24/9/16	Issued		1	NIL			
9/3/18	AEROMIST NEBULISER.		1	918/-			
9/5/18	Issued		1	NIL			
22-2-19 7-8-19	OMRON		2				
22-2-19	Issued to		5/0				
13/6/19	KARUNYA medicine depot.		3	3696.39/- 1232.12/-			
13/6/19	Issued to uphc gassaji			-1			
	" " uphc kachai			-1			
	" uphc V.R param.			-1			
							Balance - Nil
22/11/19	Health Storey						
22-11-19	Nebuliser Mask Adult		2	8500/-			17000
19/nd	Nebuliser Mask Pediatric		2	8500/-			17000
9-5-18	AEROMIST NEBULISER MASK			140.00/-			
9-5-18	Issued			NIL			

barcode - 70083622, 70083619, 0840327

① Noble





OMRON Compressor Nebulizer  
 MODEL: NE-C101  
 RATING: 230V ~ 50Hz 150VA

2018009475VF

OMRON HEALTHCARE Co., Ltd.  
 KYOTO, JAPAN

CE 0197  
 IP21

OMRON HEALTHCARE EUROPE B.V.  
 Soerenga 33, 2132 LB Hoofddorp,  
 THE NETHERLANDS

Asia Pacific HQ: OMBON HEALTHCARE SINGAPORE PTE LTD  
 MADE IN VIETNAM

CYRIX HEALTH CARE PVT.LTD  
 BEMP SERVICE PROVIDER  
 PREVENTIVE MAINTENANCE SERVICE

PMS DONE ON: 24/7/23  
 NEXT PMS DUE: 24/7/23  
 PERFORMED BY: Anu

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)  
 TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225)  
 0840327

