

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (BER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: PHC VELLANGALLUR	Name of District: Thrissur
Name of Equipment: B P Apparatus	Equipment ID & Barcode: #146238- 0842244
Make: OMRON	Model: HEM-7120
Serial Number: 20140619687VGI	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: Ticket master installation date - Nil, Stock book installation date - 20-05-2019	Present status of the equipment: Fully damaged
Date of breakdown: 12-11-2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: The machine checked and found motor and pump are defective. Need to replace these spares for further checking of equipment.	
Recommendations for repair (required service details): Not Recommending	
Cost of spares (specify parts and cost): NA	

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 2200/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix report attached.	

Reasons for recommending the equipment as BER: The machine checked and found motor and pump are defective. In ticket master installation date is Nil, As per the stock book installation date is 20-05-2019 and aged up to 5 years and 05 months. Quotation not submitted since the spares are not available in the market. So, we recommend the equipment for condemnation.

[Handwritten Signature]
27/11/2024

BIJO T JOY
Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:
both motor and pump defective

[Handwritten Signature]
27/11/24
Signature of JC BM (NHM)

Recommended for BER (Yes/ No): <i>Yes</i>
Date:



[Handwritten Signature]
MEDICAL OFFICER IN CHARGE
FAMILY HEALTH CENTRE
Vellangallur, Thrissur - 680662
Signature of Superintendent / Medical Officer (i/c)

Date: *27/11/24*
*Not mandatory

#Based on the period of life and value as per the BER guidelines * Attach Photograph



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



No : 127001

SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

<p>Health Facility..... <u>PNC</u></p> <p>Address..... <u>VELLANGALLUR</u> <u>THRASSUR</u></p> <p>Ph:..... <u>8177374210</u></p>	<p>Call Registration Date :..... <u>12/11/2024</u></p> <p>Caller ID :..... <u>146238</u></p> <p>Date of Visit :..... <u>12/11/2024</u></p> <p>Asset No :..... <u>0842244</u></p> <p>EQPT Name :..... <u>BP APPARATUS</u></p> <p>Manufacture..... <u>OMRON</u> Model : <u>NA</u></p> <p>S.No..... <u>NA</u> Dept: <u>PUBLIC HEALTH</u></p>
--	---

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : ERROR E3

Action Taken : THE MACHINE CHECKED & FOUND MOTOR & PUMP ARE DEFECTIVE. NEED TO REPLACE THESE SPARES FOR FURTHER CHECKING OF EQUIPMENT.

Completed Date : 12/11/2024 Time : 10:00am Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>AMBODY V S</u>	<u>12/11/24</u>	<u>9:00am</u>	<u>10:00am</u>

Customer Remark Completed Pending

Service Engineer Name : AMBODY V S
Signature: [Signature]
Date: 12/11/24
Contact Number: 7907996500

Customer Name : [Signature]
Signature: [Signature]
Date: [Signature]
Contact Number: [Signature]
Designation : **MEDICAL OFFICER IN CHARGE**
Hospital Seal: **FAMILY HEALTH CENTRE Vellangallur, Thrissur - 680662**



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225)
0842244



