



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 124360

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility: <u>GH Pala</u>	Call Registration Date: <u>2/11/2024</u>
Address: <u>Kottayam</u>	Caller ID: <u>144325</u>
Ph: <u>6238159232</u>	Date of Visit: <u>2/11/2024</u>
	Asset No: <u>0512690</u>
	EQPT Name: <u>Pulse Oximeter</u>
	Manufacture: <u>BPL</u> Model: <u>Cleo</u>
	S. No. <u>CMME762486</u> Dept. <u>Female Ward</u>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Main PCB and Display unit not working

Action Taken : checked and found out the problem with main PCB and Display unit. Need to replace these spare for further checking and working condition of the equipment. Enquired the spare with OEM.

Completed Date : 2/11/2024 Time : 5 pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Josekutty George</u>	<u>2/11/2024</u>	<u>3pm</u>	<u>5pm</u>

Customer Remark
Not working

Completed Pending

Service Engineer Name : Josekutty George Customer Name : MARIAMMA VARTHASE
Signature: [Signature] Signature: Mari's
Date: 2/11/2024 Date: 2/11/24
Contact Number: 8848288888 Contact Number: 9495992990



PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: GH PALA	Name of District: KOTTAYAM
Name of Equipment: Pulse Oximeter	Equipment ID:144325 & Barcode: & 0512690
Make: BPL	Model: CLEO
Serial Number:CMME7G2486	Warranty details: No warranty
Date of purchase/ Year of manufacture/ Installation Date: 22/06/2005	Present status of the equipment: Fully Damaged
Date of breakdown: 02/11/2024 (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/ CMC

Action taken: Checked and found out the problem with main PCB And Display unit .Need to replace these spares for further checking and working condition of the equipment. Enquired the spare with OEM

Recommendations for repair (required service details) : Not recommending for repair.

Cost of spares (specify parts and cost): MAIN PCB :16520/- ,DISPLAY UNIT : 8260/-
 TOTAL : 24780/-

Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 87.7% Asset Value:28250/-

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not):
 Cyrix service report and vendor quotation attached.

Reasons for recommending the equipment as RBER: Checked and found out the problem with main PCB And Display unit .Need to replace these spares for further checking and working condition of the equipment. Machine installed in 22/06/2005 and covered up to 19 + Years. The repairing cost of the equipment is 87.7%.,both criteria met .So recommended the machine for condemnation as per the tender clause 5.3.14.1

Josekutty George
 16/11/2024
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM: *Asset Value - ~~28250~~ 28250, cost of spares - 24780, life period - 19 years, As per the RBER guide lines recommend as RBER*

Recommended for BER (Yes/ No): *Yes*

Date: _____

Date: _____
 Medical Officer (i/c)

CONSULTANT BIOMEDICAL ENGINEER
 NATION: _____
 Signature of JC BM (NHM)
 Signature of Superintendent /
SUPERINTENDENT
 K. M Mani Memorial
 Govt. General Hospital, Pala



PULSE-OXY-METER

Rd

Cost

115

~~Rd from store~~

Each unit includes:-

power cord - 1

Adult probe - 1 → unserviceable - IN 22

paediatric probe - 1

user manual - 1

warranty card - 1

RS 27163+
1087 (Tax)
~~RS 29250~~

PSK

All items

Nil

Tr. out to CHwd on 15/10/2005

12/2006 Tr. in from CHwd on 15/12/2006.

Handwritten notes and signatures

2015 Rd. from store

1

RS-45252/-

2

model - Correl - plus Accusure
Sr. No. - CR14 PA 7371 DUSA Y oxygen sensor

m.s. Page 167 Vol. V

PSK

2024 Condemned as per order no.
C2-5335/21 KMMSCGH Pala
dated 2/1/2024



Handwritten signature in green ink

RMO
K.M. MANI MEMORIAL
GOVT. GENERAL HOSPITAL
PALA



QTN/BISO/081124-083

08/11/2024

QUOTATION

To,
CYRIX HEALTH CARE PVT LTD
ERNAKULAM

Dear Sir,

Please find below our lowest quotation for Pulse Oxymeter, Make-Bpl ,Model-Cleo

SL NO.	ITEM DESCRIPTION	UNIT PRICE
1	MAINBOARD	14000+18%TAX
2	DISPLAY	7000+18%TAX

Terms and conditions:

1. *Payment 100% in advance along with Purchase order*
2. *Delivery within 15 days*

Thanking you and assuring you our best attention always.

For Biomedix Solutions



[Signature]
Authorized Signatory

Mobile: 9846969564 | Email: biomedixservice@gmail.com

Branches: Kozhikode, Kannur, Kottayam-Kollam.

For enquiries contact: biomedixsales@gmail.com, biomedixservice@gmail.com



BIOMEDIX SOLUTIONS

No: 7/437- B, Mannarakayam P.O.,

Kathilankalpadi, Kanjirapally, Kottayam - 686506.

Mob:9846969564,

8/24, 11:30 AM



Anjitha T L <purchasecyrix2@gmail.com>

Quotation for spare parts 0512690 pulse oximeter

2 messages

Fri, Nov 8, 2024 at 9:58 AM

Anjitha T L <purchasecyrix2@gmail.com>
To: BIOMEDIX SOLUTIONS <biomedixsales@gmail.com>, biomedix solutions <biomedixqln@gmail.com>, BPL Medical Equipment Service <biomedixservice@gmail.com>

Dear Sir,

Please share quotation for the below spare parts

- 1) Main board
- 2) Display unit

Make-BPL
Model- cleo
Serial no-CMME7G2486
-

Thanks & Regards

Anjitha T L
Sr. Purchase Executive
PH:- 9947719111
purchasecyrix2@gmail.com
www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd
| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India
KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

BIOMEDIX SOLUTIONS <biomedixsales@gmail.com>
To: Anjitha T L <purchasecyrix2@gmail.com>

Fri, Nov 8, 2024 at 11:17 AM

Madam,
Please find the quotation attached herewith



Regards
Biomedix Solutions
Kottayam
[Quoted text hidden]

QTTN 083 CYRIX HEALTH CARE 081124.pdf
583K



[Handwritten signature]
RMO
K. M. MANI MEMORIAL
GENERAL HOSPITAL
PALA

[Handwritten signature]
08/11/24
SUPERINTENDENT
K. M Mani Memorial
Govt. General Hospital, Pala



RMO
 K.M. MANI MEMORIAL
 GOVT. GENERAL HOSPITAL
 PATA