




**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: DH PALAKKAD	Name of District: PALAKKAD
Name of Equipment: BIOCHEMISTRY ANALYZER	Equipment ID & Barcode: #147431 - 0911201
Make: ROBONIK	Model: PRIETEST TOUCH
Serial Number: AT0011011RBK	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: In ticket master installation date is 16/11/2011, As per the stock book installation date is 15-11-2011 (stock book attached for reference)	Present status of the equipment: Fully damaged
Date of breakdown: 18/11/2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked the biochemistry analyzer found result variation and display issue. Problem with main board and display. need to be replace these spares for further checking of machine.	
Recommendations for repair (required service details): No recommendations	
Cost of spares (specify parts and cost): NA	
#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 64760/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix Service report attached EOL attached.	
Reasons for recommending the equipment as BER: Checked the biochemistry analyzer found result variation and display issue. Problem with main board and display. need to be replace these spares for further checking of machine. In ticket master installation date is 16/11/2011, As per the stock book installation date is 15/11/2011 aged up to 13 years. OEM informed that this model is obsolete. End of life letter issued. As per tender clause 5.3.14.2 we recommending the equipment for condemnation.	
<p align="right">AKHILESH P  30/11/24 Name & Signature of CYRIX Authority with date</p>	
Remarks of Junior Consultant (Biomedical) NHM: Equipment installed on 15/11/2011 and aged up to 13 years. OEM submitted EOL letter.	
Recommended for BER (Yes/ No): Yes	<p align="right"> Signature of JC Biomedical Engg Consultant (Biomedical Engg) NHM (Arogyakeralam), Palakkad</p>
Date: 30/11/24.	
Date: 30/11/24	<p align="right"> Signature of Superintendent, Deputy Director of Health Services & Superintending District Hospital Palakkad</p>



Seal

Signature of Superintendent,
Deputy Director of Health Services
& Superintending
District Hospital Palakkad

*Not mandatory

*Based on the period of life and value as per the guidelines * Attach Photograph



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 126346

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>DH palakkad</u> Address..... <u>palakkad (dt)</u> Ph:..... <u>8113055076</u>	Call Registration Date :..... <u>18.11.2024</u> Caller ID :..... <u>147431</u> Date of Visit :..... <u>19.11.2024</u> Asset No :..... <u>0911201</u> EQPT Name : <u>Biochemistry Analyzer</u> Manufacture..... <u>Robonik</u> Model: <u>Prietest touch</u> S. No. <u>AT00110112BK</u> Dept..... <u>Laboratory</u>
--	---

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : main board & display issue

Action Taken : checked the equipment found result variation and display issue. problem with main board and display. Need to replace these spare for further checking.

Completed Date : 19.11.2024 Time :..... 11:00 AM Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. <u>Nil</u>			
2.			
3.			

Cyrux Engineer	Date	Start Time	End Time
<u>Dileep</u>	<u>19.11.24</u>	<u>9.00 AM</u>	<u>11.00 AM</u>

Customer Remark Completed Pending

Service Engineer Name : Dileep A
 Signature: Dileep AF
 Date: 19.11.2024
 Contact Number: 9048751909



Customer Name :
 Signature: Vayasee
 Date: 19/11/2024
 Contact Number:
 Designation :
 Hospital Seal

**Deputy Director of Health Services
& Superintendent,
District Hospital, Palakkad**

SEMI AUTO ANALYSER (BIOCHEMISTRY)

147

Date	Particulars	Receipt	IDM	Balance	Sr.
	Brought forward from previous				
	of RDC stock Book of				
	Equipments E ₁				
15-11-2011	Received from store	1		1	
	Automatic Analyser (Biochemistry)				JK
	Priced touch				
	Srno. A T 001101 RBK (1)				
	Rs 83200/-				
	Accessories				
	Sine wave UPS - 1				
	Halogen Lamp - 1				
	Thermal print paper - 2				
	Dust cover - 1				
	Variable vol Pipette - 1				
	(0-100ml)				
	Fixed volume (50ml) - 1				
	pipette				
	ROBONIK (India) Pvt Ltd				
	Transfered to Unserviceable	1		Nil	
	Register G1				JK

Vijaya Lakshmi

Deputy Director of Health Services
& Superintendent
District Hospital, Palakkad



Date: 25.11.2024

Dear Valued Customer,


Thank you for choosing **Robonik (India) Pvt. Ltd.** We appreciate the opportunity to support our business. This letter is to inform you regarding the EOL (End of Life) of the instrument Semi Automated Biochemistry Analyser (Model: Prietest Touch) having **Serial No. AT0011011RBK.**

As required, we stated herewith that the said instrument Semi Automated Biochemistry Analyser (Model : Prietest Touch) having **Serial No. AT0011011RBK** as **End of Life (EOL) Notified.** (During from installation is 10 years).

Thank you in advance for your understanding and your cooperation. If you have any questions, please contact our Customer Care Department at Toll Free No. 18005727977.

Yours faithfully,

For **ROBONIK (INDIA) PVT. LTD.**

A handwritten signature in black ink next to a circular blue stamp. The stamp contains the text 'ROBONIK (INDIA) PVT. LTD.' around the perimeter.

Authorised Signatory



Please share eol letter 9:56 am ✓✓

Ok 10:19 am

➡ Forwarded



Date: 25.11.2024

Dear Valued Customer,

EOL Letter = Prietest Touch - AT0011011RBK.pdf

1 page • 89 kB • PDF

🗨 Message

