

**PROFORMA  
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)  
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: PHC PUTHUNAGARAM	Name of District: PALAKKAD
Name of Equipment: BP APPARATUS	Equipment ID & Barcode: #147038 - 0940559
Make: INSTAPRO	Model: IP 9000
Serial Number: NA	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: In ticket master Installation date - NIL, As per stock 03/10/2019 stock book attached for reference.	Present status of the equipment: Fully damaged
Date of breakdown: 15/11/2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked the BP Apparatus found not switching on. Problem with main board, need to be replace this spare for further checking of machine.	
Recommendations for repair (required service details): No recommendations	
Cost of spares (specify parts and cost): NA	

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 2200/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix Service report attached.	

Reasons for recommending the equipment as BER: Checked the BP Apparatus found not switching on. Problem with main board, need to be replace this spare for further checking of machine. In ticket master installation date is Nil, As per the stock book Equipment installed on 03/10/2019 aged up to 5 years 1 months. Quotation not submitted since spare are not available from market. So we are recommended the equipment for condemnation .


AKHILESH P  
Name & Signature of CYRIX Authority with date *30/11/24*

Remarks of Junior Consultant (Biomedical) NHM:  
*BP apparatus installed on 3/10/2019 and aged up to 5 years. Quotation not submitted.*

Recommended for BER (Yes/ No): *Yes*  
Date: *30/11/24*

Consultant (Biomedical Engg) NHM (Arogyakeralam), Palakkad  
Signature of JC.BM (NHM) NHM (Arogyakeralam), Palakkad  
*30/11/24*

Date: *03/12/2024* Seal  
Signature of Superintendent / Medical Officer (i/c)  
\*Not mandatory #Based on the period of life and value as per the BER guidelines \*Attach Photograph





# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

No : 128674

# CYRIX<sup>®</sup>

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>PHC</u> Address..... <u>PUTHUNAGARAM</u> <u>PALAKKAD</u> Ph :..... <u>7306250300</u>	Call Registration Date : ... <u>15/11/2024</u> ..... Caller ID : ..... <u>147038</u> ..... Date of Visit : ..... <u>16/11/2024</u> ..... Asset No : ..... <u>0940559</u> ..... EQPT Name : ..... <u>BP APPARATUS</u> ..... Manufacture. <u>Instapro</u> Model : <u>1P9000</u> ..... S. No..... <u>NA</u> ..... Dept. <u>Sub Centre</u>
---	--

**Service Classification :** Breakdown call  PMS  Calibration  Cust. Training

**Problem Identified :** ..... NOT WORKING .....  
 .....  
 .....

**Action Taken :** ..... Checked the BP Apparatus found not switching on. Problem with main board. Need to replaced this spare for further checking of machine. .....  
 .....


Completed  Date : .. 16/11/24 .. Time : .. 11:00 AM .. Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			


Cyrix Engineer	Date	Start Time	End Time
<u>LIBIL LALIL</u>	<u>16/11/24</u>	<u>10:00 AM</u>	<u>11:00 AM</u>

**Customer Remark** ..... Completed  ..... Pending



**Service Engineer Name :** LIBIL LALIL  
**Signature:** [Signature]  
**Date:** 16/11/2024  
**Contact Number:** 8891382686


**Customer Name :** AKHIKA A.P  
**Signature:** [Signature]  
**Date:** 16/11/2024  
**Contact Number:** 7306250300  
**Designation :** JPHN GR II  
**Hospital Seal**



BP Apparatus - Digital

5

Date	Items	Receipt	Issue	Balance	Signature
3/10/19	BP apparatus 1 Received from PWC	1	1	-	<i>Malye</i>
	Exp. dt - 12/2028]		[00 us]		
20/1/23	2. BP apparatus received from PWC (3 to India Hd - Exp. 30/10/2027)	1	1	-	<i>[Signature]</i>
			[TO MLHP]		



MEDICAL OFFICER  
PRIMARY HEALTH CENTRE  
PUDUN GARAM  
PALARANAD  
53/1



MEDICAL OFFICER  
 PRIMARY HEALTH CENTRE  
 PUDUN GARANG  
 FALAKNAMA  
 03/12/20