

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: DH PALAKKAD	Name of District: PALAKKAD
Name of Equipment: FULLY AUTO RANDOM ACCESS CLINICAL CHEMISTRY ANALYZER	Equipment ID & Barcode: #147432 - 0911202
Make: ERBA	Model: EM 360
Serial Number: 60303	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: 28/05/2012 (stock book attached for reference)	Present status of the equipment: Fully damaged
Date of breakdown: 18/11/2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC

Action taken: Checked the machine found result variations. Problem with RGT cooling assembly, assembly of photometer with packing, total internal tubing's, RIPT assy, Big KNF pump, cuvette ring set and cuvettes damaged. need to be replace these spares for further checking of machine.

Recommendations for repair (required service details): No recommendations

Cost of spares (specify parts and cost):

RGT cooling assy with 6mm plate and spacer = 95335/-	Big knf pump assy = 52210/-
Assembly of photometer with packing = 324875/-	cuvette ring set assembly = 52210/-
Total internal tubing kit with connector = 75670/-	Cuvette 60 nos = 324300/-
RIPT assy = 72450/-	GST 18% = 179034.3/-
	Total = 1173669.3/-

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 108.4%

Asset Value: 1082525/-

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix Service report attached and quotation attached.

Reasons for recommending the equipment as BER: Checked the machine found result variations. Problem with RGT cooling assembly, assembly of photometer with packing, total internal tubings, RIPT assy, Big KNF pump, cuvette ring set and cuvettes damaged. need to be replace these spares for further checking of machine. Equipment installed on 28/05/2012 aged up to 12 years 5 months. Repairing cost is 108.4% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending for condemnation.


AKHILESH P 
Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:

Equipment installed on 28/5/2012 aged up to 12 years
Repairing cost is more than the asset value. Both criteria of RBER met.


Recommended for BER (Yes/ No): Yes

Date: 30/11/24


Signature of Consultant (Biomedical Engg) Consultant (NHM) Palakkad NHM (Arogyakeralam)

Date: 30/11/2024




Signature of Director of Health Services
Superintendent, District Hospital Palakkad
Medical Officer (i/c)
Based on the period of life and value as per the BER guidelines * Attach Photograph

*Not mandatory



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 126347

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... DH palakkad Address..... Ph:..... 8113055076	Call Registration Date : 18.11.2024 Caller ID : 147432 Date of Visit : 19.11.2024 Asset No : 09.11202 EQPT Name : Fully Auto Random access Clinical Biochemistry Analyzer Manufacture..... ERBA Model : EM360 S. No..... 6.0303 Dept..... Laboratory
---	---

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : **Result variation**

Action Taken : checked the machine found problem with RPT cooling assembly, assembly of photometer with packing, total internal Tubings, R.I.P.T assy, Big KMF pump, Cuvett ring set and Cuvette damaged. Need to replace these spare for further checking of the machine

Completed Date : **19.11.2024** Time : **3:00 PM** Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Dileep	19.11.24	11:30 AM	3:00 PM

Customer Remark Completed Pending

Service Engineer Name : **Dileep A.**
 Signature: **Dileep A**
 Date: **19.11.2024**
 Contact Number: **9048751999**



Customer Name :
 Signature:
 Date: **19/11/2024**
 Contact Number:
 Designation :
 Hospital Seal

Vijay Kumar
 Deputy Director of Health Services
 & Superintendent,
 District Hospital, Palakkad

QUOTATION

Bill To: CYRIX HEALTHCARE PVT LTD
 First Floor,petta Junction 30/641b, Poonithura,Ernakulam, Kerala,
 Pincode 682038 , 678001


Ref. No : 2425QUOTE002382
 Date : Nov 28, 2024 5:34 PM
 Validity : 30 days
 TBM GSTIN : 27AAACTP2038C1ZT

Customer GSTIN No : 32AAFCC2499H2ZM

Ship to, The Superintendent, District Hospital, Palakkad
 RDC LabCourt Road , 678001

SUB : QUOTATION FOR SUPPLY/REPLACEMENT OF SPARES IN INSTRUMENT MODEL: EM360 SR NO.: 60303

With reference to the above, following Spares / Consumables are required to be replaced / supplied. Price/s of the same is / are given below, for approval:

SR. NO	Part Code	HSN Code	Description of Spares/Consumables	Qty	Unit Price	Discount%	GST%	Amount	
1	114595	90278990	RGT COOLING ASSY WITH 6MM PLATE & SPACER_XLV2	1	95335		18	95335	
2	103093	90278990	ASSEMBLY OF PHOTOMETER WITH PACKING	1	324875		18	324875	
3	104951	90278990	TOTAL INTERNAL TUBING KIT WITH CONNECTOR (XL600/300/V2/A4/EM360)	1	75670		18	75670	
4	105281	90278990	R1PT ASSY(SPARE) FOR XL-600/300-	1	72450	0	18	72450	
5	182496	90278990	Big Knf Pump Assy.	1	52210		18	52210	
6	182530	90278990	Cuvette Ring Set Assembly (WØ Cuvettes)	1	49795		18	49795	
7	100279	70171000	CUVETTES	60	5405		18	324300	
Merchant Name :TRANSASIA BIO-MEDICALS LTD								Total (Rs)	994635
								Discount (Rs.) (-)	0
								GST (Rs.)	179034.3
Rupees in words: Eleven lakhs Seventy Three thousand Six hundred Sixty Nine rupees and Three paise only								Grand Total (Rs.)	1173669.3

TERMS AND CONDITIONS :

1. Payment : Full payment should be made in advance through NEFT / RTGS/ Demand Draft in favour of Transasia Bio-Medicals Ltd to Bank A/C no. : 00600330002538 IFSC : 00600330002538 UPI ID : transasiabiomedicals@citibank. You can also pay us through UPI apps like BHIM/ Gopay/ Paytm/ PhonePe etc. by scanning the above mentioned code. Cash Payment is strictly prohibited. No employee of the TBM is authorized to collect cash and or issue receipt against collection of Cash.
2. GST : As applicable
3. Other Cost / Local duties : Customer's responsibility
4. Delivery : Within 1-2 weeks from the receipt of your work order with payment confirmation

Thanking You



RECEIVED & ACCEPTED

Your faithfully,
 For TRANSASIA BIO-MEDICALS LTD

CUSTOMER SIGNATURE & STAMP:
 CUSTOMER NAME: _____
 CONTACT NO: _____

TRANSASIA BIO-MEDICALS LTD.

Head Office: Transasia House 8 Chandivli Studio Road, Andheri (E), Mumbai - 400 072. Tel.: (022) 4030 9000 Fax: (022) 2857 3030 Mumbai | Toll Free No. 1800 103 8226, WhatsApp No. 8451048434, SMS No. 9212433444

www.transasia.co.in
 REG. NO. 5020-05 / RS - 7



FULLY AUTOMATED BIOCHEMISTRY ANALYSER

Date	Particulars:	Received	Item	Balance	Sr
28-05-2012	Brought forward from RDC Stock Register of Equipments.				
28-05-12	Fully Automatic Biochemistry Analyser EM 360 (ERBA) Sl no. 60303 Item code 195069 Rs. 11,87,044.76		1	1	1
	② Item code 120270 System pack kit Lot no. 6071124				
	③ 120221 Erba pan (1x5ml) Lot: 2071113				
	④ 150403 water treatment plant RO+MB hybrid plant with 1520LPH, digital conductivity meter to provide out put water quality less than 100 micro Siemens. Sr no: 033				
	Item code 111277 UPS (1) Rate: 42000/- (Price Supply)				
	Item code: 15074 Refrigerator (1) Rs 23000/- (Refrigerator transferred to Unserviceable Register Cyl)				
28-05-2016	Brought forward from for 86 New Equipment Register DH Lab Palakkad. D3 (L25) purchased from Transasis Bio Medicals through KMSCL Rs. 1060815 Slno. 60573		1	2	2
	Kindo Accessories Erba Norm (1x5ml)				
	Water treatment plant 1 : 1				
	System pack demo kit : 1				
	ERBA Model ERBA EM360				

Vijayulu
Deputy Director of Health Services & Superintendent, District Hospital, Palakkad





REQUIRED QUOTATION FOR FULLY AUTOBIOCHEMISTRY ANALYZER 0911202 # 147432 @ DH PALAKKAD

4 messages

Akhilesh P <akhileshcyrix@gmail.com>

Wed, 20 Nov, 2024 at 11:02 pm

To: SANDEEP DAS M <s.das@transasia.co.in>

Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <Zm2.klbemp@cyrix.in>, transserve <transserve@transasia.co.in>

"Dear Sir,

Please visit DH Palakkad and share the price and stock availability of the required spare parts for the EM360 fully automatic analyzer."

Deatils are mentioned below

Hospital Name : DH PALAKKAD

Equipment Name : FULLY AUTOMATIC BIOCHEMISTRY ANALYZER

Make : TRANSASIA

Model : EM360

SN : 60303

AKHILESH P
DISTRICT INCHARGE - PALAKKAD
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com

Vishal Makwana <v.makwana@transasia.co.in>

Thu, 21 Nov, 2024 at 8:58 am

To: Akhilesh P <akhileshcyrix@gmail.com>, SANDEEP DAS M <s.das@transasia.co.in>

Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <Zm2.klbemp@cyrix.in>, transserve <transserve@transasia.co.in>, a.pratheesh@transasia.co.in, SAJIN NAIR (SSN) <n.sajin@transasia.co.in>

Dear Customer,

Hope this email finds you well.

This is to keep you informed that we acknowledge and need your cooperation in resolving your complaint at the earliest,

SR- 2425SR720055

Our team will review your concern and going to visit by tomorrow.

Appreciate your patience and continued support.

In case, if you wish to add more information, please feel free to respond to this email or connect us at

Toll Free: 1800 103 8226 / Whatsapp: 8451048434 / transserve@transasia.co.in

Thanks & Regards

Vishal Makwana

[Quoted text hidden]

Akhilesh P <akhileshcyrix@gmail.com>
To: SANDEEP DAS M <s.das@transasia.co.in>

Thu, 28 Nov, 2024 at 11:36 am

Dear sir
Please update the status.

Thanks & Regards

AKHILESH P
DISTRICT INCHARGE - PALAKKAD
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com
[Quoted text hidden]

SANDEEP DAS M <s.das@transasia.co.in>
To: Akhilesh P <akhileshcyrix@gmail.com>

Fri, 29 Nov, 2024 at 10:21 am


Dear Akhilesh ,

Please find the attachment for your reference

Get [Outlook for Android](#)

From: Akhilesh P <akhileshcyrix@gmail.com>
Sent: Thursday, November 28, 2024 11:36:42 AM
To: SANDEEP DAS M <s.das@transasia.co.in>
Subject: Re: REQUIRED QUOTATION FOR FULLY AUTOBIOCHEMISTRY ANALYZER 0911202 # 147432 @ DH PALAKKAD

[Quoted text hidden]



Varghese
 Deputy Director of Health Service
 & Superintendent,
 District Hospital, Palakkad

