



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 123950

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website :www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>DH</u>	Call Registration Date : ... <u>12-10-2024</u>
Address..... <u>DH KOLLAM</u>	Caller ID : <u>140242</u>
.....	Date of Visit : <u>13-10-2024</u>
Ph:..... <u>7510614125</u>	Asset No : <u>0210114</u>
	EQPT Name : <u>Pharmacy Refrigerator</u>
	Manufacture..... <u>LG</u> Model : <u>intellocool</u>
	S. No..... <u>5955</u> Dept..... <u>MICU</u>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Not cooling , Compressor Not getting On.

Action Taken : Checked the refrigerator found that not cooling, Compressor not getting on informed the refrigerator technicians to service the unit.

Completed Date : 13-10-2024 Time : 12:55 pm Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Wilfred Sunil William</u>	<u>13-10-2024</u>	<u>12:10 pm</u>	<u>12:55 pm</u>

Customer Remark Completed

Service Engineer Name : <u>Wilfred Sunil William</u>	Customer Name : <u>Govt Binu Francis</u>
Signature: <u>Wf</u>	Signature: <u>FB</u>
Date: <u>13-10-2024</u>	Date: <u>13-10-2024</u>
Contact Number: <u>9778526652</u>	Contact Number: <u>7510614125</u>
	Designation : <u>SNO</u>
	Hospital Seal



Subanta
SUPERINTENDENT
A. A. Rajendran
Dist. Hospital, Kollam

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: DH KOLLAM	Name of District: KOLLAM
Name of Equipment: PHARMACY REFRIGERATOR	Equipment ID & Barcode: 140242 & 0210114
Make: LG	Model: INTELLOCOOL
Serial Number: 5955	Warranty details: NO WARRANTY
Date of purchase/ Year of manufacture/ Installation Date: 30-07-2007	Present status of the equipment: FULLY DAMAGED
Date of breakdown: 12-10-2024 (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: NO AMC/ CAMC

Action taken:
 Checked the machine found that Compressor, heating coil, defrost timer, Bimetal control defective, need to replace these spares for further checking and working condition of the equipment., Enquire spare with OEM

Recommendations for repair (required service details)
 Not Recommending for repair

Cost of spares (specify parts and cost): Total = 10200/-
 Compressor- 4600, Gas charging, filter & Capillary – 2800, Heater coil - 1800, Defrost Timer - 650, Bimetal Control - 350

# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 63.74 %	Asset Value: 16000/-
---	----------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)
 Cyrix service report , Quotation & images attached

Reasons for recommending the equipment as BER:
 Checked and found that Compressor, Heater coil, Defrost Timer, Bimetal control are defective, thereafter gas charging, filter & Capillary need to change the machine installed on 30-07-2007 equipment outlived 17 years 2 month, repairing cost of equipment is 63.74%. As per tender clause 5.3.14.1, both criteria met. so recommending to condemnation

WILFRET SUNIL

WFS
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:
Very old machine and repairing cost is high . So recommended for Beyond economic repair

Recommended for BER (Yes/ No): *Yes*

Date: *13/11/24*

Date:
Medical Officer (i/c)



Signature
SUPERINTENDENT
 A. A. Rahim Memorial Hospital
 Govt. Dist. Hospital, Kollam

13/11/24
 Signature of JC BM (NHM)



PERFECT AIR CONDITIONER

Chemmathoor near private bus stand
punalur, contact number: 9562185059
, Kollam, Kerala, 691305

QUOTATION

Quotation No. : 801
Quotation Date : 07/11/2024
Expiry Date : 14/11/2024

BILL TO

Cyrix Health Care

S.NO.	ITEMS/SERVICES	QTY.	RATE	AMOUNT
1	COMPRESSOR CHANGE	1 PCS	4,600	4,600
2	GAS CHARGING & FILTER CHANGE & CAPILLARY CHANGE	1 PCS	2,800	2,800
3	HEATER COIL TYPE	1 PCS	1,800	1,800
4	DEFROST TIMER CHANGE	1 PCS	650	650
5	BIMETAL CONTROL CHANGE	1 PCS	350	350

SUBTOTAL

5

₹ 10,200

NOTES

DH kollam ,LG doubledoor intelli cool refrigerator

TAXABLE AMOUNT

₹ 10,200

TOTAL AMOUNT

₹ 10,200

BANK DETAILS

Name: ANIJA A J
IFSC Code: SBIN0070779
Account No: 67261447806
Bank: State Bank of India ,PSBADOOR

Total Amount (in words)

Ten Thousand Two Hundred Rupees




Authorised Signature for
PERFECT AIR CONDITIONER

Counter signed
Suhane

SUPERINTENDENT
A. A. Rahim Memorial
Govt. Dist. Hospital, Kollam





Reference
SUPERINTENDENT
A. A. Rahim Memorial
Govt. Dist. Hospital, Kollam

