

C58.



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1015612

CYRIX®

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility.....U.P.H.C...Kannamaola	Call Registration Date : ...9.10.2024.....
Address.....Nilandram.....	Caller ID : 139482.....
.....	Date of Visit : ...9.10.2024.....
Ph :9444631084.....	Asset No :0141221.....
	EQPT Name : ..Fetal Doppler.....
	Manufacture..Nilcamed.. Model : ...na.....
	S.No...na.....Dept.Nursing station

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified :Not working.....

Action Taken :Checked and found that the probe and battery socket complaint. Need to replace these parts for further checking and working condition of the equipment. Engaged Spares with the vendor.....

Completed Date : Time : Spare Required

Spare Replaced Requested


Description	Qty.	Part Number	PR Number
1.			
2. NA	NA	NA	NA
3.			

Cyrix Engineer	Date	Start Time	End Time
Jeswin Sibi	9/10/24	10:30 AM	11:30 AM

Customer Remark Completed Pending




Service Engineer Name : Jeswin Sibi
Signature: *[Signature]*
Date: 9/10/2024
Contact Number: 9833315512

Customer Name : *[Signature]* Dr Rupa Prasad
Signature:
Date: 9/10/24
Contact Number: 9562602141
Designation :
Hospital Seal



PROFORMA

RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)


Name of Hospital: UPHC KANNAMMOOLA	Name of District: Thiruvananthapuram
Name of Equipment: FETAL DOPPLER	Equipment ID & Barcode: 139482 & 0141221
Make: NISCOMED	Model: NA
Serial Number: NA	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 07/10/2020	Present status of the equipment: Fully Damaged
Date of breakdown: 08/10/2024 (Toll free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked and found that the Probe and battery socket complaint. Needs to replace these spares for further checking and working condition of the equipment. Enquired spares with vendor.	
Recommendations for repair (required service details): Not recommending for repair.	
Cost of spares (specify parts and cost): 1. probe for fetal doppler - 4500/-	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 112.5%	Asset Value: 4000/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service report and reference quotation attached.	
Reasons for recommending the equipment as BER: Checked and found Battery socket, Probe complaint. The equipment was installed on 07/10/2020 and covered up to 4+ years. The repair cost is 112.5%. As per the tender clause 5.3.14.1 both criteria met. So recommending the unit for condemnation.	
 Kasey p v Name & Signature of CYRIX Authority with date	
Remarks of Junior Consultant (Biomedical) NHM: As per tender conditions, this can be recommended beyond economic repair.	
Recommended for BER (Yes/ No):	
Date: 13-11-2024	
Date:	 Signature of Superintendent / Medical Officer (i/c)



FETAL DOPPLER

Date	Particulars	Receipt	Inv	Balance	Rem
07/10/2020	Received from NIMHJ	1		1	+
03/10/2020	Inued to Medical Officer JPHN.		1	NIL	-
24/7/23	Forwarded Surgical & Lab supplies.			2	-
					2500.00
01/1/24	Forwarded Surgical & Lab Supplies	1	1	0	-
15/5/24	Inued to HWC Medical College.		1	0	-

Rupali
 MEDICAL OFFICER
 HWC, KANNANMOOLA, TYP





974473043 974473243

cloud99biotech@gmail.com

KMC/XIII-1333
Room No: 22
CSI Golden Jubilee Complex
Baker Hill
Shastri Road
Kottayam

CLOUD99 BIOTECH

GSTIN: 32AAWPT5633Q1ZD
State: 32 - Kerala

MISME NO.: UDYAM-KL-07--0030714

Estimate

Estimate No.: C993-QTN-24
Date: 29/06/2024
Place of Supply: 32-Kerala

Estimate For:

ABHILASH A

CYRIX HEALTHCARE PRIVATE LIMITED
1 ST FLOOR 30/621B
KOCHI DHANUSHKODI ROAD POOYIT-URA
ERNAKULAM 682038

Contact No.: 6157068893

GSTIN Number: 32AA-CC2A99H2ZM

State: 32-Kerala

Quantity	Unit	Price/Unit	GST	Amount
1	Nos	₹ 4,017.86	₹ 482.14 (12%)	₹ 4,500.00
			₹ 482.14	₹ 4,500.00



Pay To:
Bank Name : UNION BANK OF INDIA KOTTAYAM
Bank Account No. : 361201010055646
Bank IFSC code : UB0536121
Account holder's name : CLOUD99 BIOTECH

Sub Total	₹ 4,017.86
SGST@6%	₹ 241.07
CGST@6%	₹ 241.07
Total	₹ 4,500.00

Description

EQUIPMENT: FETAL DOPPLER
MAKE: KAYNU IMPLEX
BARCODE: 0450613
TICKET ID: KL395
QUOTATION: VALID FOR 15 DAYS
PRICE INCLUSIVE OF TAX AMOUNT

Estimate Amount in Words
Four Thousand Five Hundred Rupees only

Terms And Conditions

Thanks for doing business with us!

For CLOUD99 BIOTECH



Authorized

Pupul

MEDICAL OFFICER
KOTTAYAM HANA SANGH & TRUST





Abhilash A <abhilashacyrix@gmail.com>

FETAL DOPPLER QUOTATION Invoice details

1 message


cloud99biotech@gmail.com <cloud99biotech@gmail.com>
To: ABHILASHACYRIX@gmail.com

Sat, Jun 29, 2024 at 4:53 PM

Dear ABHILASH A,
Thanks for doing business with us. Below are your invoice details.
Also attached are the PDF of invoice for your convenience.
Estimate:

Amount: 4500

Thanks & Regards

 vyapar_print_29_06_2024_16_45_11.pdf
196K

