



**PROFORMA**  
**RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)**  
**BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: UPHC PALKULANGARA	Name of District: Thiruvananthapuram
Name of Equipment: ENT HEADLIGHT	Equipment ID & Barcode: 139518 & 0140703
Make: ABE	Model: SL-01
Serial Number: NA	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 09/01/2020	Present status of the equipment: Fully Damaged
Date of breakdown: 08/10/2024 (Toll free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked and found that the unit is not getting ON. Identified that the Head assembly, Charging adapter and Battery pack complaint. Needs to replace these spares for further checking and working condition of the equipment. Enquired spares from vendor.	
Recommendations for repair (required service details): Not recommending for repair.	
Cost of spares (specify parts and cost): 1. Head assembly - 5264/- 2. Charging Adaptor - 840/- 3. Battery Pack - 649/- <b>Total = 6753/-</b>	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 15500/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service report and vendor quotation attached (Attached quotation is only for Ent Headlight asset value mention on the bill is for whole ear care kit)	
Reasons for recommending the equipment as BER: Checked and found Head assembly, Charging adapter and Battery pack complaint. The equipment was installed on 09/01/2020 and covered up to 4+ years. The repair cost is more than the asset value of ENT headlight. As per the tender clause 5.3.14.1 both criteria met. So recommending the unit for condemnation.	
 Name & Signature of CYRIX Authority with date	
Remarks of Junior Consultant (Biomedical) NHM: <i>Asset value of head light is almost 7000/- . Since repair cost is high . This can be RBER .          Call date : 8.10.2024 . Quotation date - 28.11.2024 . Sub: date . 1.12.24</i>	
Recommended for BER (Yes/ No):	 <b>MANJEE SHA MOHAN</b> Junior Consultant (Biomedical) National Health Mission Thiruvananthapuram
Date:	
Date:	Seal      Signature of Superintendent / Medical Officer (i/c)



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

No : 125206

# CYRIX®

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

<p>Health Facility... <u>UPHC Palkulangau</u>.....</p> <p>Address... <u>Palkulangau</u>.....</p> <p>.....<u>Thiruvananthapuram, Kerala</u>.....</p> <p>Ph:.....<u>81 86 83 80 62</u>.....</p> <p><i>Battery pack</i></p>	<p>Call Registration Date : .....<u>08/10/24</u>.....</p> <p>Caller ID : .....<u>139578</u>.....</p> <p>Date of Visit : .....<u>09/10/24</u>.....</p> <p>Asset No : .....<u>01AD703</u>.....</p> <p>EQPT Name : .....<u>ENT Head Light</u>.....</p> <p>Manufacture...<u>ABX</u>..... Model : <u>SL-01</u>.....</p> <p>S. No...<u>1250105058</u>..... Dept...<u>Op</u>.....</p>
--	--

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : Checked and found lead assembly, charging adapter and battery pack damaged

Action Taken : Need to replace head assembly, charging adapter and battery pack per further checking of the equipment

Completed  Date : 09/10/24..... Time : 05:30p..... Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2. <u>NR</u>	<u>NR</u>	<u>NR</u>	<u>NR</u>
3.			

Cyrix Engineer Mohammed Syad M Date 09/10/24 Start Time 04:30 PM End Time 07:30 P

Customer Remark Completed Pending

Service Engineer Name : Mohammed Syad M  
 Signature: [Signature]  
 Date: 09/10/24  
 Contact Number: 923499127



[Signature]  
**MEDICAL OFFICER**  
 UPHC, PALKULANGARA, TVPM

Date:  
 Contact Number:  
 Designation :  
 Hospital Seal

# MEDPLUS BIOMEDICALS

Athulraj M  
Kumaraswamy raja road  
Chelannur (P.O)  
Kozhikode

☎ 8848372192 ✉ medplusbiomedicals.clt@gmail.com  
GSTIN 32CEYPA2140F1ZV

## Quotation

To,  
Cyrix Healthcare  
Manager  
Kochi

Quotation# Quote-121  
Date: 28/11/2024

Dear Sir/Mam,

Thank you for your valuable inquiry. We are pleased to quote as below

#	DESCRIPTION	QTY	PRICE	GST	TOTAL
1	Head assembly	1 unit	₹ 4,700.00	₹ 564.00 12.0%	₹ 5,264.00
2	Charging adaptor	1 unit	₹ 750.00	₹ 90.00 12.0%	₹ 840.00
3	Battery pack	1 unit	₹ 550.00	₹ 99.00 18.0%	₹ 649.00
SUB TOTAL					₹ 6,000.00
GST					₹ 753.00
GRAND TOTAL					₹ 6,753.00

We hope you find our offer to be in line with your requirement.

### Payment Instructions

Account name : MEDPLUS BIOMEDICALS

Acc no : 07791012000500

IFSC CODE : PUNB0077910

Bank name : PUNJAB NATIONAL BANK

Branch : ERANJIPALAM

For, MEDPLUS BIOMEDICALS

### Terms & Conditions:

- 100 % advance
- Delivery : 7 days after confirmation
- Quotation validity : one month



*Dr. Am*  
MEDICAL OFFICER  
UPHC, PALKULANGARA, TVPM

AUTHORIZED SIGNATURE

REGARDING QUOTATION - 139518 - 0140703 - UPHC PALKULANGARA - ENT HEAD LIGHT

message

Thu, 28 Nov 2024 at 11:32 a

Abijith-KLBEMP <dm2.klbemp@cyrix.in>  
Athulraj m <medplusbiomedicals.clt@gmail.com>  
Afsal Yunuze <zm1.klbemp@cyrix.in>, Kasyep P V <kasyepcyrix@gmail.com>

Dear Sir,  
Kindly share quotation.

Thanks & Regards,  
Abijith.A  
District Incharge | Ernakulam  
dm2.klbemp@cyrix.in  
www.cyrixhealthcare.com  
Cyrix Healthcare Pvt Ltd  
| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India  
KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

From: DM Abijith-KLBEMP  
Sent: 24 November 2024 11:33 AM  
To: Athulraj m <medplusbiomedicals.clt@gmail.com>  
Cc: Afsal Yunuze <zm1.klbemp@cyrix.in>  
Subject: REGARDING QUOTATION - 139518 - 0140703 - UPHC PALKULANGARA - ENT HEAD LIGHT

Dear Sir,  
Kindly provide a quotation for Ent Head light spares.

- 1)Head Assembly
- 2)Charging Adaptor
- 3)Battery Pack

Thanks & Regards,  
Abijith.A  
Divisional Manager | South Zone  
dm2.klbemp@cyrix.in  
www.cyrixhealthcare.com  
Cyrix Healthcare Pvt Ltd  
| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India  
KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH




*Dr. Arp*  
MEDICAL OFFICER  
UPHC, PALKULANGARA, TVPM

147

EAR CARE KITS.

Date 9/1/21	Particulars	Receipt	Invoice	Balance	Detail.
2019 - 2020	Received from KMSCL Sn No: 18-19/A/15 1160 13822/- Mvni 3000	32	Order	No: 36	1266/
9/1/21	R1-Lx / Heine Herman, ABE Kerala & Challenges 1286, 673. warranty 91120 - 81121. Given to self Nurse	1	1	Nil	1



*[Handwritten Signature]*  
Dr. Ar.

**MEDICAL OFFICER**  
**UPHG, PALKULANGARA, TVPM**

