



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 233162

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Service Report

Health Facility : <u>PHC</u>	Call Registration Date : <u>29/07/2024</u>
Address : <u>PHC Puthuppally</u>	Caller ID : <u>126202</u>
<u>Kottayam</u>	Date of Visit : <u>31/07/2024</u>
Ph : <u>9447771517</u>	Asset No. : <u>0546469</u>
	EQPT Name : <u>Fetal doppler</u>
	Manufacture : <u>N/A</u> Model : <u>N/A</u>
	S. No. <u>072018</u> Dept. <u>OP</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Foetal doppler not Working. checked and found problem with main board and probe

Action Taken : Enquire the OEM for the quotation for spare parts.
*(The equipment details are not available in the main and sub stock register at hospital)

Completed Date : 31/7/24 Time : 11:30am Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Sanish Sivan</u>	<u>31/7/24</u>	<u>11:00am</u>	<u>11:30am</u>

Customer Remark : Not working Completed LPending
Refer to N/o PHC Puthuppally.

Service Engineer Name : Sanish Sivan
 Signature : [Signature]
 Date : 31/7/24
 Contact Number : 9946646571

Customer Name : Dr. hen Ann Jose
 Signature : [Signature]
 Date : 31/8/24
 Contact Number : 701
 Designation : Medical Officer
 Hospital Seal : PRIMARY HEALTH CENTRE PUTHUPPALY




**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOTTAYAM
2	Name of Hospital	PHC PUTHUPPALLY ✓
3	Name of Equipment with Make, Model and Serial Number	Equipment : Fetal doppler Make : Nil Model :nil SN:072018
4	Equipment ID & Barcode	126202 & 0544469 ✓
5	Date of purchase / Year of manufacture /Installation Date	28/10/2019
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	29/07/2024 (TOLL FREE)
9	Action taken	Checked and found out the problem with main pcb and probe. Enquired the spare with OEM
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	MAIN BOARD -1350/- PROBE - 3804/- TOTAL-5154/-


13	Asset Value	Rs .4000 /-
14	#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	128%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report and Quotation are attached
16	Reasons for recommending the equipment as BER	Checked and found out the problem with main pcb and probe. Machine installed in 28/10/2019 and covered up to 4+.The repairing cost 128 % both criteria met. So recommending the machine for condemnation as per the tender clause 5.3.14.1
17	Name & Signature of CYRIX Authority	Sanish Shivan 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Asset value - 4000 (as per Dashboard)
 The item not available in the stock book of PAC Puthuppally.
 Cost of spare ~~is~~ 5154 (128%)
 The spare cost is high compared to Asset value


 Signature of JC BM (NHM)
 CONSULTANT BIOMEDICAL
 NATIONAL HEALTH MISSION
 KOTTAYAM

Date




 MEDICAL OFFICER IN CHARGE
 SUPERINTENDENT
 PRIMARY HEALTH CENTRE
 PUTHUPPALLY

Fetal Doppler

Date	Received from	Rate	Receipt Issue Bal	Sig
15/11/2019	Received from molpnc	3,998.00/-	1	/
Make : EC [LADA care]				
Model : Nil				
BI NO : 072018				
Rate :				
ISSUE NO :				



Handwritten signature

MEDICAL OFFICER IN CHARGE
PRIMARY HEALTH CENTRE
PUTHUPALLY



9744473043, 9744473243

cloud99biotech@gmail.com

1 KMC/XIII-1333
Room No: 22,
CSI Golden Jubilee Complex
Baker Hill
Shastri Road
Kottayam

CLOUD99 BIOTECH

GSTIN: 32AMWPT5633Q1Z0
State: 32 - Kerala

MSME NO.: UDYAM-KL-07--0030714
DL NO.: KL/KTM/MD42/2023/000015

Estimate

Estimate For:

ARUN G C

1ST FLOOR 30/641B CYRIX HEALTHCARE PVT LTD
KOCHI DHANUSHKODI ROAD POONITHURA

Contact No.: 8921992272

GSTIN Number: 32AAFCC2499H2ZM

State: 32-Kerala

Estimate No.: C998-QTN-49

Date: 19/08/2024

Place of Supply: 32-Kerala

#	Item name	HSN/ SAC	Quantity	Unit	Price/ Unit	GST	Amount
1	BOARD FOR FETAL DOPPLER (BOARD FOR FETAL DOPPLER)	9018	1	Pcs	₹ 1,205.36	₹ 144.64 (12%)	₹ 1,350.00
2	PROBE FOR FETAL DOPPLER (PROBE FOR FETAL DOPPLER)	9018	1	Pcs	₹ 3,396.43	₹ 407.57 (12%)	₹ 3,804.00
Total			2			₹ 552.21	₹ 5,154.00

Pay To:

Bank Name : UNION BANK OF INDIA, KOTTAYAM

Bank Account No. : 361201010035646

Bank IFSC code : UBIN0536121

Account holder's name : CLOUD99 BIOTECH

Sub Total	₹ 4,601.79
SGST@6%	₹ 276.11
CGST@6%	₹ 276.11
Total	₹ 5,154.00

Description

PHC PUTHUPPALLY
EQUIPMENT - FETAL DOPPLER
MAKE: JAPAN TECHNOLOGY
LOT NO: 072018

Estimate Amount in Words

Five Thousand One Hundred Fifty Four Rupees only

Terms And Conditions

Thanks for doing business with us!

For : CLOUD99 BIOTECH



Authorized Signatory



Santhosh
MEDICAL OFFICER IN CHARGE
PRIMARY HEALTH CENTRE
PUTHUPPALLY



Suma
MEDICAL OFFICER IN CHARGE
PRIMARY HEALTH CENTRE
PUTHUPPALLY

