



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 227650

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : <u>Thgh</u>	Call Registration Date : <u>24/6/24</u>
Address : <u>Kothamangalam</u>	Caller ID : <u>118866</u>
<u>Ernakulam, Kerala</u>	Date of Visit : <u>27/6/24</u>
Ph : <u>9400030042</u>	Asset No. : <u>0722216</u>
	EQPT Name : <u>Suction Apparatus</u>
	Manufacture : <u>NA</u> Model : <u>NA</u>
	S. No. : <u>NA</u> Dept. : <u>Casualty</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : unit not working

Action Taken : checked the unit and check working and found that the suction motor damage and the suction fan defective. need to replace these spares for further checking and working condition of the Equipment

Completed Date : 27/6/24 Time : Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Akhil T. Lenin</u>	<u>27/6/24</u>	<u>10:00 AM</u>	<u>11:00 AM</u>

Customer Remark Completed Pending


Service Engineer Name : <u>Akhil T. Lenin</u>	Customer Name : <u>Ranj George</u>
Signature : <u>[Signature]</u>	Signature : <u>[Signature]</u>
Date : <u>27/6/24</u>	Date : <u>27/6/24</u>
Contact Number : <u>9605479326</u>	Contact Number : <u>9400030042</u>
	Designation : <u>SUPERINTENDENT</u>
	Hospital Seal : <u>TALUK HEAD QUARTERS HOSPITAL KOTHAMANGALAM - 686 001</u>

REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	ERNAKULAM
2	Name of Hospital	THQH KOTHAMANGALAM ERNAKULAM
3	Name of Equipment with Make, Model and Serial Number	Equipment : Suction Apparatus Make :Na Model : Na SN :Na
4	Equipment ID & Barcode	118866 & 0722216
5	Date of purchase / Year of manufacture /Installation Date	02-05-2017
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	24-06-2024 (Toll free)
9	Action taken	Checked and found suction motor and suction jar defective. Need to replace these spares for further checking and working condition of the equipment.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	Suction Jar – Rs.3600/- Suction motor – Rs.4650 Total – Rs.8250/-

13	Asset Value	Rs. 10800/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	76%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached and reference quotation attached.
16	Reasons for recommending the equipment as BER	Checked and found suction motor and suction jar defective. Machine installed on 02-05-2017 and covered up to 7+ years. The repairing cost of the equipment is 76% both criteria met so recommending the equipment for condemnation as per the tender clause 5.3.14.1.
17	Name & Signature of CYRIX Authority	Akhil T Lenin 

*Not mandatory **#Based on the period of life and value as per the BER guidelines**

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Asset value is Rs 10800/- Initially
 Quot not submitted. Submitted Quot on
 6/8/24.
 Submitted after 36 days.


 Signature of JCBM (NHM)
 CONSULTANT
 Bio-Medical Engineer
 National Health Mission (NHM)
 Aroyyakeralam Ernakulam
 09/08/2024

Date 16/8/24


 Seal
 Taluk Head Quarters Hospital
 KOTHAMANGALAM - 686 691
 686 691


 SUPERINTENDENT
 Taluk Head Quarters Hospital
 KOTHAMANGALAM (1c)
 Superintendent / Medical Officer (1c)



9744473043, 9744473243

cloud99biotech@gmail.com

1 KMC/XIII-1333
Room No: 22,
CSI Golden Jubilee Complex
Baker Hill
Shastri Road
Kottayam

CLOUD99 BIOTECH

GSTIN: 32AMWPT5633Q1Z0

State: 32 - Kerala

MSME NO: : UDYAM-KL-07-0030714

Estimate

Estimate For:

KASHYAP

1ST FLOOR 30/641B CYRIX HEALTHCARE PVT LTD
KOCHI DHANUSHKODI ROAD POONITHURA

Contact No.: 7593847134

GSTIN Number: 32AAFCC2499H2ZM

State: 32-Kerala

Estimate No.: C99B-QTN-28

Date: 08/07/2024

Place of Supply: 32-Kerala

#	Item name	HSN/ SAC	Quantity	Unit	Price/ Unit	GST	Amount
1	SUCTION JAR	90189099	1	Nos	₹ 3,214.29	₹ 385.71 (12%)	₹ 3,600.00
2	SUCTION MOTOR (MOTOR FOR HOSPITAL SUCTION)	9018	1	-	₹ 3,940.68	₹ 709.32 (18%)	₹ 4,650.00
Total			2			₹ 1,095.03	₹ 8,250.00



Pay To:

Bank Name : UNION BANK OF INDIA KOTTAYAM

Bank Account No. : 361201010035646

Bank IFSC code : UBIN0536121



Account holder's name : CLOUD99 BIOTECH

Sub Total	₹ 7,154.97
SGST@6%	₹ 192.86
CGST@6%	₹ 192.86
SGST@9%	₹ 354.66
CGST@9%	₹ 354.66
Total	₹ 8,250.00
You Saved	₹ 0.01

Description

DH NEDUMANGAD

SUCTION APPARATUS

Barcode: 0112464

Ticket id: 114956

Make : Supreme Enterprises

Estimate Amount In Words

Eight Thousand Two Hundred Fifty Rupees only

Terms And Conditions

Thanks for doing business with us!

For: CLOUD99 BIOTECH



Authorized Signatory



Quotation suction Machine - DH nedumangad (114956/0112464)

4 messages

Kasyep P V <kasyepcyrix@gmail.com>
To: CLOUD 99 <cloud99biotech@gmail.com>

Tue, 9 Jul 2024 at 5:28 pm

Dear Madam,

This mail is to enquire about the spare quotation for the suction machine;
Make : Supreme Enterprises

Quotation required;
Suction motor
Suction jar

Thanks & Regards,

KASYEP PV
District Incharge ,Trivandrum
CYRIX HEALTHCARE Pvt.Ltd
KLBEMP
Ph : +91 7593847134
kasyepcyrix@gmail.com

Kasyep P V <kasyepcyrix@gmail.com>
To: CLOUD 99 <cloud99biotech@gmail.com>

Sat, 20 Jul 2024 at 9:14 pm

Dear sir,
Consider this mail as gentle Reminder

[Quoted text hidden]

CLOUD 99 <cloud99biotech@gmail.com>
To: Kasyep P V <kasyepcyrix@gmail.com>

Sat, 20 Jul 2024 at 10:08 pm

[Quoted text hidden]

Kasyep P V <kasyepcyrix@gmail.com>
To: DM Abijith-KLBEMP <dm2.klbemp@cyrix.in>

Sun, 21 Jul 2024 at 11:38 am

[Quoted text hidden]

STOCK REGISTER OF T.H.Q. HOSPITAL, KOTHAMANGALAM

Name of Article... SUCTION APPARATUS


Date	No. and date of voucher or invoice	From whom received or to whom issued	Receipt	Issued	Balance after each transaction	Initials of Receiver	Remarks
		Brought forward - From old book page no 79					
10/1/14							
1/12/03		Tr. in from e	1		1		
		pmh S.S.R p.m.j					
		34					
28/2/14		Tr. in from e	1		2	ch	
		mm hand p.m.j					
		34					
		1 pay ward					
2/5/2017		Tr in from	1		3		
		Palliative Unit					Rs. 7950
		page no. 87					
1/3/18		Transferred to		1	2		
		unserviceable Reg					
		p.no. 15					
02.6.2020		RFS	1		3	SKO	2-71361-
29.7.2020		RFS	1		4	SKO	
5/8/2021		Transferred to					
		unserviceable Reg	1		3	SKO	
		p.no. 15					




[Handwritten Signature]

SUPERINTENDENT
TALUK HEAD QUARTERS HOSPITAL
KOTHAMANGALAM - 686 691

Carried over



Senthil
SUPERINTENDENT
TALUK HEAD QUARTERS HOSPITAL
KOTHAMANGALAM - 686 691



KOTHAMANGALAM - 686 691
Toll Free No: 686 691

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMIP)
TOLL FREE NO: 1800 - 425-7669
BAR CODE : (8004-890615225)
0722216

ON / OFF

VALVE CONTROL