


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Ernakulam
2	Name of Hospital	CHC KALADY
3	Name of Equipment with Make, Model and Serial Number	Equipment: NEBULISER Make : LIFECARE Model : CNB69009 Sr.no-: LIFEC092009230787
4	Equipment ID & Barcode	100545 & 0732810
5	Date of purchase / Year of manufacture /Installation Date	4/4/2019
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	11/03/2024(Toll-free)
9	Action taken	Checked the machine and found that motor and piston faulty.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	It is not recommend for repair
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	RS.1,353
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found that compressor motor and piston are defective. Equipment installed on 4/4/2019 and the machine outlived 5+ years. Quotation not submitted since spares are not available in the market. So it is recommending for condemnation
17	Name & Signature of CYRIX Authority	 JIBIN THOMAS GEORGE


*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

27/3/24
Signature of JC BM (NHM)


Superintendent / Medical Officer (I/c) *27/04/2024* *27/04/2024*
Medical Officer in Charge
Community Health Centre
Kalady, Ernakulam Dist. Date



**MEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 187009

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility C.H.C.....	Call Registration Date : 11/03/2024.....
Address C.H.C. Kalady..... Ernakulam, Kerala.....	Caller ID : 100545.....
Ph : 9544725107.....	Date of Visit : 12/03/2024.....
	Asset No. : 0732810.....
	EQPT Name : Nebuliser.....
	Manufacture Life Care Model : CN166004.....
	S. No. Life Care 2009230787 Dept. Store.....

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified :
Compressor working with high sound
No pressure produces.

Action Taken :
Checked and found Compressor piston
detached from rotor not repairable and
Compressor for nebuliser not available
Recommend for I.C.R.

Completed Date : 12/3/24 Time : 4:20pm Spare Required

Spare Replaced Requested


Description	Qty.	Part Number	PR Number
1. Compressor for nebuliser			
2. Lifecare :- CB1B 2009	1		
3.			

Cyrix Engineer	Date	Start Time	End Time
Yeechunrasna As	12/3/24	3:30pm	4:20pm

Customer Remark Completed Pending

Service Engineer Name : Yeechunrasna
Signature :
Date : 12/3/2024
Contact Number : 8080147236

Customer Name : Shemena N/A
Signature :
Date : 12/3/24
Contact Number : 94475
Designation : Hospital Staff
Hospital Name : Kalady - 683 57



Medical Officer In-Charge
Community Health Centre
Kalady, Ernakulam Dist.


05/04/2024

STOCK REGISTER

Name Nebuliser

61

Date	Received from or issued to	Inv. No. or Issue	Rate		Receipts	Issues	Balance	Remarks
			Rs.	P.				
4/21	BF casualty SR P.No - 27				1		3	Band
	BF ward SR P.No - 103				1		4	Band
2/8/22	RFS (Dr. Moxepan)	STOCK REGIST	1500/-		2		6	Band
1/22	Received from store (Hamdyneb Nebuliser)	of 1st time P.No 70			1		7	Band
10/3/22	RFS		Rs. 1548/-		1		8	Band
	Hajal kashani 23/7/16							
2/1/23	Given to unseevable Book				2		3	Band
2/6/23	RFS, p (Omman) (SR Vol 2 Page No. 81)		1530/-		2		5	Band



CYRIX

HEALTHCARE EQUIPMENT

CN: Ketchikan
PRO: nealuser

ID: 0732810

CAL.ON: 9-6-23

CAL.DUE: 6-6-24

356 8011-208, Certified Co. Ph. 9742465073



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)


TOLL FREE NO: 1800 - 425-7669



BAR CODE - (8004-890615225)
0732810



LifeCare™

Compressor Nebulizer  Lifecare®
Model No.: CNB69009
Input : AC~220V/50HZ
Power : 180VA
Serial No.: LIFEC092009230787
Mode of operation: Continuous
No AP/PG IPX



