



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

No : 1024210

## CYRIX®

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

### Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website :www.cyrix.com | Email : bemp.kl@cyrix.in

<p>Health Facility: <u>Family Health Center</u></p> <p>Address: <u>Ve Home Thiruvananthapuram</u></p> <p>Ph: <u>9447153397</u></p>	<p>Call Registration Date: <u>01.11.2024</u></p> <p>Caller ID: <u>144176</u></p> <p>Date of Visit: <u>01.11.2024</u></p> <p>Asset No: <u>0142837</u></p> <p>EQPT Name: <u>B.P apparatus</u></p> <p>Manufacture: <u>Life care</u> Model: <u>NA</u></p> <p>S. No: <u>NA</u> Dept: <u>FWC</u></p>
--	--

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : Mercury spillage & glass rod defective

Action Taken : Checked and found mercury spillage. Identified glass rod defective. Need to replace these spares for further checking and working condition of the equipment

Completed  Date: 01.11.2024 Time: 11.00am Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Mohammed Sherif S</u>	<u>01.11.2024</u>	<u>10:30am</u>	<u>11:00am</u>

Customer Remark: \_\_\_\_\_ Completed: \_\_\_\_\_

Service Engineer Name : Mohammed Sherif S

Signature: [Signature]

Date: 01.11.2024

Contact Number: 7994042445

Date: 01.11.2024

Customer Name: [Signature]

Signature: [Signature]

Contact Number: 9447153397




Designation : JOPHN HWC Puthencherry


Hospital Seal



Pending  
 18/11/24  
 695312  
 11/24  
 9447153397  
 JOPHN HWC Puthencherry

**PROFORMA**  
**RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)**  
**BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: PHC VETTORE	Name of District: Thiruvananthapuram
Name of Equipment: BP APPARATUS	Equipment ID & Barcode: 144176 & 0142837
Make: LIFECARE	Model: NA
Serial Number: NA	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 13/06/2007	Present status of the equipment: Fully damaged
Date of breakdown: 01/11/2024 (Toll free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked and found mercury spillage. Identified glass rod defective. Need to replace these spares for further checking and working condition of the equipment.	
Recommendations for repair (required service details) : Not recommending for repair.	
Cost of spares (specify parts and cost): NA	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 1323/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not) : Cyrix service report attached.	
Reasons for recommending the equipment as BER: Checked and found mercury spillage , Identified glass rod defective. The unit was installed on 13/06/2007 and covered up to 17+ years. Since the mercury is not available in the market, recommending the unit for condemnation.	
Kasyep.p.v  Name & Signature of CYRIX Authority with date	
Remarks of Junior Consultant (Biomedical) NHM: Mercury leakage. Since it is not <del>for</del> possible to repair economically , can be RBER	
Recommended for BER (Yes/ No):	 Signature of JC BM (NHM)
Date: 15-11-2024	
Date: 18/11/24	 Signature of Superintendent / Medical Officer (i/c)



മേധാവിയുടെ സഹായം ഇൻ ചാർജ്  
 ഡി.എം.ആർ.എ.ഒ. (i/c)  
 പിൻ: 695312


# Bp APPARATUS

Date	From Whom Issued or to Whom Issued	Receipt	Amount	Balance	Initials/Remarks
12/6/07	PAC Uthara	2		2	W
	PAC Uthara (Panchayath)	1		3	
11/7/07	PAC Uthara	Kudrime (DMS)		11	V. S. Gangaiah
10/6/08	PAC Uthara	Uthara dandi		15	W

12/11/25

അധികാരി ത.വിനോദ് ഇൻ ചാർജ്  
കുടുംബ ആരോഗ്യകേന്ദ്രം, വെട്ടൂർ  
ഫീൻ: 695312





18/11/24

മെഡിക്കൽ ഓഫീസർ ഇൻ ചാർജ്  
 കൂടുമ്പെ ആരോഗ്യകേന്ദ്രം, വെട്ടൂർ  
 ഫിൻ: 695312

