



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1024211

CYRIX®

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Service Report

Health Facility: <u>Family Health Center</u>	Call Registration Date : <u>01.11.2024</u>
Address: <u>Veddy, Thiruvananthapuram</u>	Caller ID : <u>144172</u>
Ph: <u>9447153397</u>	Date of Visit : <u>01.11.2024</u>
	Asset No : <u>0142496</u>
	EQPT Name : <u>Digital Bp apparatus</u>
	Manufacture : <u>ACON</u> Model : <u>OB11-111</u>
	S. No. <u>903A 6006D99</u> Dept. <u>FWC</u>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Not getting ON

Action Taken : Checked and found the unit is not getting on. Identified motherboard and display defective. Need to replace these spares for further checking and working condition of equipment.

Completed Date : 01.11.2024 Time : 11:20am Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Mohammed Sherif</u>	<u>01.11.2024</u>	<u>11:00am</u>	<u>11:30am</u>

Customer Remark Completed

Service Engineer Name : Mohammed Sherif
Signature: [Signature]
Date: 01.11.2024
Contact Number: 7994042447

18/11/24
Customer Name: SONI S
Signature: [Signature]
Date: 11/11/24
Contact Number: 9447153397
Designation: JOHN HWC Pathway
Hospital Seal

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: PHC VETTORE	Name of District: Thiruvananthapuram
Name of Equipment: B P APPARATUS	Equipment ID & Barcode: 144172 & 0142496
Make: ACON	Model: OB11-111
Serial Number: 903A0006D89	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 11/07/2007	Present status of the equipment: Fully Damaged
Date of breakdown: 01/11/2024 (Toll free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC

Action taken: Checked and found the unit is not getting ON. Identified Mainboard and display defective. Need to replace these spares for further checking and working condition of the equipment.


Recommendations for repair (required service details): Not recommending for repair.

Cost of spares (specify parts and cost): NA


# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 2200/-
--------------------------------------------------------------------------------------------	---------------------


Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not):
 Cyrix service report attached.

Reasons for recommending the equipment as BER: Checked and found that the Mainboard and display defective. The unit is installed on 11/07/2007 and covered up to 17+ years. Quotation not attached since the spares are not available in the market. So recommending the unit for condemnation

Kasyp.pv 
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:
 As per the tender conditions, this can be RBER.


Recommended for BER (Yes/ No):	 MANEESHA MOHAN M.R. Junior Consultant (Biomedical) National Health Mission Thiruvananthapuram Signature of BM (NHM)
Date: 12-11-2024	

Date: 18/11/24  Signature of Superintendent / Medical Officer (i/c)

Bp APPARATUS

Date	From Whom Issued	To Whom Issued	Receipt	Issued	Balance	Balance
12/6/07	PAC	McNees	2		2	✓
	PAC	McNees	1		3	
11/5/07	PAC	McNees	1		4	✓ 15/11/07
10/6/08	PAC	McNees	1		5	✓

18/11/24



മുഖ്യമന്ത്രിയുടെ കാര്യാലയം
 മന്ത്രിമാരുടെ കാര്യാലയം, മന്ത്രിമാരുടെ
 ഓഫീസ്: 695312



18/11/24.

അഡീക്കൽ തഹ്സീൽ ഇൻ ചാർജ്ജ്
കൂട്ടുബു ആരോഗ്യകേന്ദ്രം, വെട്ടൂർ
ഫിൻ: 695312



ACON® Automatic Blood Pressure Monitor
REF OB11-111



SN 903A0006D89



ACON Laboratories, Inc.
4108 Sorrento Valley Boulevard
San Diego, CA 92121, USA



5V 600mA
(4)AA (LR6)



PN20144-002

CYRIX
HEALTHCARE SYSTEMS LTD

**BIOMEDICAL EQUIPMENT MAINTENANCE
PROGRAMME (BEMP)**



6V
2-⊕

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225)

0142496



(Handwritten signature)
18/11/25

മെഡിക്കൽ ഓഫീസർ ഇൻ ചാർജ്
കുടുംബ ആരോഗ്യകേന്ദ്രം, വെട്ടൂർ
ഫീൻ: 695312

