



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 132040

CYRIX[®]
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website :www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>CHC</u>	Call Registration Date : .. <u>11-11-24</u>
Address..... <u>CHC Pathrasapuram</u>	Caller ID : .. <u>145970</u>
..... <u>Kollam</u>	Date of Visit : .. <u>12-11-24</u>
Ph: .. <u>9848524945</u>	Asset No : .. <u>0230546</u>
	EQPT Name : .. <u>Centrifuge</u>
	Manufacture <u>Rotek</u> Model: <u>PMCAA</u>
	S.No..... <u>NA</u> Dept..... <u>Laboratory</u>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : .. Not getting on

Action Taken : .. Checked and found that motor set on speed regulating transformer. also defective. Equipped with OEM

Completed Date : .. 12-11-24 Time : .. 12:30 PM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer Sudhi Jay. S

Medical Officer in Charge Completed Pending

TALUK HEAD QUARTERS HOSPITAL PATHANAPURAM

Service Engineer Name : Sudhi Jay. S

Signature: Sudhi Jay. S

Date: 12-11-24

Contact Number: 9447621489

Customer Name : Bindu-T

Signature: Bindu





Date: 12/11/2024

Contact Number: 9848524945


Designation : Senior Lab Technician

Hospital Seal

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: CHC PATHANAPURAM	Name of District: KOLLAM
Name of Equipment: CENTRIFUGE	Equipment ID & Barcode: 145970 & 0230546
Make: ROTEK	Model: PMC4A
Serial Number: 1804	Warranty details: NO WARRANTY
Date of purchase/ Year of manufacture/ Installation Date: 13-04-2007	Present status of the equipment: FULLY DAMAGED
Date of breakdown: 11-11-2024 (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: NO AMC/ CAMC
Action taken: Checked and found that the motor & speed regulator with transformer are defective. Enquired spare with OEM	
Recommendations for repair (required service details) Not Recommending for repair	
Cost of spares (specify parts and cost): 4130 Centrifuge Motor = 3000, Transformer set with switch = 500, Tax @18% = 630	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 71.42%	Asset Value: 4900/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not) Cyrix service report ,Quotation, Email communication & images attached	
Reasons for recommending the equipment as BER: Checked and found that the Centrifuge motor, Transformer set with switch are defective. Equipment installed on 13-04-2007 outlived 17-year 6 month. Repairing cost of the equipment is 71.42%. As per tender clause 5.3.14.1 both criteria met so recommend equipment for condemnation.	
SUDHIRAJ	 04/12/24 Name & Signature of CYRIX Authority with date
Remarks of Junior Consultant (Biomedical) NHM: <i>Old machine and repair cost in high . So recommended for Beyond economic repair</i>	
Recommended for BER (Yes/ No): <i>Yes</i>	 Signature of JC BM (NHM)
Date: <i>10/12/24</i>	
Date: Medical Officer (i/c) 	Seal 

For Dr Tom
MEDICAL OFFICER IN-CHARGE
UK HEAD QUARTERS HOSPITAL
PATHANAPURAM





(containing)

Date	From whom Received	Receipt	Issued
24 Oct	Brought forward	for C/R page no 5	
	Value 1		
	Centrifuge	1	
	Used 1		

Balance	Model and Make	Price	Remarks
	Model - 0784	Rs 4900	
	Code - (FHS Pipe/Analys)		



For Dr. Tom
 CHARGE
 HOSPITAL
 PATHANAPURAM



OFFICER IN-CHARGE
KALUK HEAD QUARTERS HOSPITAL
PATHANAPURAM

For Dr Tom
[Signature]

Re: PRICE ENQUIRY - Centrifuge - CHC PATHANAPURAM - 0230546
1 message

Rotek Instruments <mail@rotekinstruments.com>
To: "Abilal S [DI Kollam]" <abilalcyrix@gmail.com>

2 December 2024 at 14:23

Quotation for centrifuge spares

Sir
PFA

Thanks & Regards



ROTEK INSTRUMENTS

B&C Industries, West Vengola P.O., Perumbavoor- 683556, Ernakulum Dist, Kerala,

PII: 8593070000/8593000606 / 9446139563 web: www.rotekinstruments.com



On Mon, Dec 2, 2024 at 1:22 PM Abilal S [DI Kollam] <abilalcyrix@gmail.com> wrote:
Dear Sir ,

Kindly share the Price and Stock availability of Spares needed for Centrifuge.

Spares Needed :

- Motor-1
- Speed Regulator - 1

Hospital : CHC PATHANAPURAM
 Make : Rotek
 Model : PMC4A
 SI No : 1804
 Barcode. : 0230546



Thanks and Regards

ABILAL .S
 District incharge
 CYRIX HEALTHCARE Pvt. Ltd.
 Kollam - 691002
 Ph:+91 7593847161

For De. To. ,
 MEDICAL OFFICER IN-CHARGE
 TALUK HEAD QUARTERS HOSPITAL
 PATHANAPURAM

B&C INDUSTRIES, West Vengola P.O, Perumbavoor ,

Ernakulam , Kerala - 683556 Mail ID – mail@rotekinstruments.com

Ph: 9446139563, 8593077000, 8593000606

GST : 32AAEFB4140N1ZO

QTN- B&C /318/24-25

02/12/2024

To

Cyrix Health Care

Ernakulam

Sub: Quotation for centrifuge repair

Sir

We are sending the rate of centrifuge spares. The rate is given below for your kind consideration

Centrifuge - PMC-4

Centrifuge motor	- Rs 3000 -00
Transformer set with switch	- Rs 500 -00


3500 -00

GST @ 18% 630 -00

4130 -00
=====

Thanks & Regards
For B&C Industries



For
Dr. Tom

MEDICAL OFFICER IN CHARGE
TALUK HEAD QUARTERS HOSPITAL
PATHANAPURAM