

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: CHC MUTTOM	Name of District: IDUKKI
Name of Equipment: Micro Pipette	Equipment ID & Barcode: ID: 146763 Barcode: 0630665
Make: AGAPPAE	Model: Na
Serial Number: Na	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 23/12/2015	Present status of the equipment: Fully damaged
Date of breakdown: 14-11-2024 (toll-free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No CAMC/AMC
Action taken: Checked and found that the push button was defective. This spare needs to be replaced for further checking and the working condition of the equipment.	
Recommendations for repair (required service details) Not recommended for repair	
Cost of spares (specify parts and cost): Na	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: Na	Asset Value: 1800/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not) The Cyrix service report and calibration report are attached	
Reasons for recommending the equipment as BER: Checked and found that the push button was defective. This spare needs to be replaced for further checking and the working condition of the equipment. The equipment was installed on 23/12/2015 and covered up to 9+ years. The quotation was not submitted since spares are not available on the market. So, recommending for RBER.	
Cithu Abraham 28/11/24 Name & Signature of CYRIX Authority with date	
Remarks of Junior Consultant (Biomedical) NHM: * spare cost quotation not attached. * Installation date as stock entry and asset not matching * year of working and % of spare cost agree with RBER criteria	
Recommended for BER (Yes/ No): <u>yes</u>	P. J. Chacko Consultant Biomedical Engineer National Health Mission Arogyakeralam Signature of JC BM (NHM)
Date: 29/11/24	
Date: _____ Seal	Signature of Superintendent / Medical 02-12-2024



മെഡിക്കൽ ഓഫീസർ ഇൻ ചാർജ്
 സാമൂഹിക ആരോഗ്യകേന്ദ്രം
 മുട്ടൂർ, തൊടുപുഴ



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1018627

CYRIX®

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website :www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <i>CHC muttum</i>	Call Registration Date : ... <i>15/11/2024</i>
Address..... <i>[Phodanur 2ha, Jolekudai]</i>	Caller ID : ... <i>146 763</i>
Ph : ... <i>9496479672</i>	Date of Visit : ... <i>15/11/2024</i>
	Asset No : ... <i>0630665</i>
	EQPT Name : ... <i>micro pipette</i>
	Manufacture <i>BYAPAC</i> , Model : ... <i>N111</i>
	S. No. ... <i>N111</i> Dept. ... <i>LMS</i>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : ... *not working.*

Action Taken : ... *Checked and found that the push button was defective. This spare needs to be replaced for further holding and the working condition of the equipment.*

Completed Date : *15/11/2024* Time : ... *12pm* Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<i>Edison shine</i>	<i>15/11/2024</i>	<i>11:00am</i>	<i>12pm</i>

Customer Remark Completed Pending

Service Engineer Name : *Edison shine*
Signature: *[Signature]*
Date: *15/11/2024*
Contact Number: *703446213*

Customer Name : *Shamshir K.S.*
Signature: *[Signature]*
Date: *15/11/24*
Contact Number: *9496479672*
Designation: *[Signature]*
Hospital Seal: *[Seal]*



Handwritten notes and dates: 02-12-2024

CERTIFICATE OF CALIBRATION

CUSTOMER NAME	COMMUNITY HEALTH CENTRE MUTTOM		
ADDRESS	MUTTOM, IDUKKI, KERALA 685587	DATE OF CALIBRATION	18/07/2024
CERTIFICATE NO:	NA	VALID UP TO	NA
		DATE OF ISSUE	18/07/2024

INSTRUMENT DETAILS			
Instrument Name	MICRO PIPETTE	S.No.	NA
Make/Model No.	ENSUREPETTE	Visual Inspection	OK
LOCATION:	LAB	RANGE	1000 µL
BAR CODE : 0630665			

PROBLEM IDENTIFIED : PUSH BUTTON MISSING

ACTION TAKEN :


The pipette was received in broken status. While receiving itself, it was noticed that pipette push button is missing. Currently, pipette push button is not available in the market. So the above mentioned pipette is non-serviceable. Therefore, we recommend to decommission this Pipette.

CALIBRATION STATUS	PASS	FAIL	✓
--------------------	------	------	---

REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By
(CALIBRATION ENGINEER)



Approved By
(TECHNICAL MANAGER)
PREJITH T B






CYRIX BIOMETRICAL EQUIPMENT MAINTENANCE
 PROGRAMME (SERIAL)
 TOLL FREE NO: 1800 - 425-7669
 BAR CODE - (8004-090615225)
 0030665



09062407

Jyothi
 TR NOIC
 02-12-2024

ചെങ്കൽ താലൂക്കിലെ ഇൻ കോർപ്പറേഷൻ
 സാമൂഹിക ആരോഗ്യകേന്ദ്രം
 എടക്കാട്, തൃശ്ശൂർ



MICRO PIPETTE 1000 μ l

Balance

Transferred from Stock Book III Page no 147 on 1/7/15
Bal code 70060081 01.

23/12/15 From store ENSURE BIO741.01 02

METHOD ID F.V. 1000 DT. 10/6/2005

Temp. 20°C


Medical Officer In-charge
Community Health Ce

Bal code (6806249) Cyrix - 0680665

22/1/15

Arden 01

Medical Officer In-charge
Community Health Centre
Mullom



~~OFFICE~~
NOTE
02-12-2024

മെഡിക്കൽ ഓഫീസർ
സാമൂഹിക ആരോഗ്യ
മുട്ടം, തൊടുപുഴ