

27/11/24

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: THQH Kattappana	Name of District: IDUKKI
Name of Equipment: Suction Apparatus	Equipment ID & Barcode: id:146642 barcode:0620567
Make: MONARCH surgical	Model: Na
Serial Number: Na	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 7/5/2020	Present status of the equipment: Fully damaged
Date of breakdown: 14-11-2024 (toll-free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No CAMC/AMC

Action taken: Check and find that the suction motor and suction jar were defective. These spares need to be replaced for further checking and the working condition of the equipment. Enquired spare with OEM

Recommendations for repair (required service details) Not recommending for repair

Cost of spares (specify parts and cost): Suction jar:3600/-
Suction motor:4650/-
Total:8250/-

Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 83% Asset Value:9900/-

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)
The Cyrix service report and OEM quotation are attached

Reasons for recommending the equipment as BER: Check and find that the suction motor and suction jar were defective. The equipment was installed on 7/5/2020 and covered up to 4 + years. The repair cost of the equipment is 83%, and both criteria are met. So recommending the equipment for condemnation as per the tender clause 5.3.14.1.

Jithin Abraham
23/11/24

Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:
* Installation date in stock entry and asset not matching
* Year of working and % spare cost agree with RBER criteria. *Jithin*

Recommended for BER (Yes/ No): *yes*
Date: *29/11/2024*

Consultant, Biomedical Engineer
National Health Mission
Arogyakeralam
Signature of JC/BM (NHM)

Date:
Medical Officer (i/c)

Seal

Signature of Superintendent /



[Signature]
4/11/24
SUPERINTENDENT
TALUK HOSPITAL KATTAPPANA



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 132170

CYRIX®

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>THQH</u>	Call Registration Date : <u>14/11/24</u>
Address..... <u>Kattappana</u>	Caller ID : <u>1466212</u>
..... <u>Idukki, Kerala</u>	Date of Visit : <u>15/11/24</u>
Ph : <u>9645621527</u>	Asset No : <u>0620567</u>
	EQPT Name : <u>Suction apparatus</u>
	Manufacture <u>MONARCH</u> Model : <u>NA</u>
	S. No. <u>NA</u> Dept. <u>Casualty</u>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : NOT working

Action Taken : check and find the suction motor and suction
two were defective. These spares not to be replaced for
further checking and the working condition of the equipment
Engineer Spare with DBM

Completed Date : 15/11/24 Time : 11:30 Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2. <u>NA</u>		<u>NA</u>	<u>NA</u>
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Devakanth</u>	<u>15/11/24</u>	<u>11:00</u>	<u>11:30</u>

Customer Remark Completed Pending

Service Engineer Name : Devakanth

Signature:

Date: 15/11/24

Contact Number: 8848229091



Customer Name : Asha. P. Ravindran

Signature:

Date: 15.11.24

Contact Number: 9446225097

Designation :
Hospital Seal

SUPERINTENDENT
TALUK HOSPITAL KATTAPPANA



9744473043, 9744473243

cloud99biotech@gmail.com

1 KMC/XIII-1333
Room No: 22,
CSI Golden Jubilee Complex
Baker Hill
Shastri Road
Kottayam

CLOUD99 BIOTECH

GSTIN: 32AMWPT5633Q1ZD
State: 32 - Kerala

MSME NO.: UDYAM-KL-07-0030714

Estimate

Estimate For:

KASHYAP

1ST FLOOR 30/641B CYRIX HEALTHCARE PVT LTD
KOCHI DHANUSHKODI ROAD POONITHURA

Contact No.: 7593847134

GSTIN Number: 32AA FCC2499H2ZM

State: 32-Kerala

Estimate No.:

C99B-QTN-28

Date:

08/07/2024

Place of Supply:

32-Kerala

#	Item name	HSN/ SAC	Quantity	Unit	Price/ Unit	GST	Amount
1	SUCTION JAR	90189099	1	Nos	₹ 3,214.29	₹ 385.71 (12%)	₹ 3,600.00
2	SUCTION MOTOR (MOTOR FOR HOSPITAL SUCTION)	9018	1	-	₹ 3,940.68	₹ 709.32 (18%)	₹ 4,650.00
Total			2			₹ 1,095.03	₹ 8,250.00



Pay To:

Bank Name : UNION BANK OF INDIA, KOTTAYAM

Bank Account No. : 361201010035646

Bank IFSC code : UBIN0536121

Account holder's name : CLOUD99 BIOTECH



Sub Total	₹ 7,154.97
SGST@6%	₹ 192.86
CGST@6%	₹ 192.86
SGST@9%	₹ 354.66
CGST@9%	₹ 354.66
Total	₹ 8,250.00
You Saved	₹ 0.01

Estimate Amount In Words

Eight Thousand Two Hundred Fifty Rupees only



Terms And Conditions

Thanks for doing business with us!

For : CLOUD99 BIOTECH



Authorized Signatory



Handwritten signature and date: *As*
4/12/2019
SUPERVISOR
KATTAPPANA TALUK HOSPITAL

