



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

# CYRIX®

HEALTH CARE PVT LTD

No : 1001200

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

## Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... D.M.H.....	Call Registration Date : ..15-11-2024.....
Address..... Peroorkada.....	Caller ID : .....147010.....
..... Thiruvananthapuram.....	Date of Visit : ..15-11-2024.....
Ph : ..... 9544523909.....	Asset No : ..... 0113715.....
	EQPT Name : ..Laboratory Microscope.....
	Manufacture..... Radicals Model : Xmb.A.b.....
	S.No..... 20054..... Dept. Laboratory.....

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : ..... Not working.....

Action Taken : ..... checked and found that the microscope head assembly condenser, objective lenses defective. Needs to replace these for further checking and working condition of the equipment.....


Completed  Date : 15-11-2024 Time : ..... 2:00 p.m ..... Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2. NA	NA	NA	NA
3.			

Cyrix Engineer	Date	Start Time	End Time
Rehna.R	15-11-2024	1:15 p.m	2:00 p.m

Customer Remark  Completed  Pending



Superintendent  
District Medical Hospital  
Peroorkada  
Thiruvananthapuram




Service Engineer Name : Rehna.R  
Signature: *Rehna*  
Date: 15-11-2024  
Contact Number: 9947239956

Customer Name : Arajivay.V.J  
Signature: *Arajivay*  
Date: 15/11/24  
Contact Number: 940022196  
Designation : Lab in-charge  
Hospital Seal



PROFORMA

RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)  
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: DISTRICT MODEL HOSPITAL PEROORKADA	Name of District: THIRUVANANTHAPURAM
Name of Equipment: LABORATORY MICROSCOPE	Equipment ID & Barcode: 147010/0113715
Make: RADICAL	Model: RMH-4B
Serial Number: 20054	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 03/12/1998	Present status of the equipment: Fully Damaged
Date of breakdown: 15/11/2024 (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked and found Prism set, Objective set and Eye piece set, Transformer and Black box defective defective. Need to replace these spares for further checking and working condition of the equipment.	
Recommendations for repair (required service details): Not recommending for repair.	
Cost of spares (specify parts and cost): NA	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 18472/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix Service report attached	
Reasons for recommending the equipment as BER: Checked and found Prism set, Objective set and Eye piece set ,Transformer and black box defective. The unit was installed on 03/12/1998 and outlived up to 26+ years. Since the microscope is too old, all the spares have been damaged and it is not repairable. So recommending the unit for condemnation.	
Kasyp.pv  Name & Signature of CYRIX Authority with date	
Remarks of Junior Consultant (Biomedical) NHM: Machine too old. All the parts including black box defective. So can be RBER as per guidelines.	
Recommended for BER (Yes/ No):	
Date: 28-11-2024	Signature of JCBM (NHM)
	MANEESHA MOHAN Superintendent (Biomedical) National Health Mission Thiruvananthapuram Signature of Superintendent / Medical Officer (i/c)



12 MICROSCOPE

	B/F purchase	Received	Issue	Bal	Seq
From old book page No (5)	MSBL 96 (1)				
3.12.98	Binocular microscope (From TB centre) Radical SI. No 3141/2004	1	1	NIL	8/100
5.5.02	U-17-1185/06 dt 22-5-02 Universal agencies Palakkad Angadi Trichur 680001 Ph. 2410319 2427031 Unitam monocular microscope	1	1	NIL	8/100
6.6.07	Binocular Labomed vision 2006 SI. No. 070410804	1	1	NIL	
24/7/11	Binocular microscope SI. No: 13639 AIMICO (STD surgical sets maternal kit)	2	2		
30-03-2019	Received from Bore	2 no			2 no.
	a Model RXL-SB (LED) B 190807				
	TRANSFERRED TO CANCER CARE Please return it urgently	on 1.1.2023			
	b Model RXL SB (LED) B 190808				
	Radical Scientific Equipments Pvt Ltd Price 3784/-				

Considered as per order no: A1/804/2015 of Dept DMH Perinth

Binocular Microscope see order no: A1/804/2015 of Dept DMH Perinth

Received by customer's  
Signature  
21/11/23  
Returned by  
Signature



Superintendent  
District Medical Hospital  
Perinth

CR to Page no-160



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE  
PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225)  
0113715



KMSCL

