



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 124979

CYRIX[®]
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility... <i>U.P.H.C. secretariate</i>	Call Registration Date : ... <i>18/11/2024</i>
Address... <i>Rajaji Nagar</i>	Caller ID : ... <i>147514</i>
..... <i>Chinnandram</i>	Date of Visit : ... <i>23/11/2024</i>
Ph : ... <i>9633315512</i>	Asset No : ... <i>0142962</i>
	EQPT Name : ... <i>Nebulizer</i>
	Manufacture... <i>NA</i> Model : ... <i>NA</i>
	S. No..... <i>NA</i> Dept. ... <i>OP</i>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : *Not working*

Action Taken : *Checked and found that the nebulizer nozzle damaged and motor weak need to replace the spares for further checking and working condition of the equipment*

Completed Date : Time : Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrux Engineer	Date	Start Time	End Time
<i>Jerwin Sibi</i>	<i>23/11/24</i>	<i>1:00 PM</i>	<i>2:00 PM</i>

Customer Remark: Completed Pending



MEDICAL OFFICER
HDHC RAJAJI NAGAR, TVPM

[Signature]

Service Engineer Name : *Jerwin Sibi*
Signature: *[Signature]*
Date: *23/11/2024*
Contact Number: *9633315512*

Customer Name : *AKHILA M.S.*
Signature: *[Signature]*
Date: *23/11/24*
Contact Number: *8086598087*
Designation : *Staff Nurse*
Hospital Seal

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: UPHC SECRETERIATE (RAJAJI NAGAR)	Name of District: THIRUVANANTHAPURAM
Name of Equipment: NEBULIZER	Equipment ID & Barcode: 147514/0142962
Make: NA	Model: NA
Serial Number: 69520757	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 12/12/2014	Present status of the equipment: Fully Damaged
Date of breakdown: 18/11/2024 (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC

Action taken: Checked and found Nebulizer motor weak and nozzle damaged. Need to replace these spares for further checking and working condition of the equipment.

Recommendations for repair (required service details): Not recommending for repair.

Cost of spares (specify parts and cost): NA

# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 1353/-
--	---------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not):
 Cyrix Service report attached

Reasons for recommending the equipment as BER: Checked and found Nebulizer motor weak and nozzle damage. The unit is installed on 12/12/2014 and covered up to 9+ years. Since the spares are not available in the market, recommending the unit for condemnation.

[Signature]
 Name & Signature of CYRIX Authority with date



Remarks of Junior Consultant (Biomedical) NHM:
Motor weak and nozzle damaged. . . This can be RBER as per criteria.

Recommended for BER (Yes/ No):
 Date: 7.12.2024

[Signature]
MARISHA MOHAN R.R.
 Junior Consultant (Biomedical)
 National Health Mission
 Thiruvananthapuram
 Signature of JC BM (NHM)

Date:  Seal

[Signature]
MEDICAL OFFICER
 UPHC, RAJAJI NAGAR, TVPM
 Officer (i/c)



MEDICAL OFFICER
UPHC, RAJAJI NAGAR, TVPM