



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No: WD-37/2021-2022/595

No: 135058

**CYRIX**<sup>(R)</sup>  
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrilx.com | Email : bomp.kl@cyrilx.in

Health Facility: <u>W &amp; C Hospital</u>	Call Registration Date: <u>27-11-2024</u>
Address: <u>Thy cad</u>	Caller ID: <u>149287</u>
<u>Thiruvananthapuram</u>	Date of Visit: <u>27-11-2024</u>
Ph: <u>9495243350</u>	Asset No: <u>0115036</u>
	EQPT Name: <u>BP Apparatus</u>
	Manufacture Life Line Model: <u>na</u>
	S.No: <u>na</u> Dept: <u>Ward 5</u>

Service Classification : Breakdown call  PMS  Callbration  Cust. Training

Problem Identified : Not working

Action Taken : Checked and found body damaged tank defective, hinges damaged mercury spillage. Need to replace these spares for further checking and working condition of the equipment.

Completed  Date : 27-11-2024 Time : 1:30pm Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.	-	-	-
2.	-	-	-
3.	-	-	-

Cyrilx Engineer	Date	Start Time	End Time
<u>Rehna R</u>	<u>27-11-2024</u>	<u>1:10 P.m</u>	<u>1:30 P.m</u>


Customer Remark Completed

Service Engineer Name : Rehna R

Signature: Rehna R

Date: 27-11-2024

Contact Number: 944723995



Customer Name: Shamla Malik.M.  
Signature: [Signature]  
Date: 27/11/2024  
Contact Number: 9495243350  
Designation: Senior Nursing Officer  
Hospital Seal

*Handwritten notes and stamps:*  
11/2/24  
SUPERINTENDENT  
Ward 5  
Thiruvananthapuram  
14

PROFORMA

RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (BER)  
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: WOMEN & CHILD HOSPITAL THYCAUD	Name of District: THIRUVANANTHAPURAM
Name of Equipment: BP APPARATUS	Equipment ID & Barcode: 149287/0115036
Make: DIAMOND	Model: DELUXE
Serial Number: NA	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 22/06/2016	Present status of the equipment: Fully Damaged
Date of breakdown: 27/11/2024 (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked and found Mercury spillage. Identified mercury tank, hinges and body damage. Need to replace these spares for further checking and working condition of the equipment.	
Recommendations for repair (required service details): Not recommending for repair.	
Cost of spares (specify parts and cost): NA	

# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: NA	Asset Value: 1323/-
--	---------------------


Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not):  
Cyrix Service report attached

Reasons for recommending the equipment as BER: Checked and found mercury tank, hinges and body damage. The unit is installed on 22/06/2016 and covered up to 8+ years. Since the mercury is not available in the market, recommending the unit for condemnation.

*[Handwritten Signature]*

Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:  
Mercury leak. So can be BER.

Recommended for BER (Yes/ No):	 <b>MANESHA MOHAN M.R.</b> Junior Consultant (Biomedical) National Health Mission Thiruvananthapuram Signature of JC BM (NHM)
Date: 7-12-2024	

Date: \_\_\_\_\_ Seal \_\_\_\_\_ Signature of Superintendent Officer (i/c)  
**SUPERINTENDENT**  
 Govt. W & C Hospital, Thycaud  
 Thiruvananthapuram - 14



STOCK REGISTER OF B.P. Apparatus  
 Name of Article B.P. Apparatus

Date	No. and date of voucher or invoice	From whom received or to whom issued	Receipt	Issued	Balance after each transaction	Initials of Receiver	Remarks
1/9/12		Rd from <sup>Receipt forwarded</sup> Old Stock Register Page No. 128 (2 Nos & Repaired MAR cont.)			4		
11/2/14		Rd from store MSRP Page No 390 V1 (Local purchase)	1		5		
5/1/15		Condemed as per order No. 7154/2014/WdCH dt 5/1/2015		②	3		
6/10/2015		Condemed as per order No Ag 1978/2015/WdCHT dt 6/10/2015		①	2		
26/2/2016		Rd from store (Local purchase)	1		3		
22/6/2016		Rd from stock. (local purchase)	1		4		
		Bar code 005036 coil code 5/7/22 blue code 4/4/22					

Carried over



11/12/2016  
 Govt. W & C Hospital, Thiruvananthapuram

22  
11/2/24

SUPERINTL. ENI  
Govt. W & C Hospital, Thycano  
Thiruvananthapuram - 14

